


# TECHNICAL PROPOSAL EVALUATION REPORT

<b>Name of Procuring Agency</b>	<b>ESTABLISHMENT OF 50 BEDDED CARDIAC HOSPITAL GILGIT DEPARTMENT OF HEALTH GOVERNMENT OF GILGIT-BALTISTAN</b>
<b>Title of Procurement</b>	<b>INVITATION TO BIDS FOR PROCUREMENT OF BIOMEDICAL EQUIPMENT</b>



**50 Bedded Cardiac Hospital Gilgit**  
**Department of Health**  
**Government of Gilgit-Baltistan**

A handwritten signature in blue ink, slanted diagonally across the page.

<b>1</b>	<b><u>BACKGROUND:</u></b>
<b>1.1</b>	Department of Health, Gilgit Baltistan invited bids from the eligible suppliers for the procurement of biomedical equipment under the PSDP Project titled "Establishment of 50 bedded Cardiac Hospital, Gilgit."







<b>2. SUMMARY OF KEY INFORMATION OF PROCUREMENT PROCESS</b>		
<b>TECHNICAL EVALUATION BASIC DATA</b>		
(i)	<b>Name of Procuring Agency:</b>	Establishment of 50 Bedded Cardiac Hospital Gilgit Department of Health Government of Gilgit-Baltistan
(ii)	<b>Method of Procurement</b>	Single Stage - Two Envelopes Procedure.
(iii)	<b>Title of Procurement</b>	INVITATION TO BIDS FOR PROCUREMENT OF BIOMEDICAL EQUIPMENT
(iv)	<b>Tender Publication:</b>	Daily Newspapers, GB-PPRA & PPRA Websites
	<b>DETAIL OF TENDER PUBLICATION</b>	
		Leading daily newspapers English, Urdu:





(a)	<b>Tender Notice in the Newspapers</b>	<b>DAILY BUSINESS RECORDER (NATIONAL, ENGLISH)</b>
		<i>DATED</i> 13 <sup>th</sup> September, 2024
		<b>DAILY DUNYA, ISLAMABAD (NATIONAL, URDU)</b>
		<i>DATED</i> 13 <sup>th</sup> September, 2024
		<b>DAILY K-2 GB (LOCAL, URDU)</b>
		<i>DATED</i> 13 <sup>th</sup> September, 2024
		<b>Daily Himalaya Today GB (LOCAL, URDU)</b>
		<i>DATED</i> 13 <sup>th</sup> September, 2024
		<b>Daily Inhisar, GB</b>
		<i>DATED</i> 13 <sup>th</sup> September, 2024
(b)	<b>Advertisement on GB-PPRA website</b>	<b>GB-PPRA TENDER NO. TSE-20240913215</b>
		<b>DATED OF PUBLICATION: 14<sup>TH</sup> SEPTEMBER, 2024</b>
(c)	<b>Advertisement on PPRA website</b>	<b>PPRA TENDER NO. TS-547679E</b>
		<b>DATED OF PUBLICATION: 12<sup>TH</sup> SEPTEMBER, 2024</b>

(v)	Date and Time of Bid Submission Closing	10 <sup>TH</sup> October, 2024 at 01:30 p.m																
(vi)	Date and Time of Technical Bid Opening	10 <sup>TH</sup> October, 2024 at 02:00 p.m																
(vii)	Pre-Bid Meeting	23-09-2024 at 10:00 am																
(viii)	Extended Date and Time of Bid Submission Closing	29 <sup>th</sup> October, 2024 at 01:30 p.m																
(ix)	Extended Date and Time of Technical Bid Opening	29 <sup>th</sup> October, 2024 at 02:00 p.m																
(x)	Publication Of Addendum for Extension of Bid Submission & Opening	<table><tr><td colspan="2">DAILY BUSINESS RECORDER (NATIONAL, ENGLISH)</td></tr><tr><td>DATED</td><td>09<sup>th</sup> October, 2024</td></tr><tr><td colspan="2">DAILY DUNYA, ISLAMABAD (NATIONAL, URDU)</td></tr><tr><td>DATED</td><td>09<sup>th</sup> October, 2024</td></tr><tr><td colspan="2">DAILY K-2 GB (LOCAL, URDU)</td></tr><tr><td>DATED</td><td>09<sup>th</sup> October, 2024</td></tr><tr><td colspan="2">Daily Himalaya Today GB (LOCAL, URDU)</td></tr><tr><td>DATED</td><td>09<sup>th</sup> October, 2024</td></tr></table>	DAILY BUSINESS RECORDER (NATIONAL, ENGLISH)		DATED	09 <sup>th</sup> October, 2024	DAILY DUNYA, ISLAMABAD (NATIONAL, URDU)		DATED	09 <sup>th</sup> October, 2024	DAILY K-2 GB (LOCAL, URDU)		DATED	09 <sup>th</sup> October, 2024	Daily Himalaya Today GB (LOCAL, URDU)		DATED	09 <sup>th</sup> October, 2024
DAILY BUSINESS RECORDER (NATIONAL, ENGLISH)																		
DATED	09 <sup>th</sup> October, 2024																	
DAILY DUNYA, ISLAMABAD (NATIONAL, URDU)																		
DATED	09 <sup>th</sup> October, 2024																	
DAILY K-2 GB (LOCAL, URDU)																		
DATED	09 <sup>th</sup> October, 2024																	
Daily Himalaya Today GB (LOCAL, URDU)																		
DATED	09 <sup>th</sup> October, 2024																	



		Daily Inhisar, GB	
		DATED	09 <sup>th</sup> October, 2024
(xi)	Addendum on GB-PPRA website	09 <sup>th</sup> October, 2024	
(xii)	Addendum on PPRA website	09 <sup>th</sup> October, 2024	
(xiii)	Total Bids Received against each package on Closing Date & Time and No. of Bids Qualified in Technical Proposal Evaluation;		
Package No.	Package Name	No. Of Bids Received	No. Of Bids Qualified in Technical Evaluation
1	Anesthesia machine with ventilator	03	01
2	Bipap & CPAP	02	02
3	Cerebral oximeter	02	01
4	Docking Station, infusion & syringe pump	04	01
5	ICU Ventilator	04	01
6	Neonatal Ventilator	02	02
7	ETT Machine	04	02
8	Holter Monitor	02	01
9	Telemetry System	01	01
10	Cardiac Monitor	03	02
11	Defibrillator	03	01
12	ECG Machine	04	01
13	Signal Average Machine	02	01
14	Tilt system	02	01
15 - 19	Scopes	02	00
20	Modular OT	02	02
21	Heart Lung Machine	02	02
22	IABP	02	01
23	Hypothermia	04	01
24	Blood warmer	03	01
25	Fabricator with alligators	Nil	00
26-29	OT Lights	05	01
30	Head Lights	Nil	00
31	Operation table	03	00
32	Gyne OT table	01	00

33	Vacuum extractor	Nil	00
34	Delivery table	02	00
35	Suction machine	03	00
36	Diathermy machine	02	00
37	Sternum saw	02	01
38	Sternum saw RS	02	01
39	ACT Machine	02	01
40	Gama camera	01	01
41	Echo	02	01
42	Ultrasound machine	02	00
43	Temporary pace maker	01	01
44	Blood gas analyzer	04	00
45	Elisa system	02	00
46	ESR machine	01	00
47	Cell Separator	01	00
48	Chemistry analyzer	03	00
49	Coagulation Meter	02	00
50	Spectro photo meter	02	00
51	Hematology analyzer	03	00
52	Table top autoclave	03	01
53	Laboratory sterilizer	02	00
54	CSSD	02	02
55	Laundry	04	01
56	Medical gasses	04	01
57	Biomed test instrument	Nil	00
58	Dental treatment unit	04	01
59	Complete set of surgical instruments	04	00
60	Gynae Instrument Complete	05	01

03	<b>PREPERATION OF THE BIDDING DOCUMENTS:</b>
3.1	In order to prepare the bidding documents meeting of the Procurement committee notified vide Health Secretariat Notification No. Sec-H-6(104)/2022-23n dated 12 <sup>th</sup> February, 2024 (attached vide <b>Annex: A</b> ) was convened at GB house Islamabad from 24 <sup>th</sup> to 27 <sup>th</sup> May 2024, minutes of meeting attached vide <b>Annex: B</b> .
3.2	The draft bidding documents were submitted to Health Secretariat for approval and in response Health Secretariat GB vide letter no. Sec-H-6(104)/2024-25/688 dated 28 <sup>th</sup> August, 2024 issued directives to amend the draft bidding documents. Letter attached vide <b>Annex: C</b> .



<b>3.3</b>	In order to comply with the directives of Health Secretariat, meeting of the procurement committee was convened from 31 <sup>st</sup> August, 2024 to 01 <sup>st</sup> September, 2024 at GB House Islamabad to finalize the bidding document. Minutes of meeting attached vide <b>Annex: D</b> .
<b>04</b>	<b>APPROVAL OF THE BIDDING DOCUMENTS:</b>
	Draft Bidding documents incorporating the directives of Health Secretariat were submitted to Health Secretariat GB for approval and the same was approved by the PAO/Secretary Health GB vide Health Secretariat letter no. Sec-H-6(104)/2024-25 dated 6 <sup>th</sup> September, 2024. Letter attached vide <b>Annex: E</b>
<b>05</b>	<b>PRE-BID MEETING:</b>
	As per the advertisement & approved bidding document, the Pre-Bid meeting was held on 23-09-2024 at 10:00 am at the conference Hall of Gilgit Cardiac Hospital. Minutes of meeting attached vide <b>Annex: F</b>
<b>06</b>	<b>EXTENSION IN BID SUBMISSION &amp; CLOSING DATE</b>
	Initially <u>10<sup>TH</sup> October, 2024 at 01:30 p.m</u> was set as last date for submission of bids. During the pre-bid meeting majority of the bidders requested for the extension of bid submission date therefore the last date of submission of bids were extended to <u>29<sup>th</sup> October, 2024 at 01:30 p.m and bid opening was fixed on 29<sup>th</sup> October, 2024 at 02:00 p.m.</u>
<b>07</b>	<b>MEETING OF GREVIANCE REDRESSAL COMMITTEE:</b>
<b>7.1</b>	In order to address the grievances of bidders on the bidding documents the meeting of GRC notified by the Health Secretariat vide Notification No. Sec-H-6(104)/2023-24 dated 20 <sup>th</sup> January, 2024 (attached vide <b>Annex: G</b> ) was convened on 15 <sup>th</sup> October, 2024 at EPI Conference Hall.
<b>7.2</b>	Recommendations/Decisions of the GRC were made part of the Bidding Documents and minutes were issued to all bidders. Minutes of the GRC attached vide <b>Annex: H</b> .
<b>08</b>	<b>OPENING OF PROPOSALS:</b>
	In order to open the bids meeting of the Procurement Committee was convened on <b>29<sup>th</sup> October 2024 at 1:30PM</b> in the conference hall of Gilgit Cardiac Hospital. Minutes of the meeting attached vide <b>Annex: I</b>

Handwritten signatures and initials in blue ink are present at the bottom of the page, including a large signature on the left, several smaller initials in the center, and a signature on the right with the number '7' written above it.



<b>09</b>	<b>MEETING OF THE PROCUREMENT COMMITTEE FOR EVALUATION OF THE TECHNICAL PROPOSAL:</b>
<b>9.1</b>	In order to evaluate the technical proposal meeting of the procurement committee was convened from 10 <sup>th</sup> December, 2024 to 17 <sup>th</sup> December, 2024 at Conference Hall of Cardiac Hospital Gilgit.
<b>9.2</b>	Following members of the procurement committee attended the meeting: Attendance Sheet attached vide <b>Annex: J</b> .

<b>S#</b>	<b>Name</b>	<b>Designation</b>
1	Dr Syed Sadiq Shah	Director, EPI-GB
2	Engr. Shahid Masood	PD, 50 bedded Cardiac Hospital Gilgit
3	Syed Ishtiaq Hussain	Director, SHPI-GB
4	Dr. Faheem Iqbal	Consultant Cardiologist, 50 bedded Cardiac Hospital Gilgit
5	Aziz Haider	DD(P), DHS Baltistan
6	Engr. Shabbir Hussain	BME, DHS(P) Gilgit
7	Muhammad Ali	Finance Officer, FD-GB
8	Muhammad Harith	RO, P&D D GB
9	Engr. Mohsin Ali	BME, Shifa Int Hosp Islamabad
10	Engr. Rizwan	BME, RIC Rawalpindi
11	Engr. Umair Kabeer	BME, PIC Lahore

Engr. Irshad Wali, also attended the proceedings and facilitated the Technical Evaluation as representative of the Consultant PEPAC in compliance with the Health Secretariat Letter No. Sec-H-6(104)/2024-25 dated 16<sup>th</sup> September, 2024. Letter Attached vide **Annex: K**.

<b>10</b>	<b>PROCEEDINGS OF TECHNICAL EVALUATION:</b>
<b>10.1</b>	Package wise evaluation of the technical proposals was carried out by the Procurement committee.
<b>10.2</b>	Procurement committee thoroughly examined the technical proposals in accordance with the Technical Evaluation criteria mentioned in the Bidding Documents.
<b>11</b>	<b>VERIFICATIONS OF DOCUMENTARY EVIDENCES INCLUDING CDR:</b>
	The Procurement committee decided that the documentary evidences including CDRs provided by the bidder may be verified if deemed necessary at any stage of the procurement process. If any of the documentary evidences provided by the bidders is found fake the bidder will be disqualified at any stage during the procurement process.
<b>12</b>	<b>TECHICAL EVALUATION RESULTS:</b>
	Package wise Technical Evaluation result sheets are as follows:



Item No. 01  
Item Name: Anesthesia Machine with Ventilator  
Tender Name: Procurement Of Bio-Medical Equipment Under PSDP Project "Establishment of 50 Bedded Cardiac Hospital, Gilgit".

PART - I KNOCKED DOWN CRITERIA (BID EVALUATION & VENDOR EVALUATION)				
Sr. #	Evaluation Parameters	M/s Noor International	M/s Total Technologies Pvt Ltd.	M/s Shirazi Trading Pvt. Ltd.
1.	Bid Form	Yes	Yes	Yes
2.	Bidder's Information Form	Yes	Yes	Yes
3.	Documentary proof active of National Tax Number (NTN) and General Sales Tax (GST) Number and Registration Professional Tax Certificate.	Yes	Yes	Yes
4.	Original Bid Security (In form of CDR) & Form	No	Yes	Yes
5.	Bid validity time is as per demand	Yes	Yes	Yes
6.	Warranty Form as per Tender	Yes	Yes	Yes
7.	Delivery as per Schedule of Requirement	Yes	Yes	Yes
8.	Undertaking on legal valid and attested stamp paper Rs. 100/- that the firm is not blacklisted by any of Provincial or Federal Government Department, Agency, Organization or autonomous body or Private Sector Organization anywhere in Pakistan.	Yes	Yes	Yes
9.	Embassy Attested Exclusively Authorized Agent / Sole Distributor Certificate by the Original Equipment Manufacturer (OEM).	Yes	Yes	Yes
10.	Sub-Distributor Certificate from Sole Distributor having consent of OEM. (Where Applicable)	N/A	N/A	N/A
11.	Proof for having minimum 5 years of experience in supply, installation, commissioning of biomedical equipment.	Yes	Yes	Yes
12.	At least 05 verifiable Past Performance Certificates by the Procuring Agency of quoted products issued against the same Supply Orders.	Yes	Yes	Yes
13.	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes	Yes	Yes

14.	Originally signed documentary evidence from the Manufacturer that the quoted brand is the Original Equipment Manufacturer (OEM) with indication of manufacturing site and its location.	Yes	Yes	Yes
15.	Up to Date After Sales Service Certificate from OEM/Principal	Yes	Yes	Yes
16.	Availability of relevant Tools and Testing / Calibration Equipment.	Yes	Yes	Yes
17.	A Certificate from the OEM that the installation will be conducted in conformity with the system requirements.	Yes	Yes	Yes
18.	Consortiums or Joint Ventures (JVs) (Where applicable)	N/A	N/A	N/A
19.	Complete Package / Tender	No	Yes	Yes
Remarks:		Not Eligible	Eligible	Eligible
<b>PART – II</b> <b>KNOCKED DOWN CRITERIA</b> <b>(PRODUCT EVALUATION)</b>				
Item Ad. No.	Anesthesia Machine with Ventilator	M/s Noor International	M/s Total Technologies Pvt Ltd.	M/s Shirazi Trading Pvt. Ltd.
	Brand	Penlon	Lowenstein Medical	GE HealthCare
01	Quoted Model	Anesthesia: Prima 460, Vital Sign Monitor: Any View P-15	Anesthesia: Leon Plus, Vaporizer: Sigma Delta, Vital Sign Monitor: Efficia CM 150	Anesthesia: Carestation 650, Vital Sign Monitor: CS B650
1.	Country of Manufacturer	UK	Germany & UK	USA
2.	Country of Origin of Product	UK	Germany/(ROC/India)/UK	USA & Finland
3.	Compliance with defined Quality Standards	(Non Complied) Not Attached with Technical Offer	(Compliance) CE (MDD) & MHLW	(Compliance) CE (MDR), MHLW and FDA 510k
4.	Specifications Compliance Feature Wise	(Non Complied)  1- Neonate patients mode not verified on catalogue. 2- Built in Oxygen	1- The Vital Sign Monitor Efficia CM 150 is an Indian Origin Monitor, while in technical offer the make is declared as	Compliance



		<p>Monitor with Paramagnetic technology or Non consumable Oxygen Sensor not verified from catalogues not available.</p> <p>3- Work Surface is of aluminum and plastic instead of Stainless steel or fiber.</p> <p>4- Self-test mode not available as per catalogue. (Technically not Acceptable)</p>	<p>2- Lowenstein Medical. Stainless steel or fiber work surface cannot be verified from catalogue.</p>	
5.	Technical Eligibility of Product	Not Eligible	Not Eligible	Eligible
6.	Technical Eligibility of Firm	Not Eligible	Eligible	Eligible
<b>Technical Status:</b>		<b>NON RESPONSIVE</b>	<b>NON RESPONSIVE</b>	<b>RESPONSIVE</b>

Item No.  
Item Name:  
Tender Name:

02  
Bi-PAP Unit and CPAP  
Procurement of Bio-Medical Equipment Under PSDP Project  
"Establishment of 50 Bedded Cardiac Hospital, Gilgit".

PART – I KNOCKED DOWN CRITERIA (BID EVALUATION & VENDOR EVALUATION)			
Sr. #	Evaluation Parameters	M/s Total Technologies Pvt. Ltd.	M/s Jamil Enterprises
1	Bid Form	Yes	Yes
2	Bidder's Information Form	Yes	Yes
3	Documentary proof active of National Tax Number (NTN) and General Sales Tax (GST) Number and Registration Professional Tax Certificate.	Yes	Yes
4	Original Bid Security (In form of CDR) & Form	Yes	Yes
5	Bid validity time is as per demand	Yes	Yes
6	Warranty Form as per Tender	Yes	Yes
7	Delivery as per Schedule of Requirement	Yes	Yes
8	Undertaking on legal valid and attested stamp paper Rs. 100/- that the firm is not blacklisted by any of Provincial or Federal Government Department, Agency, Organization or autonomous body or Private Sector Organization anywhere in Pakistan.	Yes	Yes
9	Embassy Attested Exclusively Authorized Agent / Sole Distributor Certificate by the Original Equipment Manufacturer (OEM).	Yes	Yes (For M/s Electro Med Cooperation)
10	Sub-Distributor Certificate from Sole Distributor having consent of OEM. (Where Applicable)	N/A	Yes
11	Proof for having minimum 5 years of experience in supply, installation, commissioning of biomedical equipment.	Yes	Yes
12	At least 05 verifiable Past Performance Certificates by the Procuring Agency of quoted products issued against the same Supply Orders.	Yes	Yes
13	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes	Yes
14	Originally signed documentary evidence from the Manufacturer that the quoted brand is the Original Equipment Manufacturer (OEM) with indication of manufacturing site and its location.	Yes	Yes
15	Up to Date After Sales Service Certificate from OEM/Principal	Yes	Yes
16	Availability of relevant Tools and Testing / Calibration Equipment.	Yes	Yes
17	A Certificate from the OEM that the installation will be conducted in conformity with the system requirements.	Yes	Yes
18	Consortiums or Joint Ventures (JVs) (Where applicable)	N/A	N/A





19	Complete Package / Tender	Yes	Yes
Remarks:		Eligible	Eligible
<b>PART – II</b> <b>KNOCKED DOWN CRITERIA</b> <b>(PRODUCT EVALUATION)</b>			
Item Ad. No.	Bi-PAP Unit and CPAP	M/s Total Technologies Pvt. Ltd.	M/s Jamil Enterprises
	Brand	Lowenstein Medical	Philips Respironics
02	Quoted Model	Prisma 30ST & Prisma 20A	Dream Station BiPAP AVAPS & Dream Station Auto CPAP
1	Country of Manufacturer	Germany	USA
2	Country of Origin of Product	Germany	USA
3	Compliance with defined Quality Standards	(Compliance) CE (MDD) & FDA 510K	(Compliance) CE (MDD) & FDA 510K
4	Specifications Compliance Feature Wise	Compliance	Compliance
5	Technical Eligibility of Product	Eligible	Eligible
6	Technical Eligibility of Firm	Eligible	Eligible
Technical Status:		RESPONSIVE	RESPONSIVE

The block contains several handwritten signatures and initials in blue ink. There are approximately 10-12 distinct marks, including long horizontal strokes, loops, and stylized letters, scattered across the lower half of the page.

Item No. 03  
Item Name: Cerebral Oximeter  
Tender Name: Procurement of Bio-Medical Equipment Under PSDP Project  
"Establishment of 50 Bedded Cardiac Hospital, Gilgit".

PART – I KNOCKED DOWN CRITERIA (BID EVALUATION & VENDOR EVALUATION)			
Sr. #	Evaluation Parameters	M/s Radiant Medical Pvt. Ltd.	M/s Hospicare Systems
1	Bid Form	Yes	Yes
2	Bidder's Information Form	Yes	Yes
3	Documentary proof active of National Tax Number (NTN) and General Sales Tax (GST) Number and Registration Professional Tax Certificate.	Yes	Yes
4	Original Bid Security (In form of CDR) & Form	Yes	No
5	Bid validity time is as per demand	Yes	Yes
6	Warranty Form as per Tender	Yes	Yes
7	Delivery as per Schedule of Requirement	Yes	Yes
8	Undertaking on legal valid and attested stamp paper Rs. 100/- that the firm is not blacklisted by any of Provincial or Federal Government Department, Agency, Organization or autonomous body or Private Sector Organization anywhere in Pakistan.	Yes	Yes
9	Embassy Attested Exclusively Authorized Agent / Sole Distributor Certificate by the Original Equipment Manufacturer (OEM).	Yes	No (Embassy attested certificate not provided)
10	Sub-Distributor Certificate from Sole Distributor having consent of OEM. (Where Applicable)	N/A	N/A
11	Proof for having minimum 5 years of experience in supply, installation, commissioning of biomedical equipment.	Yes	Yes
12	At least 05 verifiable Past Performance Certificates by the Procuring Agency of quoted products issued against the same Supply Orders.	Yes	No
13	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes	Yes
14	Originally signed documentary evidence from the Manufacturer that the quoted brand is the Original Equipment Manufacturer (OEM) with indication of manufacturing site and its location.	Yes	Yes
15	Up to Date After Sales Service Certificate from OEM/Principal	Yes	Yes
16	Availability of relevant Tools and Testing / Calibration Equipment.	Yes	Yes



17	A Certificate from the OEM that the installation will be conducted in conformity with the system requirements.	Yes	Yes
18	Consortiums or Joint Ventures (JVs) (Where applicable)	N/A	N/A
19	Complete Package / Tender	Yes	No
Remarks:		Eligible	Not Eligible
<b>PART – II</b> <b>KNOCKED DOWN CRITERIA</b> <b>(PRODUCT EVALUATION)</b>			
Item Ad. No.	Cerebral Oximeter	M/s Radiant Medical Pvt. Ltd.	M/s Hospicare Systems
	Brand	Nonin Medical Inc.	Masimo
03	Quoted Model	Sense Smart X-100	Roots
1	Country of Manufacturer	USA	USA
2	Country of Origin of Product	USA	Maxico
3	Compliance with defined Quality Standards	(Compliance) CE (MDD) & FDA 510K	(Compliance) CE (MDD) & FDA 510K
4	Specifications Compliance Feature Wise	Compliance	(Non Complied) Country of Origin and Country manufacturer are not from USA, Europe or Japan.
5	Technical Eligibility of Product	Eligible	Not Eligible
6	Technical Eligibility of Firm	Eligible	Not Eligible
Technical Status:		RESPONSIVE	NON RESPONSIVE

Item No. 04

Item Name: Docking Station, Infusion Pump and Syringe Pump

Tender Name: Procurement Of Bio-Medical Equipment Under PSDP Project "Establishment of 50 Bedded Cardiac Hospital, Gilgit".

PART – I KNOCKED DOWN CRITERIA (BID EVALUATION & VENDOR EVALUATION)					
Sr. #	Evaluation Parameters	M/s Total Technologies Pvt. Ltd.	M/s Jamil Enterprises	M/s Irfan Brothers International	M/s Medifa Enterprises
1	Bid Form	Yes	Yes	Yes	Yes
2	Bidder's Information Form	Yes	Yes	Yes	Yes
3	Documentary proof active of National Tax Number (NTN) and General Sales Tax (GST) Number and Registration Professional Tax Certificate.	Yes	Yes	Yes	Yes
4	Original Bid Security (In form of CDR) & Form	Yes	Yes	Yes	No
5	Bid validity time is as per demand	Yes	Yes	Yes	Yes
6	Warranty Form as per Tender	Yes	Yes	Yes	Yes
7	Delivery as per Schedule of Requirement		Yes	Yes	Yes
8	Undertaking on legal valid and attested stamp paper Rs. 100/- that the firm is not blacklisted by any of Provincial or Federal Government Department, Agency, Organization or autonomous body or Private Sector Organization anywhere in Pakistan.	Yes	Yes	Yes	Yes
9	Embassy Attested Exclusively Authorized Agent / Sole Distributor Certificate by the Original Equipment Manufacturer (OEM).	No (Embassy attested Certificate not provided)	Yes (For M/s Medical Equipment and Systems)	Yes (For M/s Human Healthcare)	No (Embassy attested Certificate not provided)
10	Sub-Distributor Certificate from Sole Distributor having consent of OEM. (Where Applicable)	N/A	Yes	Yes	N/A
11	Proof for having minimum 5 years of experience in supply, installation, commissioning of biomedical equipment.	Yes	Yes	Yes	Yes



12	At least 05 verifiable Past Performance Certificates by the Procuring Agency of quoted products issued against the same Supply Orders.	No	Yes	Yes	Yes
13	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes	Yes	Yes	Yes
14	Originally signed documentary evidence from the Manufacturer that the quoted brand is the Original Equipment Manufacturer (OEM) with indication of manufacturing site and its location.	Yes	Yes	Yes	Yes
15	Up to Date After Sales Service Certificate from OEM/Principal	Yes	Yes	Yes	Yes
16	Availability of relevant Tools and Testing / Calibration Equipment.	Yes	Yes	Yes	Yes
17	A Certificate from the OEM that the installation will be conducted in conformity with the system requirements.	Yes	Yes	Yes	Yes
18	Consortiums or Joint Ventures (JVs) (Where applicable)	N/A	N/A	N/A	N/A
19	Complete Package / Tender	No	Yes	Yes	No
Remarks:		Not Eligible	Eligible	Eligible	Not Eligible

**PART – II**  
**KNOCKED DOWN CRITERIA**  
**(PRODUCT EVALUATION)**

Item Ad. No.	Docking Station, Infusion Pump and Syringe Pump	M/s Total Technologies Pvt. Ltd.	M/s Jamil Enterprises	M/s Irfan Brothers International	M/s Medifa Enterprises
	Brand	Terumo Corporation	Fresenius Medical Care	Medicinos Gija	Sternmed
04	Quoted Model	Docking Station: TE-RS700, Infusion Pump: TE-LM730 and Syringe Pump: TE-SS830	Docking Station: Agila Link 6, Infusion Pump: Infusia VP7s and Syringe Pump: Infusia SP7s	Docking Station: Unifusion 06, Infusion Pump: Medifusion IP and Syringe Pump: Medifusion SP	Docking Station: 6 Channel, Infusion Pump: IPN 56 and Syringe Pump: SPN 58

1	Country of Manufacturer	Japan	Germany	Lithuania	Germany
2	Country of Origin of Product	Japan	France & China	Lithuania	Germany
3	Compliance with defined Quality Standards	(Compliance) CE (MDD)	(Compliance) CE (MDD)	(Non Complied) No (CE (MDD) from authorized EU body not available.)	(Compliance) CE (MDD)
4	Specifications Compliance Feature Wise	1. Embassy attested Certificate not provided). 2. 05 verifiable Past Performance Certificates of quoted brand not attached.	Compliance	Compliance	1. Bolus rate: 0.1 to 1200 / hr. cannot be verified from catalogue. 2. 5 and 60 ml syringe sizes not available in Syringe Pump. 3. 1 ml increment not available in Infusion Pump.
5	Technical Eligibility of Product	Eligible	Eligible	Not Eligible	Not Eligible
6	Technical Eligibility of Firm	Not Eligible	Eligible	Eligible	Not Eligible
Technical Status:		NON RESPONSIVE	RESPONSIVE	NON RESPONSIVE	NON RESPONSIVE



Item No. 05

Item Name: ICU Ventilator

Tender Name: Procurement Of Bio-Medical Equipment Under PSDP Project "Establishment of 50 Bedded Cardiac Hospital, Gilgit".

PART – I KNOCKED DOWN CRITERIA (BID EVALUATION & VENDOR EVALUATION)					
Sr. #	Evaluation Parameters	M/s Total Technologies Pvt. Ltd.	M/s Sarim Enterprises (JV with Chinara International)	M/s Sahaab International	M/s Noor International
1	Bid Form	Yes	As per BD ITB Clause 5, a bidder is required to submit only one bid for a specific item/package. Additionally, the tender advertisement explicitly stated the requirement to submit separate proposals/bids. This was further recorded in the minutes of the pre-bid meeting and clarified by the GRC, which specified that firms must submit an original proposal (Technical and Financial) along with	Yes	Yes

			one photocopy in hard form and one in soft form (in USB) for each item/package. Contrary to these requirements, the firm has submitted a single proposal for multiple items (Item No. 05 & 06).		
2	Bidder's Information Form	Yes	Yes	Yes	Yes
3	Documentary proof active of National Tax Number (NTN) and General Sales Tax (GST) Number and Registration Professional Tax Certificate.	Yes	Yes	Yes	Yes
4	Original Bid Security (In form of CDR) & Form	Yes	Yes	Yes	No
5	Bid validity time is as per demand	Yes	Yes	Yes	Yes
6	Warranty Form as per Tender	Yes	Yes	Yes	Yes
7	Delivery as per Schedule of Requirement		Yes	Yes	Yes
8	Undertaking on legal valid and attested stamp paper Rs. 100/- that the firm is not blacklisted by any of Provincial or Federal Government Department, Agency, Organization or autonomous body or Private Sector Organization anywhere in Pakistan.	Yes	No	No	Yes
9	Embassy Attested Exclusively Authorized Agent / Sole Distributor Certificate by the Original Equipment Manufacturer (OEM).	Yes	No (For M/s Med Express is not a sole agent of quoted brand and unattested exclusive authorization attached.)	No (Embassy attested certificate not provided)	Yes
10	Sub-Distributor Certificate from Sole Distributor having consent of OEM. (Where Applicable)	N/A	N/A	N/A	N/A



11	Proof for having minimum 5 years of experience in supply, installation, commissioning of biomedical equipment.	Yes	Yes	Yes	Yes
12	At least 05 verifiable Past Performance Certificates by the Procuring Agency of quoted products issued against the same Supply Orders.	Yes	Yes	No	Yes
13	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes	Yes	Yes	Yes
14	Originally signed documentary evidence from the Manufacturer that the quoted brand is the Original Equipment Manufacturer (OEM) with indication of manufacturing site and its location.	Yes	Yes	Yes	Yes
15	Up to Date After Sales Service Certificate from OEM/Principal	Yes	Yes	Yes	Yes
16	Availability of relevant Tools and Testing / Calibration Equipment.	Yes	Yes	Yes	Yes
17	A Certificate from the OEM that the installation will be conducted in conformity with the system requirements.	Yes	Yes	Yes	Yes
18	Consortiums or Joint Ventures (JVs) (Where applicable)	N/A	No (Not as per the format of BD) M/s Chinar International and Sarim Enterprises)	N/A	N/A
19	Complete Package / Tender	Yes	Yes	No	No
Remarks:		Eligible	Not Eligible	Not Eligible	Not Eligible
PART - II KNOCKED DOWN CRITERIA (PRODUCT EVALUATION)					
Item Ad. No.	ICU Ventilator	M/s Total Technologies Pvt. Ltd.	M/s Sarim Enterprises (JV with Chinar International)	M/s Sahaab International	M/s Noor International

	Brand	Lowenstein Medical	N/A	Medallion Equipment	Hamilton Medical AG
05	Quoted Model	Elisa 300	N/A	MV-M70	Hamilton-C3
1.	Country of Manufacturer	Germany	N/A	UK	Switzerland
2.	Country of Origin of Product	Germany	N/A	Not Mentioned	Switzerland
3.	Compliance with defined Quality Standards	(Compliance) CE (MDD) & MHLW	N/A	(Non complied) (Expired Certificate and (CE (MDD) from authorized EU body not available.) and other certification not provided	(Compliance) CE (MDD) & FDA 510K
4.	Specifications Compliance Feature Wise	Compliance	N/A	The required specifications were copied on product catalog attached to the technical offer; moreover, the product information also could not be verified from the manufacturer's website.	Compliance
5.	Technical Eligibility of Product	Eligible	N/A	Not Eligible	Eligible
6.	Technical Eligibility of Firm	Eligible	Not Eligible	Not Eligible	Not Eligible
Technical Status:		RESPONSIVE	NON RESPONSIVE	NON RESPONSIVE	NON RESPONSIVE




Item No.  
Item Name:  
Tender Name:

06  
Neonatal Ventilator  
Procurement of Bio-Medical Equipment Under PSDP Project  
"Establishment of 50 Bedded Cardiac Hospital, Gilgit".

PART – I KNOCKED DOWN CRITERIA (BID EVALUATION & VENDOR EVALUATION)			
Sr. #	Evaluation Parameters	M/s Total Technologies Pvt. Ltd.	M/s Mediland Pvt. Ltd.
1.	Bid Form	Yes	Yes
2.	Bidder's Information Form	Yes	Yes
3.	Documentary proof active of National Tax Number (NTN) and General Sales Tax (GST) Number and Registration Professional Tax Certificate.	Yes	Yes
4.	Original Bid Security (In form of CDR) & Form	Yes	Yes
5.	Bid validity time is as per demand	Yes	Yes
6.	Warranty Form as per Tender	Yes	Yes
7.	Delivery as per Schedule of Requirement	Yes	Yes
8.	Undertaking on legal valid and attested stamp paper Rs. 100/- that the firm is not blacklisted by any of Provincial or Federal Government Department, Agency, Organization or autonomous body or Private Sector Organization anywhere in Pakistan.	Yes	Yes
9.	Embassy Attested Exclusively Authorized Agent / Sole Distributor Certificate by the Original Equipment Manufacturer (OEM).	Yes	Yes
10.	Sub-Distributor Certificate from Sole Distributor having consent of OEM. (Where Applicable)	N/A	N/A
11.	Proof for having minimum 5 years of experience in supply, installation, commissioning of biomedical equipment.	Yes	Yes
12.	At least 05 verifiable Past Performance Certificates by the Procuring Agency of quoted products issued against the same Supply Orders.	Yes	Yes
13.	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes	Yes
14.	Originally signed documentary evidence from the Manufacturer that the quoted brand is the Original Equipment Manufacturer (OEM) with indication of manufacturing site and its location.	Yes	Yes
15.	Up to Date After Sales Service Certificate from OEM/Principal	Yes	Yes
16.	Availability of relevant Tools and Testing / Calibration Equipment.	Yes	Yes

17.	A Certificate from the OEM that the installation will be conducted in conformity with the system requirements.	Yes	Yes
18.	Consortiums or Joint Ventures (JVs) (Where applicable)	N/A	N/A
19.	Complete Package / Tender	Yes	Yes
<b>Remarks:</b>		<b>Eligible</b>	<b>Eligible</b>
<b>PART – II KNOCKED DOWN CRITERIA (PRODUCT EVALUATION)</b>			
Item Ad. No.	Neonatal Ventilator	M/s Total Technologies Pvt. Ltd.	M/s Mediland Pvt. Ltd.
	Brand	Lowenstein Medical	Getinge
06	Quoted Model	Leoni Plus	Servo N
1.	Country of Manufacturer	Germany	Sweden
2.	Country of Origin of Product	Germany	Sweden
3.	Compliance with defined Quality Standards	(Non Complied) CE (MDD) and other certification not available.	CE (MDD) & FDA 510K
4.	Specifications Compliance Feature Wise	Compliance	Compliance
5.	Technical Eligibility of Product	Not Eligible	Eligible
6.	Technical Eligibility of Firm	Eligible	Eligible
<b>Technical Status:</b>		<b>NON RESPONSIVE</b>	<b>RESPONSIVE</b>

2





Item No. 07

Item Name: ETT Machine

Tender Name: Procurement Of Bio-Medical Equipment Under PSDP Project "Establishment of 50 Bedded Cardiac Hospital, Gilgit".

PART – I KNOCKED DOWN CRITERIA (BID EVALUATION & VENDOR EVALUATION)					
Sr. #	Evaluation Parameters	M/s Radiant Medical	M/s Jamil Enterprises	M/s Hospicare Systems	M/s Shirazi Trading Pvt. Ltd.
1.	Bid Form	Yes	Yes	Yes	Yes
2.	Bidder's Information Form	Yes	Yes	Yes	Yes
3.	Documentary proof active of National Tax Number (NTN) and General Sales Tax (GST) Number and Registration Professional Tax Certificate.	Yes	Yes	Yes	Yes
4.	Original Bid Security (In form of CDR) & Form	Yes	Yes	No	No
5.	Bid validity time is as per demand	Yes	Yes	Yes	Yes
6.	Warranty Form as per Tender	Yes	Yes	Yes	Yes
7.	Delivery as per Schedule of Requirement		Yes	Yes	Yes
8.	Undertaking on legal valid and attested stamp paper Rs. 100/- that the firm is not blacklisted by any of Provincial or Federal Government Department, Agency, Organization or autonomous body or Private Sector Organization anywhere in Pakistan.	Yes	Yes	Yes	Yes
9.	Embassy Attested Exclusively Authorized Agent / Sole Distributor Certificate by the Original Equipment Manufacturer (OEM).	Yes	Yes (For Nova Tek)	Yes	Yes
10.	Sub-Distributor Certificate from Sole Distributor having consent of OEM. (Where Applicable)	N/A	Yes	N/A	N/A
11.	Proof for having minimum 5 years of experience in supply, installation, commissioning of biomedical equipment.	Yes	Yes	Yes	Yes
12.	At least 05 verifiable Past Performance Certificates by the Procuring Agency of quoted products issued against the same Supply Orders.	Yes	No	Yes	Yes

13	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes	Yes	Yes	Yes
14	Originally signed documentary evidence from the Manufacturer that the quoted brand is the Original Equipment Manufacturer (OEM) with indication of manufacturing site and its location.	Yes	Yes	Yes	Yes
15	Up to Date After Sales Service Certificate from OEM/Principal	Yes	Yes	Yes	Yes
16	Availability of relevant Tools and Testing / Calibration Equipment.	Yes	Yes	Yes	Yes
17	A Certificate from the OEM that the installation will be conducted in conformity with the system requirements.	Yes	Yes	Yes	Yes
18	Consortiums or Joint Ventures (JVs) (Where applicable)	N/A	N/A	N/A	N/A
19	Complete Package / Tender	Yes	No	No	Yes
Remarks:		Eligible	Not Eligible	Not Eligible	Eligible
<b>PART – II</b> <b>KNOCKED DOWN CRITERIA</b> <b>(PRODUCT EVALUATION)</b>					
Item Ad. No.	Docking Station, Infusion Pump and Syringe Pump	M/s Radiant Medical	M/s Jamil Enterprises	M/s Hospicare Systems	M/s Shirazi Trading Pvt. Ltd.
	Brand	Baxter	Lab Tech	Cardioline	GE Healthcare
07	Quoted Model	X-Scribe with TMX 428	EC-12S with EC-12LT	HD+, Cube Stress Software, XR 600	Cardio Soft with T2100 ST2 Treadmill
1.	Country of Manufacturer	USA	Hungary	Italy	Germany
2.	Country of Origin of Product	USA	Hungary	Italy	Germany
3.	Compliance with defined Quality Standards	(Compliance) CE (MDD) & FDA 510K	(Compliance) CE (MDD)	(Compliance) CE (MDD)	(Compliance) CE (MDD) & FDA 510K
4.	Specifications Compliance Feature Wise	Compliance	(Non Complied)	Compliance	Compliance



			1. QRS Detection for selected ECG Leads not available. 2. Complete integrated full functional workstation not available instead of a retrofitted solution. 3. Dedicated imported cart supplied by from same manufacturer could not verified from website.		
5.	Technical Eligibility of Product	Eligible	Not Eligible	Eligible	Eligible
6.	Technical Eligibility of Firm	Eligible	Not Eligible	Not Eligible	Eligible
<b>Technical Status:</b>		<b>RESPONSIVE</b>	<b>NON RESPONSIVE</b>	<b>NON RESPONSIVE</b>	<b>RESPONSIVE</b>

Handwritten signatures and initials in blue ink, including a large stylized 'P', a signature resembling 'JGSM', a circled 'R', and several other scribbles and initials.

Item No. 08  
Item Name: Holter Monitor with 06 Recorder (Wireless)  
Tender Name: Procurement of Bio-Medical Equipment Under PSDP Project  
"Establishment of 50 Bedded Cardiac Hospital, Gilgit".

PART – I KNOCKED DOWN CRITERIA (BID EVALUATION & VENDOR EVALUATION)			
Sr. #	Evaluation Parameters	M/s Hospicare Systems	M/s Jamil Enterprises
1.	Bid Form	Yes	Yes
2.	Bidder's Information Form	Yes	Yes
3.	Documentary proof active of National Tax Number (NTN) and General Sales Tax (GST) Number and Registration Professional Tax Certificate.	Yes	Yes
4.	Original Bid Security (In form of CDR) & Form	No	Yes
5.	Bid validity time is as per demand	Yes	Yes
6.	Warranty Form as per Tender	Yes	Yes
7.	Delivery as per Schedule of Requirement	Yes	Yes
8.	Undertaking on legal valid and attested stamp paper Rs. 100/- that the firm is not blacklisted by any of Provincial or Federal Government Department, Agency, Organization or autonomous body or Private Sector Organization anywhere in Pakistan.	Yes	Yes
9.	Embassy Attested Exclusively Authorized Agent / Sole Distributor Certificate by the Original Equipment Manufacturer (OEM).	Yes	Yes (For M/s Nova Tech)
10.	Sub-Distributor Certificate from Sole Distributor having consent of OEM. (Where Applicable)	N/A	Yes
11.	Proof for having minimum 5 years of experience in supply, installation, commissioning of biomedical equipment.	Yes	Yes
12.	At least 05 verifiable Past Performance Certificates by the Procuring Agency of quoted products issued against the same Supply Orders.	Yes	Yes
13.	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes	Yes
14.	Originally signed documentary evidence from the Manufacturer that the quoted brand is the Original Equipment Manufacturer (OEM) with indication of manufacturing site and its location.	Yes	Yes
15.	Up to Date After Sales Service Certificate from OEM/Principal	Yes	Yes
16.	Availability of relevant Tools and Testing / Calibration Equipment.	Yes	Yes



17.	A Certificate from the OEM that the installation will be conducted in conformity with the system requirements.	Yes	Yes
18.	Consortiums or Joint Ventures (JVs) (Where applicable)	N/A	N/A
19.	Complete Package / Tender	Yes	Yes
Remarks:		Not Eligible	Eligible
<b>PART – II</b> <b>KNOCKED DOWN CRITERIA</b> <b>(PRODUCT EVALUATION)</b>			
Item Ad. No.	Holter Monitor With 06 Recorder (Wireless)	M/s Hospicare Systems	M/s Jamil Enterprises
	Brand	Cardioloinc	LabTech
08	Quoted Model	Walk 400H with Cardioloinc Cube Holter	EC-12H
1.	Country of Manufacturer	Italy	Hungary
2.	Country of Origin of Product	Italy	Hungary
3.	Compliance with defined Quality Standards	(Compliance) CE (MDD)	(Compliance) CE (MDD) & FDA 510K
4.	Specifications Compliance Feature Wise	Compliance	Compliance
5.	Technical Eligibility of Product	Eligible	Eligible
6.	Technical Eligibility of Firm	Not Eligible	Eligible
Technical Status:		NON RESPONSIVE	RESPONSIVE

Handwritten signatures and marks in blue ink are present below the table. The page number 29 is visible in the bottom right corner.

Item No. 09  
Item Name: Telemetry System for 08 Patients  
Tender Name: Procurement Of Bio-Medical Equipment Under PSDP Project  
"Establishment of 50 Bedded Cardiac Hospital, Gilgit".

PART – I KNOCKED DOWN CRITERIA (BID EVALUATION & VENDOR EVALUATION)		
Sr. #	Evaluation Parameters	M/s Jamil Enterprises
1.	Bid Form	Yes
2.	Bidder's Information Form	Yes
3.	Documentary proof active of National Tax Number (NTN) and General Sales Tax (GST) Number and Registration Professional Tax Certificate.	Yes
4.	Original Bid Security (In form of CDR) & Form	Yes
5.	Bid validity time is as per demand	Yes
6.	Warranty Form as per Tender	Yes
7.	Delivery as per Schedule of Requirement	Yes
8.	Undertaking on legal valid and attested stamp paper Rs. 100/- that the firm is not blacklisted by any of Provincial or Federal Government Department, Agency, Organization or autonomous body or Private Sector Organization anywhere in Pakistan.	Yes
9.	Embassy Attested Exclusively Authorized Agent / Sole Distributor Certificate by the Original Equipment Manufacturer (OEM).	Yes (For M/s Medequips (SMC) Pvt. Ltd.)
10.	Sub-Distributor Certificate from Sole Distributor having consent of OEM. (Where Applicable)	Yes
11.	Proof for having minimum 5 years of experience in supply, installation, commissioning of biomedical equipment.	Yes
12.	At least 05 verifiable Past Performance Certificates by the Procuring Agency of quoted products issued against the same Supply Orders.	Yes
13.	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes
14.	Originally signed documentary evidence from the Manufacturer that the quoted brand is the Original Equipment Manufacturer (OEM) with indication of manufacturing site and its location.	Yes
15.	Up to Date After Sales Service Certificate from OEM/Principal	Yes
16.	Availability of relevant Tools and Testing / Calibration Equipment.	Yes
17.	A Certificate from the OEM that the installation will be conducted in conformity with the system requirements.	Yes
18.	Consortiums or Joint Ventures (JVs) (Where applicable)	N/A
19.	Complete Package / Tender	Yes
Remarks:		Eligible
PART – II		



KNOCKED DOWN CRITERIA (PRODUCT EVALUATION)		
Item Ad. No.	Telemetry System For 08 Patients	M/s Jamil Enterprises
	Brand	Space Lab Healthcare
09	Quoted Model	XHIBIT 96102 & ARIATELE 96281
1.	Country of Manufacturer	USA
2.	Country of Origin of Product	USA
3.	Compliance with defined Quality Standards	(Compliance) CE MDD & FDA10K
4.	Specifications Compliance Feature Wise	Compliance
5.	Technical Eligibility of Product	Eligible
6.	Technical Eligibility of Firm	Eligible
Technical Status:		RESPONSIVE

Item No. 10

Item Name: Bedside Monitor, Bedside Monitor (15 inches), Portable Cardiac Monitors with central station, Cardiac Monitors with one central monitoring system, ICU Monitors with once Central Station Unit, ICU Monitors with once Central Station Unit, Portable Cardiac Monitor, Portable, Cardiac Monitors with central station, Pulse Oximeter

Tender Name: Procurement Of Bio-Medical Equipment Under PSDP Project "Establishment of 50 Bedded Cardiac Hospital, Gilgit".

PART – I KNOCKED DOWN CRITERIA (BID EVALUATION & VENDOR EVALUATION)				
Sr. #	Evaluation Parameters	M/s Irfan Brothers International	M/s Shirazi Trading Pvt. Ltd.	M/s Medequips (SMC) Pvt. Ltd.
1.	Bid Form	Yes	Yes	Yes
2.	Bidder's Information Form	Yes	Yes	Yes
3.	Documentary proof active of National Tax Number (NTN) and General Sales Tax (GST) Number and Registration Professional Tax Certificate.	Yes	Yes	Yes
4.	Original Bid Security (In form of CDR) & Form	Yes	Yes	Yes
5.	Bid validity time is as per demand	Yes	Yes	Yes
6.	Warranty Form as per Tender	Yes	Yes	Yes
7.	Delivery as per Schedule of Requirement	Yes	Yes	Yes
8.	Undertaking on legal valid and attested stamp paper Rs. 100/- that the firm is not blacklisted by any of Provincial or Federal Government Department, Agency, Organization or autonomous body or Private Sector Organization anywhere in Pakistan.	No (Unsigned and Unstamped)	Yes	Yes
9.	Embassy Attested Exclusively Authorized Agent / Sole Distributor Certificate by the Original Equipment Manufacturer (OEM).	No (Embassy Attested Certificate not provided)	Yes	Yes
10.	Sub-Distributor Certificate from Sole Distributor having consent of OEM. (Where Applicable)	No (The attached form is not as per format of sub-distribution form)	N/A	N/A
11.	Proof for having minimum 5 years of experience in supply, installation, commissioning of biomedical equipment.	Yes	Yes	Yes



12.	At least 05 verifiable Past Performance Certificates by the Procuring Agency of quoted products issued against the same Supply Orders.	No	Yes	Yes
13.	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes	Yes	Yes
14.	Originally signed documentary evidence from the Manufacturer that the quoted brand is the Original Equipment Manufacturer (OEM) with indication of manufacturing site and its location.	No	Yes	Yes
15.	Up to Date After Sales Service Certificate from OEM/Principal	No	Yes	Yes
16.	Availability of relevant Tools and Testing / Calibration Equipment.	Yes	Yes	Yes
17.	A Certificate from the OEM that the installation will be conducted in conformity with the system requirements.	Yes	Yes	Yes
18.	Consortiums or Joint Ventures (JVs) (Where applicable)	N/A	N/A	N/A
19.	Complete Package / Tender	No Bid Pages are not signed and stamped as required in BD ITB Clause 19.	Yes	Yes
Remarks:		Not Eligible	Eligible	Eligible
PART – II KNOCKED DOWN CRITERIA (PRODUCT EVALUATION)				
Item Ad. No.	Bedside Monitor, Bedside Monitor (15 inches), Portable Cardiac Monitors with central station, Cardiac Monitors with one central monitoring system, ICU Monitors with once Central Station Unit, ICU Monitors with once Central Station Unit, Portable, Cardiac Monitor, Portable, Cardiac Monitors with central station, Pulse Oximeter	M/s Irfan Brothers International	M/s Shirazi Trading Pvt. Ltd.	M/s Medequips (SMC) Pvt. Ltd.
	Brand	Infinium Medical	GE Healthcare	Space Lab Healthcare

11	Quoted Model	Monitor: Omni II & Omni III, Central Station: Model not quoted Pulse Oximeter: CLEO	Monitor: B105M, B155 III, Carescape 1000 with PDM, Central Station: E-COP+ E-miniC+CSCS, Pulse Oximeter: Portrait VSM	Monitor: CP-50, Xprezzon, Central Station: Star8800, Xhibit, Pulse Oximeter: Masimo RAD-97
1.	Country of Manufacturer	USA	USA	USA
2.	Country of Origin of Product	USA	Finland & China	USA & China
3.	Compliance with defined Quality Standards	(Non Complied) FDA CFG (Dual certificate for ICU Monitors not attached.	(Complied) CE(MDD), FDA 510K & MHLW	(Complied) CE (MDD) & FDA 510K
4.	Specifications Compliance Feature Wise	1- CMS model not quoted. 2- The Modular ICU Monitor is of 15 inches, while the requirement was of 17". 3- In sub item 10 (iv) Cardiac Monitoring for (8 beds) is quoted instead of 30 beds. 4- FDA 510k certificate is issued without Arrhythmia Detection or Alarms)	Compliance	Compliance
5.	Technical Eligibility of Product	Not Eligible	Eligible	Eligible
6.	Technical Eligibility of Firm	Not Eligible	Eligible	Eligible
Technical Status:		NON RESPONSIVE	RESPONSIVE	RESPONSIVE



Item No. 11  
Item Name: Defibrillator, Defibrillator with external Pacing Marker and Defibrillator with paddles (internal & external) Paeds and Adults  
Tender Name: Procurement Of Bio-Medical Equipment Under PSDP Project "Establishment of 50 Bedded Cardiac Hospital, Gilgit".

PART – I KNOCKED DOWN CRITERIA (BID EVALUATION & VENDOR EVALUATION)				
Sr. #	Evaluation Parameters	M/s Radiant Medical Pvt Ltd.	M/s Shirazi Trading Pvt. Ltd.	M/s Jamil Enterprises
1.	Bid Form	Yes	Yes	Yes
2.	Bidder's Information Form	Yes	Yes	Yes
3.	Documentary proof active of National Tax Number (NTN) and General Sales Tax (GST) Number and Registration Professional Tax Certificate.	Yes	Yes	Yes
4.	Original Bid Security (In form of CDR) & Form	Yes	Yes	Yes
5.	Bid validity time is as per demand	Yes	Yes	Yes
6.	Warranty Form as per Tender	No (Conditional offer)	Yes	Yes
7.	Delivery as per Schedule of Requirement	Yes	Yes	Yes
8.	Undertaking on legal valid and attested stamp paper Rs. 100/- that the firm is not blacklisted by any of Provincial or Federal Government Department, Agency, Organization or autonomous body or Private Sector Organization anywhere in Pakistan.	Yes	Yes	Yes
9.	Embassy Attested Exclusively Authorized Agent / Sole Distributor Certificate by the Original Equipment Manufacturer (OEM).	No (Baxen Cardio is not an OEM) Statement regarding declaration of OEM varies as per Bid Documents.	No (Embassy Attestation is not clear)	Yes (For M/s Medical Equipment and System)
10.	Sub-Distributor Certificate from Sole Distributor having consent of OEM. (Where Applicable)	N/A	N/A	Yes
11.	Proof for having minimum 5 years of experience in supply, installation, commissioning of biomedical equipment.	Yes	Yes	Yes

12.	At least 05 verifiable Past Performance Certificates by the Procuring Agency of quoted products issued against the same Supply Orders.	Yes	No	Yes
13.	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes	Yes	Yes
14.	Originally signed documentary evidence from the Manufacturer that the quoted brand is the Original Equipment Manufacturer (OEM) with indication of manufacturing site and its location.	Yes	Yes	Yes
15.	Up to Date After Sales Service Certificate from OEM/Principal	Yes	Yes	Yes
16.	Availability of relevant Tools and Testing / Calibration Equipment.	Yes	Yes	Yes
17.	A Certificate from the OEM that the installation will be conducted in conformity with the system requirements.	Yes	Yes	Yes
18.	Consortiums or Joint Ventures (JVs) (Where applicable)	N/A	N/A	N/A
19.	Complete Package / Tender	Yes	Yes	Yes
Remarks:		Not Eligible	Not Eligible	Eligible
<b>PART – II</b> <b>KNOCKED DOWN CRITERIA</b> <b>(PRODUCT EVALUATION)</b>				
Item Ad. No.	Defibrillator, Defibrillator with external Pacing Marker and Defibrillator with paddles (internal & external) Paeds and Adults	M/s Radiant Medical Pvt Ltd.	M/s Shirazi Trading Pvt. Ltd.	M/s Jamil Enterprises
	Brand	Baxen Cardio	Progetti Medical	Zoll
11	Quoted Model	Reanibex 800	Rescue Life	M2
1.	Country of Manufacturer	Spain	Italy	USA
2.	Country of Origin of Product	Spain	Italy	China
3.	Compliance with defined Quality Standards	(Non complied) CE (MDD) on the name of Osatu Group S. Coop	(Compliance) CE (MDD)	(Compliance) CE (MDD) & FDA CFG
4.	Specifications Compliance Feature Wise	(Non complied)	(Non complied)	Compliance



		<p>1- Reusable internal paddles by OEM cannot be verified from product data sheet.</p> <p>2- Quoted only one set of Internal Paddles instead of both which was required in each.</p>	<p>5- Disposable defibrillation pads and internal defibrillation paddles instead of reusable internal paddles.</p> <p>6- AED mode has fixed energy at 200 J.</p> <p>7- Pacing / Defibrillation Pads (Reusable / Disposables) with packing marker (Adult &amp; Paediatric not quoted).</p>	
5.	Technical Eligibility of Product	Not Eligible	Not Eligible	Eligible
6.	Technical Eligibility of Firm	Not Eligible	Not Eligible	Eligible
<b>Technical Status:</b>		<b>NON RESPONSIVE</b>	<b>NON RESPONSIVE</b>	<b>RESPONSIVE</b>

Item No. 12

Item Name: ECG Machine 3 channel and ECG Machine Multi Channel (12)

Tender Name: Procurement Of Bio-Medical Equipment Under PSDP Project "Establishment of 50 Bedded Cardiac Hospital, Gilgit".

PART – I KNOCKED DOWN CRITERIA (BID EVALUATION & VENDOR EVALUATION)					
Sr. #	Evaluation Parameters	M/s Radiant Medical Pvt Ltd.	M/s Jamil Enterprises	M/s Sahaab International Pvt. Ltd.	M/s Hospicare Systems
1.	Bid Form	Yes	Yes	Yes	Yes
2.	Bidder's Information Form	Yes	Yes	Yes	Yes
3.	Documentary proof active of National Tax Number (NTN) and General Sales Tax (GST) Number and Registration Professional Tax Certificate.	Yes	Yes	Yes	Yes
4.	Original Bid Security (In form of CDR) & Form	Yes	Yes	Yes	No
5.	Bid validity time is as per demand	Yes	Yes	Yes	Yes
6.	Warranty Form as per Tender	Yes	Yes	Yes	Yes
7.	Delivery as per Schedule of Requirement	Yes	Yes	Yes	Yes
8.	Undertaking on legal valid and attested stamp paper Rs. 100/- that the firm is not blacklisted by any of Provincial or Federal Government Department, Agency, Organization or autonomous body or Private Sector Organization anywhere in Pakistan.	Yes	Yes	No	Yes
9.	Embassy Attested Exclusively Authorized Agent / Sole Distributor Certificate by the Original Equipment Manufacturer (OEM).	Yes	Yes (For M/s Biotech Services)	No (The attached Certificate is not Embassy Attested.)	Yes
10.	Sub-Distributor Certificate from Sole Distributor having consent of OEM. (Where Applicable)	N/A	Yes	N/A	N/A



11.	Proof for having minimum 5 years of experience in supply, installation, commissioning of biomedical equipment.	Yes	Yes	Yes	Yes
12.	At least 05 verifiable Past Performance Certificates by the Procuring Agency of quoted products issued against the same Supply Orders.	Yes	No	No (Unverifiable performance certificates)	Yes
13.	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes	Yes	Yes	Yes
14.	Originally signed documentary evidence from the Manufacturer that the quoted brand is the Original Equipment Manufacturer (OEM) with indication of manufacturing site and its location.	Yes	Yes	Yes	Yes
15.	Up to Date After Sales Service Certificate from OEM/Principal	Yes	Yes	Yes	Yes
16.	Availability of relevant Tools and Testing / Calibration Equipment.	Yes	Yes	No	Yes
17.	A Certificate from the OEM that the installation will be conducted in conformity with the system requirements.	Yes	Yes	Yes	Yes
18.	Consortiums or Joint Ventures (JVs) (Where applicable)	N/A	N/A	N/A	N/A
19.	Complete Package / Tender	Yes	No	No	No
Remarks:		Eligible	Not Eligible	Not Eligible	Not Eligible
<b>PART – II</b> <b>KNOCKED DOWN CRITERIA</b> <b>(PRODUCT EVALUATION)</b>					
Item Ad. No.	ECG Machine 3 channel and ECG Machine Multi Channel (12)	M/s Radiant Medical Pvt Ltd.	M/s Jamil Enterprises	M/s Sahaab International Pvt. Ltd.	M/s Hospicare Systems
	Brand	Baxter (Hillrom)	Innomed Medical	Medallion Equipment	Cardioline
12	Quoted Model	ELI-230 & CP-150	Heart Screen 80G-L1 & Heart Screen 112C-1	ECG-M3 & ECG-M12	ECG 100S & ECG 100S

1.	Country of Manufacturer	USA	Hungary	UK	Italy
2.	Country of Origin of Product	USA	Hungary	Not Mentioned	Italy
3.	Compliance with defined Quality Standards	(Compliance) CE MDD & FDA10K	(Compliance) CE (MDD)	(Non Complied) No (Certificate could not verified online) and (CE (MDD) from authorized EU body not available.)	(Compliance) CE (MDD)
4.	Specifications Compliance Feature Wise	Compliance	(Non Complied) 1- Interpretation Software not available for 3-channel. 2- Overall battery backup with 30 mins instead of 60 mins, which also include printing. 3- Accessories not quoted in Technical Offer with both systems. 4- Compatibility with PACS not	(Non Complied) 1- The quoted model is not available on official website of quoted manufacturer.	(Non Complied) 1- Manual & Automatic Operation not available. 2- Provision of battery charging slot in case of mobile / ambulance usage not offered in 3- and 12- channel ECG. 3- Compatibility with PACS not available.



			available.		
5.	Technical Eligibility of Product	Eligible	Not Eligible	Not Eligible	Not Eligible
6.	Technical Eligibility of Firm	Eligible	Not Eligible	Not Eligible	Not Eligible
<b>Technical Status:</b>		<b>RESPONSIVE</b>	<b>NON RESPONSIVE</b>	<b>NON RESPONSIVE</b>	<b>NON RESPONSIVE</b>

Handwritten signatures and initials in blue ink, including a circled 'A', a checkmark, a stylized 'S', a large signature, a signature with '11/9/84', a signature with '12/10', and several other scribbles and initials.

Item No. 13  
Item Name: Signal Average ECG System  
Tender Name: Procurement of Bio-Medical Equipment Under PSDP Project  
"Establishment of 50 Bedded Cardiac Hospital, Gilgit".

PART – I KNOCKED DOWN CRITERIA (BID EVALUATION & VENDOR EVALUATION)			
Sr. #	Evaluation Parameters	M/s Radiant Medical Pvt. Ltd.	M/s Jamil Enterprises
1.	Bid Form	Yes	Yes
2.	Bidder's Information Form	Yes	Yes
3.	Documentary proof active of National Tax Number (NTN) and General Sales Tax (GST) Number and Registration Professional Tax Certificate.	Yes	Yes
4.	Original Bid Security (In form of CDR) & Form	Yes	Yes
5.	Bid validity time is as per demand	Yes	Yes
6.	Warranty Form as per Tender	Yes	Yes
7.	Delivery as per Schedule of Requirement	Yes	Yes
8.	Undertaking on legal valid and attested stamp paper Rs. 100/- that the firm is not blacklisted by any of Provincial or Federal Government Department, Agency, Organization or autonomous body or Private Sector Organization anywhere in Pakistan.	Yes	Yes
9.	Embassy Attested Exclusively Authorized Agent / Sole Distributor Certificate by the Original Equipment Manufacturer (OEM).	Yes	Yes (For M/s Amtronech)
10.	Sub-Distributor Certificate from Sole Distributor having consent of OEM. (Where Applicable)	N/A	Yes
11.	Proof for having minimum 5 years of experience in supply, installation, commissioning of biomedical equipment.	Yes	Yes
12.	At least 05 verifiable Past Performance Certificates by the Procuring Agency of quoted products issued against the same Supply Orders.	Yes	Yes
13.	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes	Yes
14.	Originally signed documentary evidence from the Manufacturer that the quoted brand is the Original Equipment Manufacturer (OEM) with indication of manufacturing site and its location.	No	Yes
15.	Up to Date After Sales Service Certificate from OEM/Principal	Yes	Yes
16.	Availability of relevant Tools and Testing / Calibration Equipment.	Yes	Yes

*[Handwritten signatures and marks at the bottom of the page]*



17.	A Certificate from the OEM that the installation will be conducted in conformity with the system requirements.	Yes	Yes
18.	Consortiums or Joint Ventures (JVs) (Where applicable)	N/A	N/A
19.	Complete Package / Tender	Yes	Yes
Remarks:		Not Eligible	Eligible
<b>PART – II</b> <b>KNOCKED DOWN CRITERIA</b> <b>(PRODUCT EVALUATION)</b>			
Item Ad. No.	Signal Average ECG System	M/s Radiant Medical Pvt. Ltd.	M/s Jamil Enterprises
	Brand	Baxter	Schiller AG
13	Quoted Model	ELI-380	Cardiovit AT-180.1
1.	Country of Manufacturer	USA	Switzerland
2.	Country of Origin of Product	USA	Switzerland
3.	Compliance with defined Quality Standards	(Compliance) FDA 510K & CE (MDD)	(Compliance) CE (MDR)
4.	Specifications Compliance Feature Wise	(Non Complied) Baxter ELI-380 is a 12-lead resting ECG system, while the "Signal Average" ECG System was demanded. Hence, it does not fulfill the criteria of specifications.	Compliance
5.	Technical Eligibility of Product	Not Eligible	Eligible
6.	Technical Eligibility of Firm	Not Eligible	Eligible
Technical Status:		<b>NON RESPONSIVE</b>	<b>RESPONSIVE</b>

Item No. 14  
Item Name: Head Up Tilt (HUTT) System  
Tender Name: Procurement of Bio-Medical Equipment Under PSDP Project  
"Establishment of 50 Bedded Cardiac Hospital, Gilgit".

PART – I KNOCKED DOWN CRITERIA (BID EVALUATION & VENDOR EVALUATION)			
Sr. #	Evaluation Parameters	M/s Mediland Pakistan	M/s Jamil Enterprises
1.	Bid Form	Yes	Yes
2.	Bidder's Information Form	Yes	Yes
3.	Documentary proof active of National Tax Number (NTN) and General Sales Tax (GST) Number and Registration Professional Tax Certificate.	Yes	Yes
4.	Original Bid Security (In form of CDR) & Form	Yes	Yes
5.	Bid validity time is as per demand	Yes	Yes
6.	Warranty Form as per Tender	Yes	Yes
7.	Delivery as per Schedule of Requirement	Yes	Yes
8.	Undertaking on legal valid and attested stamp paper Rs. 100/- that the firm is not blacklisted by any of Provincial or Federal Government Department, Agency, Organization or autonomous body or Private Sector Organization anywhere in Pakistan.	Yes	Yes
9.	Embassy Attested Exclusively Authorized Agent / Sole Distributor Certificate by the Original Equipment Manufacturer (OEM).	No (Embassy Attested Certificate not provided.)	Yes (For M/s Amtronech)
10.	Sub-Distributor Certificate from Sole Distributor having consent of OEM. (Where Applicable)	N/A	Yes
11.	Proof for having minimum 5 years of experience in supply, installation, commissioning of biomedical equipment.	Yes	Yes
12.	At least 05 verifiable Past Performance Certificates by the Procuring Agency of quoted products issued against the same Supply Orders.	No	Yes
13.	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes	Yes
14.	Originally signed documentary evidence from the Manufacturer that the quoted brand is the Original Equipment Manufacturer (OEM) with indication of manufacturing site and its location.	Yes	Yes
15.	Up to Date After Sales Service Certificate from OEM/Principal	Yes	Yes
16.	Availability of relevant Tools and Testing / Calibration Equipment.	Yes	Yes



17.	A Certificate from the OEM that the installation will be conducted in conformity with the system requirements.	Yes	Yes
18.	Consortiums or Joint Ventures (JVs) (Where applicable)	N/A	N/A
19.	Complete Package / Tender	No	Yes
Remarks:		Not Eligible	Eligible
<b>PART – II</b> <b>KNOCKED DOWN CRITERIA</b> <b>(PRODUCT EVALUATION)</b>			
Item Ad. No.	Head Up Tilt (HUTT) System	M/s Mediland Pakistan	M/s Jamil Enterprises
	Brand	Finapres Medical System (FMS)	CN Systems
14	Quoted Model	Nova Plus	Task Force Touch Cardio
1.	Country of Manufacturer	Netherland	Austria
2.	Country of Origin of Product	Netherland & Germany	Austria
3.	Compliance with defined Quality Standards	(Non Complied) Not Attached	(Complied) CE (MDR)
4.	Specifications Compliance Feature Wise	(Non Complied) 1. The quoted model is not available on official website of manufacturer. 2. The country of origin and country of manufacturer is not same.	Compliance
5.	Technical Eligibility of Product	Not Eligible	Eligible
6.	Technical Eligibility of Firm	Not Eligible	Eligible
Technical Status:		NON RESPONSIVE	RESPONSIVE

Item No. 15-19  
Item Name: Broncho Scope, Video Laryngoscope, Laryngoscope, Colposcope and Laparoscope  
Tender Name: Procurement of Bio-Medical Equipment Under PSDP Project "Establishment of 50 Bedded Cardiac Hospital, Gilgit".

PART – I KNOCKED DOWN CRITERIA (BID EVALUATION & VENDOR EVALUATION)			
Sr. #	Evaluation Parameters	M/s Medifa Enterprises	M/s All Med Solutions
1.	Bid Form	As per GRC meeting dated 15-10-2024 held Health Secretariat GB, it was decided that firms shall quote items (15-19) in a single package, while M/s Medifa Enterprises quoted only Item No. 18 (Colposcope), which does not fulfill the pre-requisites.	Yes
2.	Bidder's Information Form	N/A	Yes
3.	Documentary proof active of National Tax Number (NTN) and General Sales Tax (GST) Number and Registration Professional Tax Certificate.	N/A	Yes
4.	Original Bid Security (In form of CDR) & Form	N/A	No
5.	Bid validity time is as per demand	N/A	Yes
6.	Warranty Form as per Tender	N/A	Yes
7.	Delivery as per Schedule of Requirement	N/A	Yes
8.	Undertaking on legal valid and attested stamp paper Rs. 100/- that the firm is not blacklisted by any of Provincial or Federal Government Department, Agency, Organization or autonomous body or Private Sector Organization anywhere in Pakistan.	N/A	Yes
9.	Embassy Attested Exclusively Authorized Agent / Sole Distributor Certificate by the Original Equipment Manufacturer (OEM).	N/A	Yes
10.	Sub-Distributor Certificate from Sole Distributor having consent of OEM. (Where Applicable)	N/A	N/A
11.	Proof for having minimum 5 years of experience in supply, installation, commissioning of biomedical equipment.	N/A	Yes




12.	At least 05 verifiable Past Performance Certificates by the Procuring Agency of quoted products issued against the same Supply Orders.	N/A	Yes
13.	Technical & Engineering capability (As defined for the specific Bid in specifications)	N/A	Yes
14.	Originally signed documentary evidence from the Manufacturer that the quoted brand is the Original Equipment Manufacturer (OEM) with indication of manufacturing site and its location.	N/A	Yes
15.	Up to Date After Sales Service Certificate from OEM/Principal	N/A	Yes
16.	Availability of relevant Tools and Testing / Calibration Equipment.	N/A	Yes
17.	A Certificate from the OEM that the installation will be conducted in conformity with the system requirements.	N/A	Yes
18.	Consortiums or Joint Ventures (JVs) (Where applicable)	N/A	Yes
19.	Complete Package / Tender	No (In-complete offer, as only quoted Item No. 15.)	No
Remarks:		Not Eligible	Not Eligible

**PART – II  
KNOCKED DOWN CRITERIA  
(PRODUCT EVALUATION)**

Item Ad. No.	Broncho Scope, Video Laryngoscope, Laryngoscope, Colposcope and Laparoscope	M/s Medifa Enterprises	M/s All Med Solutions
	Brand	N/A	Karl Storz
15-19	Quoted Model	N/A	Broncho Scope: 11910D, Monitor: TM220, Telecam C3, Video Laryngoscope and Laryngoscope: C-MAC, Colposcope: Vitcom Hopkins, Floor Stand: UG- 800&UG801, LED Nova 150, Laparoscope: Hopkins, Camera: IMAGE1 S Connect, Light Source: Cold Light Fountain Power LED 175

			SCB, Electronic CO <sub>2</sub> Insufflator: Endoflator 40 SCB, Smoke Evaluator System: S-PILOT (UP501)
1.	Country of Manufacturer	N/A	Germany
2.	Country of Origin of Product	N/A	Germany
3.	Compliance with defined Quality Standards	N/A	(Compliance) CE (MDD) & FDA CFG
4.	Specifications Compliance Feature Wise	N/A	Compliance
5.	Technical Eligibility of Product	N/A	Eligible
6.	Technical Eligibility of Firm	N/A	Not Eligible
<b>Technical Status:</b>		<b>NON RESPONSIVE</b>	<b>NON RESPONSIVE</b>





Item No. 20  
Item Name: Modular Operation Theatre  
Tender Name: Procurement of Bio-Medical Equipment Under PSDP Project  
"Establishment of 50 Bedded Cardiac Hospital, Gilgit".

PART – I KNOCKED DOWN CRITERIA (BID EVALUATION & VENDOR EVALUATION)			
Sr. #	Evaluation Parameters	M/s Radiant Medical Pvt. Ltd.	M/s Total Technologies Pvt. Ltd.
1.	Bid Form	Yes	Yes
2.	Bidder's Information Form	Yes	Yes
3.	Documentary proof active of National Tax Number (NTN) and General Sales Tax (GST) Number and Registration Professional Tax Certificate.	Yes	Yes
4.	Original Bid Security (In form of CDR) & Form	Yes	Yes
5.	Bid validity time is as per demand	Yes	Yes
6.	Warranty Form as per Tender	Yes	Yes
7.	Delivery as per Schedule of Requirement	Yes	Yes
8.	Undertaking on legal valid and attested stamp paper Rs. 100/- that the firm is not blacklisted by any of Provincial or Federal Government Department, Agency, Organization or autonomous body or Private Sector Organization anywhere in Pakistan.	Yes	Yes
9.	Embassy Attested Exclusively Authorized Agent / Sole Distributor Certificate by the Original Equipment Manufacturer (OEM).	Yes	Yes
10.	Sub-Distributor Certificate from Sole Distributor having consent of OEM. (Where Applicable)	N/A	N/A
11.	Proof for having minimum 5 years of experience in supply, installation, commissioning of biomedical equipment.	Yes	Yes
12.	At least 05 verifiable Past Performance Certificates by the Procuring Agency of quoted products issued against the same Supply Orders.	Yes	Yes
13.	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes	Yes
14.	Originally signed documentary evidence from the Manufacturer that the quoted brand is the Original Equipment Manufacturer (OEM) with indication of manufacturing site and its location.	Yes	Yes
15.	Up to Date After Sales Service Certificate from OEM/Principal	Yes	Yes
16.	Availability of relevant Tools and Testing / Calibration Equipment.	Yes	Yes
17.	A Certificate from the OEM that the installation will be conducted in conformity with the system requirements.	Yes	Yes
18.	Consortiums or Joint Ventures (JVs) (Where applicable)	N/A	N/A
19.	Complete Package / Tender	Yes	Yes

		Remarks:	Eligible	Eligible
<b>PART – II</b>				
<b>KNOCKED DOWN CRITERIA</b>				
<b>(PRODUCT EVALUATION)</b>				
Item Ad. No.	Modular Operation Theatre	M/s Radiant Medical Pvt. Ltd.	M/s Total Technologies Pvt. Ltd.	
	Brand	HT Group	Operamed	
<b>20</b>	Quoted Model	As per compliance Sheet	As per compliance Sheet	
<b>1.</b>	Country of Manufacturer			
<b>2.</b>	Country of Origin of Product			
<b>3.</b>	Compliance with defined Quality Standards	(Compliance) ISO and other relevant standards	(Compliance) ISO and other relevant standards	
<b>4.</b>	Specifications Compliance Feature Wise	Compliance	Compliance	
<b>5.</b>	Technical Eligibility of Product	Eligible	Eligible	
<b>6.</b>	Technical Eligibility of Firm	Eligible	Eligible	
<b>Technical Status:</b>		<b>RESPONSIVE</b>	<b>RESPONSIVE</b>	

<b>COMPLIANCE SHEET</b>			
No.	Parameters	M/s Radiant Medical Pvt. Ltd.	M/s Total Technologies Pvt. Ltd.
1	Wall Paneling System	Model: SSP01 HT Group Germany	Model: Teue Space Wall Panels Operamed Italy
2	Air Tight Clean Room Suspended Ceiling System	Model: GSC-01 HT Group Germany	Model: Teue Space Wall Panels Operamed Italy
3	Peripheral Lights for Operating Room	Model: Illuminator C-L HT Group Germany	Model: True LED LC Lighting Operamed Italy
4	X-Ray Viewer for Operating Room	Model: BF LED Weiko Medical Italy (Supplied by HT group)	Model: BF LED Weiko Medical Italy
5	Clean Room Flooring Air Tight Automatic Sliding Door For	Model: Static Deceptive (SD Type) Forbo Netherland	Model: Sphera EC/SD Forbo Netherland
6	Operating Rooms with Electronic Privacy Glass	Model: ASD – AD HT Group Germany	Model: True Space Doors Operamed Italy
7	Medical Cabinet for Operating Room	Model: SSC -C HT Group Germany	Model: True Space Stainless Steel Cabinet Operamed Italy
8	Ceiling Mounted Anesthesia Pendant	Model: TRUPOINT 5000 Baxter Medical Germany	Model: OKI ISOLA LM Medical Davison Italy
9	Ceiling Mounted Surgical Pendant	Model: TRUPOINT 7000 Baxter Medical Germany	Model: OKI ISOLA LM Medical Davison Italy



10	<b>Double Dome LED Operating Light with HD Camera</b>	Model: EPURE (Double Dome) (Wireless Camera Ready) Epure USA	Model: Mach LED 6MC/6MC Dr Mach Germany
11	<b>Laminar Air Flow Ceiling System</b>	Model: Type F A21 France	Model: LF-ZP 2.8*2.7M Ecco Line Marcchart Austria
12	<b>Dedicated Hygienic Air Handling Unit</b>	Model: H8000 (Direct Expansion – H Line) A21 France	Operamed Italy
13	<b>Surgeon Control Panel</b>	Model: TCP-22 DAGRI Slovenia	Model: Operamed Connect – Touch Screen Control Panel Operamed Italy
14	<b>Digital Integration System for Audio Video Conferencing</b>	Quoted	Quoted
15	<b>Online Sinewave imported UPS of at least 8 KVA of operational back-up of at least 30 minutes</b>	Model: APC / equivalent USA / equivalent	Model: 9PX 800VA / equivalent Eaton Italy
16	<b>Isolated Power Supply 10 kVA</b>	Model: ES 710/8000 Bender Germany	Model: I-Panel (10Kva) LM Medical Italy
17	<b>Stainless Steel AISI 304 or better Double Door Pass-Through Cabinet towards dirty corridor</b>	HT Group Germany	Operamed Italy
18	<b>2 Bay Surgical Scrub (Automatic Stainless Steel Scrub System with manual provision)</b>	Model: Type B-2 HT Group Germany	Model: 8221 Operamed From Bawer Italy

Item No. 21  
Item Name: Heart Lung Machine  
Tender Name: Procurement of Bio-Medical Equipment Under PSDP Project  
"Establishment of 50 Bedded Cardiac Hospital, Gilgit".

PART – I KNOCKED DOWN CRITERIA (BID EVALUATION & VENDOR EVALUATION)			
Sr. #	Evaluation Parameters	M/s Vertex Medical Pvt. Ltd.	M/s Mediland Pakistan Pvt. Ltd.
1.	Bid Form	Yes	Yes
2.	Bidder's Information Form	Yes	Yes
3.	Documentary proof active of National Tax Number (NTN) and General Sales Tax (GST) Number and Registration Professional Tax Certificate.	Yes	Yes
4.	Original Bid Security (In form of CDR) & Form	Yes	Yes
5.	Bid validity time is as per demand	Yes	Yes
6.	Warranty Form as per Tender	Yes	Yes
7.	Delivery as per Schedule of Requirement	Yes	Yes
8.	Undertaking on legal valid and attested stamp paper Rs. 100/- that the firm is not blacklisted by any of Provincial or Federal Government Department, Agency, Organization or autonomous body or Private Sector Organization anywhere in Pakistan.	Yes	Yes
9.	Embassy Attested Exclusively Authorized Agent / Sole Distributor Certificate by the Original Equipment Manufacturer (OEM).	Yes	Yes
10.	Sub-Distributor Certificate from Sole Distributor having consent of OEM. (Where Applicable)	N/A	N/A
11.	Proof for having minimum 5 years of experience in supply, installation, commissioning of biomedical equipment.	Yes	Yes
12.	At least 05 verifiable Past Performance Certificates by the Procuring Agency of quoted products issued against the same Supply Orders.	Yes	Yes
13.	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes	Yes
14.	Originally signed documentary evidence from the Manufacturer that the quoted brand is the Original Equipment Manufacturer (OEM) with indication of manufacturing site and its location.	Yes	Yes
15.	Up to Date After Sales Service Certificate from OEM/Principal	Yes	Yes
16.	Availability of relevant Tools and Testing / Calibration Equipment.	Yes	Yes
17.	A Certificate from the OEM that the installation will be conducted in conformity with the system requirements.	Yes	Yes
18.	Consortiums or Joint Ventures (JVs) (Where applicable)	N/A	N/A



19.	Complete Package / Tender	Yes	Yes
Remarks:		Eligible	Eligible
PART – II KNOCKED DOWN CRITERIA (PRODUCT EVALUATION)			
Item Ad. No.	Echocardiography Machines	M/s Vertex Medical Pvt. Ltd.	M/s Mediland Pakistan Pvt. Ltd.
	Brand	Liva Nova	Getinge
21	Quoted Model	S5	HL-40
1.	Country of Manufacturer	Germany	Germany
2.	Country of Origin of Product	Germany	Germany/Sweden
3.	Compliance with defined Quality Standards	Compliance CE MDD & FDA10K	Compliance MHLW & FDA10K
4.	Specifications Compliance Feature Wise	Compliance	Compliance
5.	Technical Eligibility of Product	Eligible	Eligible
6.	Technical Eligibility of Firm	Eligible	Eligible
Technical Status:		RESPONSIVE	RESPONSIVE

Item No. 22  
Item Name: Intra-Aortic Balloon Pump with All Accessories  
Tender Name: Procurement of Bio-Medical Equipment Under PSDP Project  
"Establishment of 50 Bedded Cardiac Hospital, Gilgit".

PART – I KNOCKED DOWN CRITERIA (BID EVALUATION & VENDOR EVALUATION)			
Sr. #	Evaluation Parameters	M/s Shirazi Trading Pvt. Ltd.	M/s Mediland Pakistan (Pvt) Ltd.
1.	Bid Form	Yes	Yes
2.	Bidder's Information Form	Yes	Yes
3.	Documentary proof active of National Tax Number (NTN) and General Sales Tax (GST) Number and Registration Professional Tax Certificate.	Yes	Yes
4.	Original Bid Security (In form of CDR) & Form	Yes	Yes
5.	Bid validity time is as per demand	Yes	Yes
6.	Warranty Form as per Tender	Yes	Yes
7.	Delivery as per Schedule of Requirement	Yes	Yes
8.	Undertaking on legal valid and attested stamp paper Rs. 100/- that the firm is not blacklisted by any of Provincial or Federal Government Department, Agency, Organization or autonomous body or Private Sector Organization anywhere in Pakistan.	Yes	Yes
9.	Embassy Attested Exclusively Authorized Agent / Sole Distributor Certificate by the Original Equipment Manufacturer (OEM).	Yes	Yes
10.	Sub-Distributor Certificate from Sole Distributor having consent of OEM. (Where Applicable)	N/A	N/A
11.	Proof for having minimum 5 years of experience in supply, installation, commissioning of biomedical equipment.	Yes	Yes
12.	At least 05 verifiable Past Performance Certificates by the Procuring Agency of quoted products issued against the same Supply Orders.	Yes	Yes
13.	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes	Yes
14.	Originally signed documentary evidence from the Manufacturer that the quoted brand is the Original Equipment Manufacturer (OEM) with indication of manufacturing site and its location.	Yes	Yes
15.	Up to Date After Sales Service Certificate from OEM/Principal	Yes	Yes
16.	Availability of relevant Tools and Testing / Calibration Equipment.	Yes	Yes
17.	A Certificate from the OEM that the installation will be conducted in conformity with the system requirements.	Yes	Yes
18.	Consortiums or Joint Ventures (JVs) (Where applicable)	N/A	N/A



19.	Complete Package / Tender	Yes	Yes
	Remarks:	Eligible	Eligible
PART – II KNOCKED DOWN CRITERIA (PRODUCT EVALUATION)			
Item Ad. No.	Intra-Aortic Balloon Pump with All Accessories	M/s Shirazi Trading Pvt. Ltd.	M/s Mediland Pakistan (Pvt) Ltd.
	Brand	Teleflex	Getinge
22	Quoted Model	Arrow AC3 Optimus	Cardio Save Hybrid
1.	Country of Manufacturer	USA	Germany
2.	Country of Origin of Product	USA	Germany
3.	Compliance with defined Quality Standards	(Compliance) CE MDD & FDA10K	(Non Complied) CE MDD & FDA10K (With Class-I recalls)
4.	Specifications Compliance Feature Wise	Compliance	Non Complied
5.	Technical Eligibility of Product	Eligible	Not Eligible
6.	Technical Eligibility of Firm	Eligible	Eligible
Technical Status:		RESPONSIVE	NON RESPONSIVE

Item No. 23

Item Name: Hypothermia Machine

Tender Name: Procurement Of Bio-Medical Equipment Under PSDP Project "Establishment of 50 Bedded Cardiac Hospital, Gilgit".

PART - I KNOCKED DOWN CRITERIA (BID EVALUATION & VENDOR EVALUATION)					
Sr. #	Evaluation Parameters	M/s Hospicare Systems	M/s Jamil Enterprises	M/s Mediland Pakistan Pvt. Ltd.	M/s Vertex medical Pvt. Ltd.
20	Bid Form	Yes	Yes	Yes	Yes
21	Bidder's Information Form	Yes	Yes	Yes	Yes
22	Documentary proof active of National Tax Number (NTN) and General Sales Tax (GST) Number and Registration Professional Tax Certificate.	Yes	Yes	Yes	Yes
23	Original Bid Security (In form of CDR) & Form	No	Yes	Yes	No
24	Bid validity time is as per demand	Yes	Yes	Yes	Yes
25	Warranty Form as per Tender	Yes	Yes	Yes	Yes
26	Delivery as per Schedule of Requirement		Yes	Yes	Yes
27	Undertaking on legal valid and attested stamp paper Rs. 100/- that the firm is not blacklisted by any of Provincial or Federal Government Department, Agency, Organization or autonomous body or Private Sector Organization anywhere in Pakistan.	Yes	Yes	Yes	Yes
28	Embassy Attested Exclusively Authorized Agent / Sole Distributor Certificate by the Original Equipment Manufacturer (OEM).	Yes	No (A certificate from Gentherm was issued to Terumo Middle East FZCO, authorizing the distribution of Hemotherm 400 CE. Terumo Middle East FZCO subsequently	No (Expired Certificate)	Yes

56



			authorized M/s Universal Enterprises, which later issued a sub-distributorship letter to M/s Jamil Enterprises.)		
29	Sub-Distributor Certificate from Sole Distributor having consent of OEM. (Where Applicable)	N/A	N/A	N/A	N/A
30	Proof for having minimum 5 years of experience in supply, installation, commissioning of biomedical equipment.	Yes	Yes	Yes	Yes
31	At least 05 verifiable Past Performance Certificates by the Procuring Agency of quoted products issued against the same Supply Orders.	Yes	Yes	Yes	Yes
32	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes	Yes	Yes	Yes
33	Originally signed documentary evidence from the Manufacturer that the quoted brand is the Original Equipment Manufacturer (OEM) with indication of manufacturing site and its location.	Yes	Yes	Yes	Yes
34	Up to Date After Sales Service Certificate from OEM/Principal	Yes	Yes	Yes	Yes
35	Availability of relevant Tools and Testing / Calibration Equipment.	Yes	Yes	Yes	Yes
36	A Certificate from the OEM that the installation will be conducted in conformity with the system requirements.	Yes	Yes	Yes	Yes
37	Consortiums or Joint Ventures (JVs) (Where applicable)	N/A	N/A	N/A	N/A
38	Complete Package / Tender	Yes	No	No	Yes
Remarks:		Not Eligible	Not Eligible	Not Eligible	Eligible

PART – II KNOCKED DOWN CRITERIA (PRODUCT EVALUATION)					
Item Ad. No.	Hypothermia Machine	M/s Hospicare Systems	M/s Jamil Enterprises	M/s Mediland Pakistan Pvt. Ltd.	M/s Vertex medical Pvt. Ltd.
	Brand	Gentherm	Terumo Middle East FZCO	Getinge	Liva Nova
23	Quoted Model	Hemotherm 400 CE	Hemotherm 400 CE	HCU 40	Heater Cooler System 3T
7.	Country of Manufacturer	USA	USA	Germany	Germany
8.	Country of Origin of Product	USA	UAE	Germany	Germany
9.	Compliance with defined Quality Standards	(Compliance) FDA 510K	(Compliance) FDA 510K	(Compliance) FDA 510K & CE (MDD)	(Compliance) FDA 510K & CE (MDD)
10	Specifications Compliance Feature Wise	Compliance	(Non Complied)  The country of origin is from UAE as Terumo Middle East FZCO, who is distributor of Gnetherm products in UAE.	Compliance	Compliance
11	Technical Eligibility of Product	Eligible	Not Eligible	Eligible	Eligible
12	Technical Eligibility of Firm	Not Eligible	Not Eligible	Not Eligible	Eligible
Technical Status:		NON RESPONSIVE	NON RESPONSIVE	NON RESPONSIVE	RESPONSIVE



Item No. 24

Item Name: Blood Warmer

Tender Name: Procurement Of Bio-Medical Equipment Under PSDP Project "Establishment of 50 Bedded Cardiac Hospital, Gilgit".

PART – I KNOCKED DOWN CRITERIA (BID EVALUATION & VENDOR EVALUATION)				
Sr. #	Evaluation Parameters	M/s Radiant Medical Pvt Ltd.	M/s Hospicare Systems	M/s Irfan Brothers International
1.	Bid Form	Yes	Yes	Bid is not signed and stamped as required in BD Clause ITB 19.
2.	Bidder's Information Form	Yes	Yes	Yes
3.	Documentary proof active of National Tax Number (NTN) and General Sales Tax (GST) Number and Registration Professional Tax Certificate.	Yes	Yes	Yes
4.	Original Bid Security (In form of CDR) & Form	Yes	No (Not Attached with Technical Offer)	Yes
5.	Bid validity time is as per demand	Yes	Yes	Yes
6.	Warranty Form as per Tender	Yes	Yes	Yes
7.	Delivery as per Schedule of Requirement			
8.	Undertaking on legal valid and attested stamp paper Rs. 100/- that the firm is not blacklisted by any of Provincial or Federal Government Department, Agency, Organization or autonomous body or Private Sector Organization anywhere in Pakistan.	Yes	Yes	Yes
9.	Embassy Attested Exclusively Authorized Agent / Sole Distributor Certificate by the Original Equipment Manufacturer (OEM).	Yes	Yes	Exclusively Authorized Agent / Sole Distributor for the quoted item not available and the

				attached Certificate is not Embassy Attested.
10.	Sub-Distributor Certificate from Sole Distributor having consent of OEM. (Where Applicable)	N/A	N/A	Yes
11.	Proof for having minimum 5 years of experience in supply, installation, commissioning of biomedical equipment.	Yes	Yes	Yes
12.	At least 05 verifiable Past Performance Certificates by the Procuring Agency of quoted products issued against the same Supply Orders.	Yes	Yes	No
13.	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes	Yes	Yes
14.	Originally signed documentary evidence from the Manufacturer that the quoted brand is the Original Equipment Manufacturer (OEM) with indication of manufacturing site and its location.	Yes	Yes	No
15.	Up to Date After Sales Service Certificate from OEM/Principal	Yes	Yes	Yes
16.	Availability of relevant Tools and Testing / Calibration Equipment.	Yes	Yes	Yes
17.	A Certificate from the OEM that the installation will be conducted in conformity with the system requirements.	Yes	Yes	Yes
18.	Consortiums or Joint Ventures (JVs) (Where applicable)	N/A	N/A	N/A
19.	Complete Package / Tender	Yes	No	No
Remarks:		Eligible	Not Eligible	Not Eligible
<b>PART – II</b> <b>KNOCKED DOWN CRITERIA</b> <b>(PRODUCT EVALUATION)</b>				
Item Ad. No.	Blood Warmer	M/s Radiant Medical Pvt Ltd.	M/s Hospicare Systems	M/s Irfan Brothers International
	Brand	Barkey	Gentherm Medical	KW Apparecchi Scientifici
24	Quoted Model	PlasmaTherm	Astotherm Plus 220	DPFD 3
1.	Country of Manufacturer	Germany	USA	Italy
2.	Country of Origin of Product	Germany	USA	Italy
3.	Compliance with defined Quality Standards	Compliance	Non Complied	Non Complied



		CE MDD & FDA10K		
4.	Specifications Compliance Feature Wise	Compliance	(Non Complied) 3- Temperature setting: +37°C to +42°C not available instead 39/41/43°C. 4- Minimum three to four bags capacity not available. 5- CE MDD certificate issued on the Stihler Electronics GmbH. (Technically not Acceptable)	(Non Complied) Quoted Plasma Thawer instead of Blood Warmer.
5.	Technical Eligibility of Product	Eligible	Not Eligible	Not Eligible
6.	Technical Eligibility of Firm	Eligible	Eligible	Not Eligible
<b>Technical Status:</b>		<b>RESPONSIVE</b>	<b>NON RESPONSIVE</b>	<b>NON RESPONSIVE</b>

Item No. 26-29

Item Name: Mobile Light, Portable OT Light, OT Light, Operation Theater Lights with Satellite (Ceiling)

Tender Name: Procurement of Bio-Medical Equipment Under PSDP Project "Establishment of 50 Bedded Cardiac Hospital, Gilgit".

PART – I KNOCKED DOWN CRITERIA (BID EVALUATION & VENDOR EVALUATION)						
Sr. #	Evaluation Parameters	M/s Total Technologies Pvt. Ltd.	M/s Jamil Enterprises	M/s Irfan Brothers International	M/s Sahaab International	M/s GMED
1.	Bid Form	Yes	Yes	Yes	Yes	As per BD ITB Clause 5, a bidder is required to submit only one bid for a specific item/package. Additionally, the tender advertisement explicitly stated the requirement to submit separate proposals/bids. This was further recorded in the minutes of the pre-bid meeting and clarified by the GRC, which specified that firms must submit an original proposal (Technical and Financial) along with one photocopy in hard form and one in soft



						form (in USB) for each item/package. Contrary to these requirements, the firm has submitted a single proposal for multiple items (Item Nos. 26-29, 31, 32, 34, 36, 37, and 38).
2.	Bidder's Information Form	Yes	Yes	Yes	Yes	N/A
3.	Documentary proof active of National Tax Number (NTN) and General Sales Tax (GST) Number and Registration Professional Tax Certificate.	Yes	Yes	Yes	Yes	N/A
4.	Original Bid Security (In form of CDR) & Form	Yes	Yes	Yes	No	N/A
5.	Bid validity time is as per demand	Yes	Yes	Yes	Yes	N/A
6.	Warranty Form as per Tender	Yes	Yes	Yes	Yes	N/A
7.	Delivery as per Schedule of Requirement		Yes	Yes	Yes	N/A
8.	Undertaking on legal valid and attested stamp paper Rs. 100/- that the firm is not blacklisted by any of Provincial or Federal Government Department, Agency, Organization or autonomous body or Private Sector Organization anywhere in Pakistan.	Yes	Yes	Yes	No	N/A
9.	Embassy Attested Exclusively Authorized Agent / Sole Distributor Certificate by the Original Equipment Manufacturer (OEM).	Yes	Yes (For M/s Sahar International Pvt. Ltd.)	Yes (For M/s Human Healthcare)	No (Embassy attested certificate not provided)	N/A

10.	Sub-Distributor Certificate from Sole Distributor having consent of OEM. (Where Applicable)	N/A	Yes	Yes	N/A	N/A
11.	Proof for having minimum 5 years of experience in supply, installation, commissioning of biomedical equipment.	Yes	Yes	Yes	Yes	N/A
12.	At least 05 verifiable Past Performance Certificates by the Procuring Agency of quoted products issued against the same Supply Orders.	Yes	Yes	No	No	N/A
13.	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes	Yes	Yes	Yes	N/A
14.	Originally signed documentary evidence from the Manufacturer that the quoted brand is the Original Equipment Manufacturer (OEM) with indication of manufacturing site and its location.	Yes	Yes	Yes	Yes	N/A
15.	Up to Date After Sales Service Certificate from OEM/Principal	Yes	Yes	Yes	Yes	N/A
16.	Availability of relevant Tools and Testing / Calibration Equipment.	Yes	Yes	Yes	Yes	N/A
17.	A Certificate from the OEM that the installation will be conducted in conformity with the system requirements.	Yes	Yes	Yes	Yes	N/A
18.	Consortiums or Joint Ventures (JVs) (Where applicable)	N/A	N/A	N/A	N/A	N/A
19.	Complete Package / Tender	Yes	Yes	No	No Bid Pages are not signed and	No



					stamped as required in BD ITB Clause 19.	
Remarks:		Eligible	Eligible	Not Eligible	Not Eligible	Not Eligible
PART – II KNOCKED DOWN CRITERIA (PRODUCT EVALUATION)						
Item Ad. No.	Mobile Light, Portable OT Light, OT Light, Operation Theater Lights with Satellite (Ceiling)	M/s Total Technologies Pvt. Ltd.	M/s Jamil Enterprises	M/s Irfan Brothers International	M/s Sahaab International	M/s GMED
	Brand	Terumo Corporation	DR MACH	Medicinos Gija	Medallion Equipment	N/A
26-29	Quoted Model	Mobile OT Light: EMALED 300M+, Portable Light: EMALED 300M+, OT Light: EMALED 560/560VM, Operation Theater Lights with Satellite (Ceiling): EMALED 560/560	Mobile OT Light: 150 FP (with battery), Portable Light: 150 FP (with battery), OT Light: LED 300 DF SC with Satellite LED 300 with HD Camera, Operation Theater Lights with Satellite (Ceiling): LED 300 DF SC with Satellite LED 300 DF SC	Mobile OT Light: LUMEN 160 LEM, Portable Light: LUMEN 160 LEM, OT Light: LUMEN C160/160, Operation Theater Lights with Satellite (Ceiling): LUMEN C160/160,	Mobile OT Light: 700M, Portable Light: 700M, OT Light: 700, Operation Theater Lights with Satellite (Ceiling): 700	N/A
1.	Country of Manufacturer	Japan	Germany	Lithuania	UK	N/A
2.	Country of Origin of Product	Japan	Germany	Lithuania	Not Mentioned	N/A

3.	Compliance with defined Quality Standards	(Compliance) CE (MDD)	(Compliance) CE (MDD)	(Non Complied)  (CE (MDD) from authorized EU body not available.)	(Non Complied) (Certificate could not have verified online) and (CE (MDD) from authorized EU body not available.) (Technically not Acceptable.)	N/A
4.	Specifications Compliance Feature Wise	Compliance	(Non Complied)  1. The quoted model for OT light Item No. 28 (LED 300 DF SC) has camera preparation instead of integrated camera and also does not contain monitor carried on standard axis.  2. The video monitor model not quoted.	(Non Complied)  1. The quoted model 160 LEM is not available on official website of quoted manufacturer.  2. (CE (MDD) from authorized EU body not available.) (Technically not Acceptable)	(Non Complied)  4. The required specifications were copied on product catalog attached to the technical offer; moreover, the product information also could not be verified from the	N/A



					<p>manufacturer's website.</p> <p>5. The quoted model is not available on official website of quoted manufacturer and the required specifications were copied on product catalog attached to the technical offer.</p>	
5.	Technical Eligibility of Product	Eligible	Not Eligible	Not Eligible	Not Eligible	N/A
6.	Technical Eligibility of Firm	Eligible	Eligible	Not Eligible	Not Eligible	Not Eligible
<b>Technical Status:</b>		<b>RESPONSIVE</b>	<b>NON RESPONSIVE</b>	<b>NON RESPONSIVE</b>	<b>NON RESPONSIVE</b>	<b>NON RESPONSIVE</b>

Item No.  
Item Name:  
Tender Name:

31  
Operation Table  
Procurement of Bio-Medical Equipment Under PSDP Project "Establishment of 50 Bedded Cardiac Hospital, Gilgit".

PART - I KNOCKED DOWN CRITERIA (BID EVALUATION & VENDOR EVALUATION)				
Sr. #	Evaluation Parameters	M/s GMED	M/s Irfan Brothers International	M/s Sahaab International
1.	Bid Form	As per BD ITB Clause 5, a bidder is required to submit only one bid for a specific item/package. Additionally, the tender advertisement explicitly stated the requirement to submit separate proposals/bids. This was further recorded in the minutes of the pre-bid meeting and clarified by the GRC, which specified that firms must submit an original proposal (Technical and Financial) along with one photocopy in hard form and one in soft form (in USB) for each item/package. Contrary to these requirements, the firm has submitted a single proposal for multiple items (Item Nos. 26-29, 31, 32, 34, 36, 37, and 38).	Yes	Yes
2.	Bidder's Information Form	N/A	Yes	Yes



3.	Documentary proof active of National Tax Number (NTN) and General Sales Tax (GST) Number and Registration Professional Tax Certificate.	N/A	Yes	Yes
4.	Original Bid Security (In form of CDR) & Form	N/A	Yes	Yes
5.	Bid validity time is as per demand	N/A	Yes	Yes
6.	Warranty Form as per Tender	N/A	Yes	Yes
7.	Delivery as per Schedule of Requirement	N/A	Yes	Yes
8.	Undertaking on legal valid and attested stamp paper Rs. 100/- that the firm is not blacklisted by any of Provincial or Federal Government Department, Agency, Organization or autonomous body or Private Sector Organization anywhere in Pakistan.	N/A	Yes	No
9.	Embassy Attested Exclusively Authorized Agent / Sole Distributor Certificate by the Original Equipment Manufacturer (OEM).	N/A	Yes (For M/s Human Healthcare)	No (Embassy attested certificate not provided)
10.	Sub-Distributor Certificate from Sole Distributor having consent of OEM. (Where Applicable)	N/A	No (Not provided as per format of BD)	N/A
11.	Proof for having minimum 5 years of experience in supply, installation, commissioning of biomedical equipment.	N/A	No	Yes
12.	At least 05 verifiable Past Performance Certificates by the Procuring Agency of quoted products issued against the same Supply Orders.	N/A	Yes	No
13.	Technical & Engineering capability (As defined for the specific Bid in specifications)	N/A	Yes	Yes
14.	Originally signed documentary evidence from the Manufacturer that the quoted brand is the Original Equipment Manufacturer (OEM) with indication of manufacturing site and its location.	N/A	Yea	Yes
15.	Up to Date After Sales Service Certificate from OEM/Principal	N/A	Yes	Yes


16.	Availability of relevant Tools and Testing / Calibration Equipment.	N/A	Yes	Yes
17.	A Certificate from the OEM that the installation will be conducted in conformity with the system requirements.	N/A	Yes	Yes
18.	Consortiums or Joint Ventures (JVs) (Where applicable)	N/A	N/A	N/A
19.	Complete Package / Tender	No	No	No
Remarks:		Not Eligible	Not Eligible	Not Eligible
<b>PART – II</b> <b>KNOCKED DOWN CRITERIA</b> <b>(PRODUCT EVALUATION)</b>				
Item Ad. No.	Operation Table	M/s GMED	M/s Irfan Brothers International	M/s Sahaab International
	Brand	N/A	Medicinos Gija	Medallion Equipment
31	Quoted Model	N/A	Kepler – T5000	MT-T1
1.	Country of Manufacturer	N/A	Lithuania	UK
2.	Country of Origin of Product	N/A	Lithuania	Not Mentioned
3.	Compliance with defined Quality Standards	N/A	(Non Complied) (CE (MDD) from authorized EU body not available.)	(Non Complied) (Expired Certificate and (CE (MDD) from authorized EU body not available.)



4.	Specifications Compliance Feature Wise	N/A	Compliance	<p>T(Non Complied)</p> <p>The required specifications were copied on product catalog attached to the technical offer; moreover, the product information also could not be verified from the manufacturer's website.</p>
5.	Technical Eligibility of Product	N/A	Not Eligible	Not Eligible
6.	Technical Eligibility of Firm	Not Eligible	Not Eligible	Not Eligible
<b>Technical Status:</b>		<b>NON RESPONSIVE</b>	<b>NON RESPONSIVE</b>	<b>NON RESPONSIVE</b>

Item No. 32  
Item Name: Gyne OT Table  
Tender Name: Procurement of Bio-Medical Equipment Under PSDP Project  
"Establishment of 50 Bedded Cardiac Hospital, Gilgit".

PART – I KNOCKED DOWN CRITERIA (BID EVALUATION & VENDOR EVALUATION)		
Sr. #	Evaluation Parameters	M/s GMED
1.	Bid Form	As per BD ITB Clause 5, a bidder is required to submit only one bid for a specific item/package. Additionally, the tender advertisement explicitly stated the requirement to submit separate proposals/bids. This was further recorded in the minutes of the pre-bid meeting and clarified by the GRC, which specified that firms must submit an original proposal (Technical and Financial) along with one photocopy in hard form and one in soft form (in USB) for each item/package. Contrary to these requirements, the firm has submitted a single proposal for multiple items (Item Nos. 26-29, 31, 32, 34, 36, 37, and 38).
2.	Bidder's Information Form	N/A
3.	Documentary proof active of National Tax Number (NTN) and General Sales Tax (GST) Number and Registration Professional Tax Certificate.	N/A
4.	Original Bid Security (In form of CDR) & Form	N/A
5.	Bid validity time is as per demand	N/A
6.	Warranty Form as per Tender	N/A
7.	Delivery as per Schedule of Requirement	N/A
8.	Undertaking on legal valid and attested stamp paper Rs. 100/- that the firm is not blacklisted by any of Provincial or Federal Government Department, Agency, Organization or autonomous body or Private Sector Organization anywhere in Pakistan.	N/A
9.	Embassy Attested Exclusively Authorized Agent / Sole Distributor Certificate by the Original Equipment Manufacturer (OEM).	N/A
10.	Sub-Distributor Certificate from Sole Distributor having consent of OEM. (Where Applicable)	N/A
11.	Proof for having minimum 5 years of experience in supply, installation,	N/A



Handwritten signatures and initials in blue ink are present at the bottom of the page, including a large signature on the left, several smaller initials in the center, and a signature on the right next to the number 72.



	commissioning of biomedical equipment.	
12.	At least 05 verifiable Past Performance Certificates by the Procuring Agency of quoted products issued against the same Supply Orders.	N/A
13.	Technical & Engineering capability (As defined for the specific Bid in specifications)	N/A
14.	Originally signed documentary evidence from the Manufacturer that the quoted brand is the Original Equipment Manufacturer (OEM) with indication of manufacturing site and its location.	N/A
15.	Up to Date After Sales Service Certificate from OEM/Principal	N/A
16.	Availability of relevant Tools and Testing / Calibration Equipment.	N/A
17.	A Certificate from the OEM that the installation will be conducted in conformity with the system requirements.	N/A
18.	Consortiums or Joint Ventures (JVs) (Where applicable)	N/A
19.	Complete Package / Tender	No
Remarks:		Not Eligible

PART – II KNOCKED DOWN CRITERIA (PRODUCT EVALUATION)		
Item Ad. No.	Gyne OT Table	M/s GMED
	Brand	N/A
32	Quoted Model	N/A
1.	Country of Manufacturer	N/A
2.	Country of Origin of Product	N/A
3.	Compliance with defined Quality Standards	N/A
4.	Specifications Compliance Feature Wise	N/A
5.	Technical Eligibility of Product	N/A
6.	Technical Eligibility of Firm	Not Eligible
Technical Status:		NON RESPONSIVE

Item No. 34  
Item Name: Delivery Table  
Tender Name: Procurement of Bio-Medical Equipment Under PSDP Project  
"Establishment of 50 Bedded Cardiac Hospital, Gilgit".

PART – I KNOCKED DOWN CRITERIA (BID EVALUATION & VENDOR EVALUATION)			
Sr. #	Evaluation Parameters	M/s Irfan Brothers	M/s GMED
1.	Bid Form	Bid is not signed and stamped as required in BD Clause ITB 19.	As per BD ITB Clause 5, a bidder is required to submit only one bid for a specific item/package. Additionally, the tender advertisement explicitly stated the requirement to submit separate proposals/bids. This was further recorded in the minutes of the pre-bid meeting and clarified by the GRC, which specified that firms must submit an original proposal (Technical and Financial) along with one photocopy in hard form and one in soft form (in USB) for each item/package. Contrary to these requirements, the firm has submitted a single proposal for multiple items (Item Nos. 26-29, 31, 32, 34, 36, 37, and 38).
2.	Bidder's Information Form	Yes	N/A
3.	Documentary proof active of National Tax Number (NTN) and General Sales Tax (GST) Number and Registration Professional Tax Certificate.	Yes	N/A
4.	Original Bid Security (In form of CDR) & Form	Yes	N/A
5.	Bid validity time is as per demand	Yes	N/A
6.	Warranty Form as per Tender	Yes	N/A
7.	Delivery as per Schedule of Requirement	Yes	N/A
8.	Undertaking on legal valid and attested stamp paper Rs. 100/- that the firm is not blacklisted by any of Provincial or Federal Government Department, Agency, Organization or autonomous body or Private	No Unsigned & Un Stamped	N/A



	Sector Organization anywhere in Pakistan.		
9.	Embassy Attested Exclusively Authorized Agent / Sole Distributor Certificate by the Original Equipment Manufacturer (OEM).	No (For M/s Human Healthcare)	N/A
10	Sub-Distributor Certificate from Sole Distributor having consent of OEM. (Where Applicable)	No Not as per the format of BD. Moreover, the sole distributor claiming that they are the OEM of quoted manufacturer.	N/A
11	Proof for having minimum 5 years of experience in supply, installation, commissioning of biomedical equipment.	Yes	N/A
12	At least 05 verifiable Past Performance Certificates by the Procuring Agency of quoted products issued against the same Supply Orders.	No	N/A
13	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes	N/A
14	Originally signed documentary evidence from the Manufacturer that the quoted brand is the Original Equipment Manufacturer (OEM) with indication of manufacturing site and its location.	No	N/A
15	Up to Date After Sales Service Certificate from OEM/Principal	No	N/A
16	Availability of relevant Tools and Testing / Calibration Equipment.	Yes	N/A
17	A Certificate from the OEM that the installation will be conducted in conformity with the system requirements.	Yes	N/A
18	Consortiums or Joint Ventures (JVs) (Where applicable)	N/A	N/A
19	Complete Package / Tender	No	N/A
Remarks:		Not Eligible	Not Eligible
PART - II KNOCKED DOWN CRITERIA			

(PRODUCT EVALUATION)			
Item Ad. No.	Delivery Table	M/s Irfan Brothers	M/s GMED
	Brand	CLINIQON HC	N/A
34	Quoted Model	Grace Mother Care E	N/A
1.	Country of Manufacturer	UK	N/A
2.	Country of Origin of Product	Turkey	N/A
3.	Compliance with defined Quality Standards	(Non Complied) (CE (MDD) from authorized EU body not available.)	N/A
4.	Specifications Compliance Feature Wise	1. Overall width is 82 cm, while the demanded was 95 cm (950 mm). 2. The offered height is 51 cm to 82, instead 64.5 cm to 87 cm (645 mm to 870 mm). 3. Built in / remote control to not available.	N/A
5.	Technical Eligibility of Product	Not Eligible	N/A
6.	Technical Eligibility of Firm	Not Eligible	Not Eligible
Technical Status:		NON RESPONSIVE	NON RESPONSIVE



Item No. 35  
Item Name: Suction Machine  
Tender Name: Procurement of Bio-Medical Equipment Under PSDP Project "Establishment of 50 Bedded Cardiac Hospital, Gilgit".

PART - I KNOCKED DOWN CRITERIA (BID EVALUATION & VENDOR EVALUATION)				
Sr. #	Evaluation Parameters	M/s Sahaab International	M/s Irfan Brothers International	M/s Medifa Enterprises
1.	Bid Form	No	Yes	Yes
2.	Bidder's Information Form	Yes	Yes	Yes
3.	Documentary proof active of National Tax Number (NTN) and General Sales Tax (GST) Number and Registration Professional Tax Certificate.	Yes	Yes	Yes
4.	Original Bid Security (In form of CDR) & Form	Yes	Yes	Yes
5.	Bid validity time is as per demand	Yes	Yes	Yes
6.	Warranty Form as per Tender	Yes	Yes	Yes
7.	Delivery as per Schedule of Requirement	Yes	Yes	Yes
8.	Undertaking on legal valid and attested stamp paper Rs. 100/- that the firm is not blacklisted by any of Provincial or Federal Government Department, Agency, Organization or autonomous body or Private Sector Organization anywhere in Pakistan.	No	Yes	No (Not as per format of BD)
9.	Embassy Attested Exclusively Authorized Agent / Sole Distributor Certificate by the Original Equipment Manufacturer (OEM).	No	Yes (For M/s Human Healthcare)	No
10.	Sub-Distributor Certificate from Sole Distributor having consent of OEM. (Where Applicable)	N/A	Yes	Yes
11.	Proof for having minimum 5 years of experience in supply, installation, commissioning of biomedical equipment.	Yes	Yes	Yes
12.	At least 05 verifiable Past Performance Certificates by the Procuring Agency of quoted products issued against the same Supply Orders.	No	Yes	Yes

Handwritten signatures and initials in blue ink are present at the bottom of the page, including a large signature on the left, several smaller initials in the center, and a signature with the number '77' on the right.

13.	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes	Yes	Yes
14.	Originally signed documentary evidence from the Manufacturer that the quoted brand is the Original Equipment Manufacturer (OEM) with indication of manufacturing site and its location.	No	Yes	Yes
15.	Up to Date After Sales Service Certificate from OEM/Principal	No	Yes	Yes
16.	Availability of relevant Tools and Testing / Calibration Equipment.	Yes	Yes	Yes
17.	A Certificate from the OEM that the installation will be conducted in conformity with the system requirements.	No	Yes	Yes
18.	Consortiums or Joint Ventures (JVs) (Where applicable)	N/A	N/A	N/A
19.	Complete Package / Tender	No	Yes	No
Remarks:		Not Eligible	Eligible	Not Eligible
<b>PART – II</b> <b>KNOCKED DOWN CRITERIA</b> <b>(PRODUCT EVALUATION)</b>				
Item Ad. No.	Suction Machine	M/s Sahaab International	M/s Irfan Brothers	M/s Medifa Enterprises
	Brand	Medallion Equipment	Medicinos Gija	CA-MI Srl
35	Quoted Model	M23-MII	Prima 50	New Hospivac 400 Full PSU 5
1.	Country of Manufacturer	UK	Lithuania	Italy
2.	Country of Origin of Product	UK	Lithuania	Italy
3.	Compliance with defined Quality Standards	(Non Complied) (Certificate could not verified online) and (CE (MDD) from authorized EU body not available.)	(Non Complied) No (CE (MDD) from authorized EU body not available.)	(Compliance) CE (MDD)
4.	Specifications Compliance Feature Wise	(Non Complied)	Compliance	(Non Complied)



		4. Triple or over flow safety device not available. 5. Change over valve not available. 6. Duplex offer (ACT Machine also quoted in the same bid.) 7. The required specifications were copied on product catalog attached to the technical offer; moreover, the product information also could not be verified from the manufacturer's website.		1. (Polysulfide or Polycarbonate type) not verified. 2. Triple or over flow safety device not available.
5.	Technical Eligibility of Product	Not Eligible	Not Eligible	Not Eligible
6.	Technical Eligibility of Firm	Not Eligible	Eligible	Not Eligible
<b>Technical Status:</b>		<b>NON RESPONSIVE</b>	<b>NON RESPONSIVE</b>	<b>NON RESPONSIVE</b>

Item No.  
Item Name:  
Tender Name:

36  
Diathermy Machine  
Procurement of Bio-Medical Equipment Under PSDP Project  
"Establishment of 50 Bedded Cardiac Hospital, Gilgit".

PART – I KNOCKED DOWN CRITERIA (BID EVALUATION & VENDOR EVALUATION)			
Sr. #	Evaluation Parameters	M/s Irfan Brothers	M/s GMED
1.	Bid Form	Bid is not signed and stamped as required in BD Clause ITB 19.	As per BD ITB Clause 5, a bidder is required to submit only one bid for a specific item/package. Additionally, the tender advertisement explicitly stated the requirement to submit separate proposals/bids. This was further recorded in the minutes of the pre-bid meeting and clarified by the GRC, which specified that firms must submit an original proposal (Technical and Financial) along with one photocopy in hard form and one in soft form (in USB) for each item/package. Contrary to these requirements, the firm has submitted a single proposal for multiple items (Item Nos. 26-29, 31, 32, 34, 36, 37, and 38).
2.	Bidder's Information Form	Yes	N/A
3.	Documentary proof active of National Tax Number (NTN) and General Sales Tax (GST) Number and Registration Professional Tax Certificate.	Yes	N/A
4.	Original Bid Security (In form of CDR) & Form	Yes	N/A
5.	Bid validity time is as per demand	Yes	N/A
6.	Warranty Form as per Tender	Yes	N/A
7.	Delivery as per Schedule of Requirement	Yes	N/A
8.	Undertaking on legal valid and attested stamp paper Rs. 100/- that the firm is not blacklisted by any of Provincial or Federal Government Department, Agency, Organization or autonomous body or Private	No (Unsigned & Un Stamped)	N/A



	Sector Organization anywhere in Pakistan.		
9.	Embassy Attested Exclusively Authorized Agent / Sole Distributor Certificate by the Original Equipment Manufacturer (OEM).	Yes	N/A
10.	Sub-Distributor Certificate from Sole Distributor having consent of OEM. (Where Applicable)	No Not as per the format of BD. Moreover, the sole distributor claiming that they are the OEM of quoted manufacturer.	N/A
11.	Proof for having minimum 5 years of experience in supply, installation, commissioning of biomedical equipment.	Yes	N/A
12.	At least 05 verifiable Past Performance Certificates by the Procuring Agency of quoted products issued against the same Supply Orders.	No	N/A
13.	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes	N/A
14.	Originally signed documentary evidence from the Manufacturer that the quoted brand is the Original Equipment Manufacturer (OEM) with indication of manufacturing site and its location.	Yes	N/A
15.	Up to Date After Sales Service Certificate from OEM/Principal	Yes	N/A
16.	Availability of relevant Tools and Testing / Calibration Equipment.	Yes	N/A
17.	A Certificate from the OEM that the installation will be conducted in conformity with the system requirements.	Yes	N/A
18.	Consortiums or Joint Ventures (JVs) (Where applicable)	N/A	N/A
19.	Complete Package / Tender	No	No
Remarks:		Not Eligible	Not Eligible

PART – II KNOCKED DOWN CRITERIA (PRODUCT EVALUATION)			
Item Ad. No.	Diathermy Machine	M/s Irfan Brothers	M/s GMED
	Brand	Medicinos Gija	N/A
36	Quoted Model	Energia 400T	N/A
1.	Country of Manufacturer	Lithuania	N/A
2.	Country of Origin of Product	Lithuania	N/A
3.	Compliance with defined Quality Standards	(Non Complied) (CE (MDD) from authorized EU body not available.)	N/A
4.	Specifications Compliance Feature Wise	Compliance	N/A
5.	Technical Eligibility of Product	Not Eligible	N/A
6.	Technical Eligibility of Firm	Not Eligible	Not Eligible
Technical Status:		NON RESPONSIVE	NON RESPONSIVE



Item No.  
Item Name:  
Tender Name:

37  
Sternum Saw  
Procurement of Bio-Medical Equipment Under PSDP Project  
"Establishment of 50 Bedded Cardiac Hospital, Gilgit".

PART – I KNOCKED DOWN CRITERIA (BID EVALUATION & VENDOR EVALUATION)			
Sr. #	Evaluation Parameters	M/s Jamil Enterprises	M/s GMED
1.	Bid Form	Yes	As per BD ITB Clause 5, a bidder is required to submit only one bid for a specific item/package. Additionally, the tender advertisement explicitly stated the requirement to submit separate proposals/bids. This was further recorded in the minutes of the pre-bid meeting and clarified by the GRC, which specified that firms must submit an original proposal (Technical and Financial) along with one photocopy in hard form and one in soft form (in USB) for each item/package. Contrary to these requirements, the firm has submitted a single proposal for multiple items (Item Nos. 26-29, 31, 32, 34, 36, 37, and 38). Further GMED do not have exclusive authorization for the said tender as verified by the manufacturer.
2.	Bidder's Information Form	Yes	N/A
3.	Documentary proof active of National Tax Number (NTN) and General Sales Tax (GST) Number and Registration Professional Tax Certificate.	Yes	N/A
4.	Original Bid Security (In form of CDR) & Form	Yes	N/A
5.	Bid validity time is as per demand	Yes	N/A
6.	Warranty Form as per Tender	Yes	N/A
7.	Delivery as per Schedule of Requirement	Yes	N/A
8.	Undertaking on legal valid and attested stamp paper Rs.	Yes	N/A

	100/- that the firm is not blacklisted by any of Provincial or Federal Government Department, Agency, Organization or autonomous body or Private Sector Organization anywhere in Pakistan.		
9.	Embassy Attested Exclusively Authorized Agent / Sole Distributor Certificate by the Original Equipment Manufacturer (OEM).	Yes (For M/s Claris Medical)	N/A
10.	Sub-Distributor Certificate from Sole Distributor having consent of OEM. (Where Applicable)	Yes	N/A
11.	Proof for having minimum 5 years of experience in supply, installation, commissioning of biomedical equipment.	Yes	N/A
12.	At least 05 verifiable Past Performance Certificates by the Procuring Agency of quoted products issued against the same Supply Orders.	Yes	N/A
13.	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes	N/A
14.	Originally signed documentary evidence from the Manufacturer that the quoted brand is the Original Equipment Manufacturer (OEM) with indication of manufacturing site and its location.	Yes	N/A
15.	Up to Date After Sales Service Certificate from OEM/Principal	Yes	N/A
16.	Availability of relevant Tools and Testing / Calibration Equipment.	Yes	N/A
17.	A Certificate from the OEM that the installation will be conducted in conformity with the system requirements.	Yes	N/A
18.	Consortiums or Joint Ventures (JVs) (Where applicable)	N/A	N/A
19.	Complete Package / Tender	Yes	No
Remarks:		Eligible	Not Eligible
PART - II			



KNOCKED DOWN CRITERIA (PRODUCT EVALUATION)			
Item Ad. No.	Sternum Saw	M/s Jamil Enterprises	M/s GMED
	Brand	DeSoutter Medical	N/A
37	Quoted Model	DBC-470	N/A
1.	Country of Manufacturer	UK	N/A
2.	Country of Origin of Product	UK	N/A
3.	Compliance with defined Quality Standards	(Compliance) CE (MDD)	N/A
4.	Specifications Compliance Feature Wise	Compliance	N/A
5.	Technical Eligibility of Product	Eligible	N/A
6.	Technical Eligibility of Firm	Eligible	Not Eligible
Technical Status:		RESPONSIVE	NON RESPONSIVE

2

Handwritten signatures and initials in blue ink, including a large signature on the left, a circular stamp in the center, and several smaller signatures and initials scattered around.

Item No.  
Item Name:  
Tender Name:

38  
Sternum Saw (RS)  
Procurement of Bio-Medical Equipment Under PSDP Project  
"Establishment of 50 Bedded Cardiac Hospital, Gilgit".

PART – I KNOCKED DOWN CRITERIA (BID EVALUATION & VENDOR EVALUATION)			
Sr. #	Evaluation Parameters	M/s Jamil Enterprises	M/s GMED
1.	Bid Form	Yes	As per BD ITB Clause 5, a bidder is required to submit only one bid for a specific item/package. Additionally, the tender advertisement explicitly stated the requirement to submit separate proposals/bids. This was further recorded in the minutes of the pre-bid meeting and clarified by the GRC, which specified that firms must submit an original proposal (Technical and Financial) along with one photocopy in hard form and one in soft form (in USB) for each item/package. Contrary to these requirements, the firm has submitted a single proposal for multiple items (Item Nos. 26-29, 31, 32, 34, 36, 37, and 38). Further GMED do not have exclusive authorization for the said tender as verified by the manufacturer.
2.	Bidder's Information Form	Yes	N/A
3.	Documentary proof active of National Tax Number (NTN) and General Sales Tax (GST) Number and Registration Professional Tax Certificate.	Yes	N/A
4.	Original Bid Security (In form of CDR) & Form	Yes	N/A
5.	Bid validity time is as per demand	Yes	N/A
6.	Warranty Form as per Tender	Yes	N/A
7.	Delivery as per Schedule of Requirement	Yes	N/A
8.	Undertaking on legal valid and attested stamp paper Rs. 100/- that the firm is not blacklisted by any of	Yes	N/A



	Provincial or Federal Government Department, Agency, Organization or autonomous body or Private Sector Organization anywhere in Pakistan.		
9.	Embassy Attested Exclusively Authorized Agent / Sole Distributor Certificate by the Original Equipment Manufacturer (OEM).	Yes (For M/s Claris Medical)	N/A
10.	Sub-Distributor Certificate from Sole Distributor having consent of OEM. (Where Applicable)	Yes	N/A
11.	Proof for having minimum 5 years of experience in supply, installation, commissioning of biomedical equipment.	Yes	N/A
12.	At least 05 verifiable Past Performance Certificates by the Procuring Agency of quoted products issued against the same Supply Orders.	Yes	N/A
13.	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes	N/A
14.	Originally signed documentary evidence from the Manufacturer that the quoted brand is the Original Equipment Manufacturer (OEM) with indication of manufacturing site and its location.	Yes	N/A
15.	Up to Date After Sales Service Certificate from OEM/Principal	Yes	N/A
16.	Availability of relevant Tools and Testing / Calibration Equipment.	Yes	N/A
17.	A Certificate from the OEM that the installation will be conducted in conformity with the system requirements.	Yes	N/A
18.	Consortiums or Joint Ventures (JVs) (Where applicable)	N/A	N/A
19.	Complete Package / Tender	Yes	No
Remarks:		Eligible	Not Eligible
<b>PART – II</b> <b>KNOCKED DOWN CRITERIA</b> <b>(PRODUCT EVALUATION)</b>			

Item Ad. No.	Sternum Saw (RS)	M/s Jamil Enterprises	M/s GMED
	Brand	DeSoutter Medical	N/A
38	Quoted Model	DPX-170	N/A
1.	Country of Manufacturer	UK	N/A
2.	Country of Origin of Product	UK	N/A
3.	Compliance with defined Quality Standards	(Compliance) ISO 13485:2016, CE & FDA Device Listing	N/A
4.	Specifications Compliance Feature Wise	Compliance	N/A
5.	Technical Eligibility of Product	Eligible	N/A
6.	Technical Eligibility of Firm	Eligible	Not Eligible
<b>Technical Status:</b>		<b>RESPONSIVE</b>	<b>NON RESPONSIVE</b>



Item No. 39  
Item Name: ACT Machine  
Tender Name: Procurement of Bio-Medical Equipment Under PSDP Project  
"Establishment of 50 Bedded Cardiac Hospital, Gilgit".

PART – I KNOCKED DOWN CRITERIA (BID EVALUATION & VENDOR EVALUATION)			
Sr. #	Evaluation Parameters	M/s Sahaab International	M/s Jamil Enterprises
1.	Bid Form	Yes	Yes
2.	Bidder's Information Form	Yes	Yes
3.	Documentary proof active of National Tax Number (NTN) and General Sales Tax (GST) Number and Registration Professional Tax Certificate.	Yes	Yes
4.	Original Bid Security (In form of CDR) & Form	Yes	Yes
5.	Bid validity time is as per demand	Yes	Yes
6.	Warranty Form as per Tender	Yes	Yes
7.	Delivery as per Schedule of Requirement	Yes	Yes
8.	Undertaking on legal valid and attested stamp paper Rs. 100/- that the firm is not blacklisted by any of Provincial or Federal Government Department, Agency, Organization or autonomous body or Private Sector Organization anywhere in Pakistan.	No	Yes
9.	Embassy Attested Exclusively Authorized Agent / Sole Distributor Certificate by the Original Equipment Manufacturer (OEM).	No	Yes (For M/s Iqbal & Company)
10.	Sub-Distributor Certificate from Sole Distributor having consent of OEM. (Where Applicable)	N/A	Yes
11.	Proof for having minimum 5 years of experience in supply, installation, commissioning of biomedical equipment.	Yes	Yes
12.	At least 05 verifiable Past Performance Certificates by the Procuring Agency of quoted products issued against the same Supply Orders.	No	Yes
13.	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes	Yes
14.	Originally signed documentary evidence from the Manufacturer that the quoted brand is the Original Equipment Manufacturer (OEM) with indication of manufacturing site and its location.	No	Yes
15.	Up to Date After Sales Service Certificate from OEM/Principal	No	Yes
16.	Availability of relevant Tools and Testing / Calibration Equipment.	Yes	Yes

17.	A Certificate from the OEM that the installation will be conducted in conformity with the system requirements.	No	Yes
18.	Consortiums or Joint Ventures (JVs) (Where applicable)	N/A	N/A
19.	Complete Package / Tender	No	Yes
Remarks:		Not Eligible	Eligible
<b>PART – II</b> <b>KNOCKED DOWN CRITERIA</b> <b>(PRODUCT EVALUATION)</b>			
Item Ad. No.	ACT Machine	M/s Sahaab International	M/s Jamil Enterprises
	Brand	Medallion Equipment	Werfen/II Accriva Diagnostic
39	Quoted Model	MCG-M2	Hemochron Signature Elite
1.	Country of Manufacturer	UK	USA
2.	Country of Origin of Product	UK	USA
3.	Compliance with defined Quality Standards	(Non Complied) (Certificate could not verified online) and (CE (MDD) from authorized EU body not available.)	(Compliance) FDA 510K
4.	Specifications Compliance Feature Wise	(Non Complied)  The required specifications were copied on product catalog attached to the technical offer; moreover, the product information also could not be verified from the manufacturer's website.	Compliance
5.	Technical Eligibility of Product	Not Eligible	Eligible
6.	Technical Eligibility of Firm	Not Eligible	Eligible
Technical Status:		<b>NON RESPONSIVE</b>	<b>RESPONSIVE</b>



Item No. 40  
Item Name: Gamma Camera with Hot Lab  
Tender Name: Procurement of Bio-Medical Equipment Under PSDP Project  
"Establishment of 50 Bedded Cardiac Hospital, Gilgit".

Sr. #	Evaluation Parameters	M/s Shirazi Trading Pvt. Ltd.
1.	Bid Form	Yes
2.	Bidder's Information Form	Yes
3.	Documentary proof active of National Tax Number (NTN) and General Sales Tax (GST) Number and Registration Professional Tax Certificate.	Yes
4.	Original Bid Security (In form of CDR) & Form	Yes
5.	Bid validity time is as per demand	Yes
6.	Warranty Form as per Tender	Yes
7.	Delivery as per Schedule of Requirement	Yes
8.	Undertaking on legal valid and attested stamp paper Rs. 100/- that the firm is not blacklisted by any of Provincial or Federal Government Department, Agency, Organization or autonomous body or Private Sector Organization anywhere in Pakistan.	Yes
9.	Embassy Attested Exclusively Authorized Agent / Sole Distributor Certificate by the Original Equipment Manufacturer (OEM).	Yes
10.	Sub-Distributor Certificate from Sole Distributor having consent of OEM. (Where Applicable)	N/A
11.	Proof for having minimum 5 years of experience in supply, installation, commissioning of biomedical equipment.	Yes
12.	At least 05 verifiable Past Performance Certificates by the Procuring Agency of quoted products issued against the same Supply Orders.	Yes
13.	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes
14.	Originally signed documentary evidence from the Manufacturer that the quoted brand is the Original Equipment Manufacturer (OEM) with indication of manufacturing site and its location.	Yes
15.	Up to Date After Sales Service Certificate from OEM/Principal	Yes
16.	Availability of relevant Tools and Testing / Calibration Equipment.	Yes
17.	A Certificate from the OEM that the installation will be conducted in conformity with the system requirements.	Yes
18.	Consortiums or Joint Ventures (JVs) (Where applicable)	N/A
19.	Complete Package / Tender	Yes
Remarks:		Eligible
Item Ad. No.	Gamma Camera with Hot Lab	M/s Shirazi Trading Pvt. Ltd.
	Brand	GE Healthcare
40	Quoted Model	NM 830

1.	Country of Manufacturer	USA
2.	Country of Origin of Product	China
3.	Compliance with defined Quality Standards	(Compliance) CE MDD & FDA10K
4.	Specifications Compliance Feature Wise	Compliance
5.	Technical Eligibility of Product	Eligible
6.	Technical Eligibility of Firm	Eligible
<b>Technical Status:</b>		<b>RESPONSIVE</b>

The page contains several handwritten signatures and initials in blue ink, scattered across the lower half of the document. These include a large signature on the left, a signature in the center, a signature on the right, and several smaller initials and marks at the bottom.



Item No.  
Item Name:  
Tender Name:

41  
Echocardiography Machines  
Procurement of Bio-Medical Equipment Under PSDP Project  
"Establishment of 50 Bedded Cardiac Hospital, Gilgit".

PART – I KNOCKED DOWN CRITERIA (BID EVALUATION & VENDOR EVALUATION)			
Sr. #	Evaluation Parameters	M/s Hoora Pharma	M/s Medequips (SMC) Pvt. Ltd.
1.	Bid Form	Yes	Yes
2.	Bidder's Information Form	Yes	Yes
3.	Documentary proof active of National Tax Number (NTN) and General Sales Tax (GST) Number and Registration Professional Tax Certificate.	Yes	Yes
4.	Original Bid Security (In form of CDR) & Form	Yes	Yes
5.	Bid validity time is as per demand	Yes	Yes
6.	Warranty Form as per Tender	Yes	Yes
7.	Delivery as per Schedule of Requirement	Yes	Yes
8.	Undertaking on legal valid and attested stamp paper Rs. 100/- that the firm is not blacklisted by any of Provincial or Federal Government Department, Agency, Organization or autonomous body or Private Sector Organization anywhere in Pakistan.	Yes	Yes
9.	Embassy Attested Exclusively Authorized Agent / Sole Distributor Certificate by the Original Equipment Manufacturer (OEM).	No	Yes
10.	Sub-Distributor Certificate from Sole Distributor having consent of OEM. (Where Applicable)	N/A	N/A
11.	Proof for having minimum 5 years of experience in supply, installation, commissioning of biomedical equipment.	Yes	Yes
12.	At least 05 verifiable Past Performance Certificates by the Procuring Agency of quoted products issued against the same Supply Orders.	Yes	Yes
13.	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes	Yes
14.	Originally signed documentary evidence from the Manufacturer that the quoted brand is the Original Equipment Manufacturer (OEM) with indication of manufacturing site and its location.	Yes	Yes

15.	Up to Date After Sales Service Certificate from OEM/Principal	Yes	Yes
16.	Availability of relevant Tools and Testing / Calibration Equipment.	Yes	Yes
17.	A Certificate from the OEM that the installation will be conducted in conformity with the system requirements.	Yes	Yes
18.	Consortiums or Joint Ventures (JVs) (Where applicable)	N/A	N/A
19.	Complete Package / Tender	Yes	Yes
Remarks:		Not Eligible	Eligible
<b>PART – II</b> <b>KNOCKED DOWN CRITERIA</b> <b>(PRODUCT EVALUATION)</b>			
Item Ad. No.	Echocardiography Machines	M/s Hoora Pharma	M/s Medequips (SMC) Pvt. Ltd.
	Brand	GE Healthcare	Canon Medical
41	Quoted Model	Echocardiography Machine: S60N and S70N	Echocardiography Machine: Aplio a, Aplio i-700, Probes: PET-512MB & PET-508MA
1.	Country of Manufacturer	USA	Japan
2.	Country of Origin of Product	Norway	Japan
3.	Compliance with defined Quality Standards	(Compliance) CE MDD & FDA10K	(Compliance) MHLW & FDA10K
4.	Specifications Compliance Feature Wise	(Non Complied)  As per GRC meeting minutes, country of manufacturer and origin should be same as USA or Europe or Japan, while the quoted origin is Norway which different from Country of manufacturer.	Compliance
5.	Technical Eligibility of Product	Not Eligible	Eligible
6.	Technical Eligibility of Firm	Not Eligible	Eligible
Technical Status:		NON RESPONSIVE	RESPONSIVE



Item No.  
Item Name:

42  
**General Purpose Ultrasound Machine (dplr) & Portable  
Ultrasound Machine (dplr)**  
Tender Name: Procurement of Bio-Medical Equipment Under PSDP Project  
"Establishment of 50 Bedded Cardiac Hospital, Gilgit".

PART – I KNOCKED DOWN CRITERIA (BID EVALUATION & VENDOR EVALUATION)			
Sr. #	Evaluation Parameters	M/s Biotech Services	M/s Medequips (SMC) Pvt. Ltd.
1.	Bid Form	Yes	Yes
2.	Bidder's Information Form	Yes	Yes
3.	Documentary proof active of National Tax Number (NTN) and General Sales Tax (GST) Number and Registration Professional Tax Certificate.	Yes	Yes
4.	Original Bid Security (In form of CDR) & Form	Yes	Yes
5.	Bid validity time is as per demand	Yes	Yes
6.	Warranty Form as per Tender	Yes	Yes
7.	Delivery as per Schedule of Requirement	Yes	Yes
8.	Undertaking on legal valid and attested stamp paper Rs. 100/- that the firm is not blacklisted by any of Provincial or Federal Government Department, Agency, Organization or autonomous body or Private Sector Organization anywhere in Pakistan.	Yes	Yes
9.	Embassy Attested Exclusively Authorized Agent / Sole Distributor Certificate by the Original Equipment Manufacturer (OEM).	No	No
10.	Sub-Distributor Certificate from Sole Distributor having consent of OEM. (Where Applicable)	N/A	N/A
11.	Proof for having minimum 5 years of experience in supply, installation, commissioning of biomedical equipment.	Yes	Yes
12.	At least 05 verifiable Past Performance Certificates by the Procuring Agency of quoted products issued against the same Supply Orders.	Yes	Yes
13.	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes	Yes
14.	Originally signed documentary evidence from the Manufacturer that the quoted brand is the Original Equipment Manufacturer (OEM) with indication of manufacturing site and its location.	Yes	Yes
15.	Up to Date After Sales Service Certificate from OEM/Principal	Yes	Yes
16.	Availability of relevant Tools and Testing / Calibration Equipment.	Yes	Yes

17.	A Certificate from the OEM that the installation will be conducted in conformity with the system requirements.	Yes	Yes
18.	Consortiums or Joint Ventures (JVs) (Where applicable)	N/A	N/A
19.	Complete Package / Tender	No The qty of equipment is not quoted as per ad. (I.e. Required qty of Portable Ultrasound was 03, while the firm quoted 01, on the other hand General Purpose Ultrasound Machine (dplr) required qty was 03, the firm quoted 01.	No The qty of equipment is not quoted as per ad. (I.e. Required qty of Portable Ultrasound was 03, while the firm quoted 01, on the other hand General Purpose Ultrasound Machine (dplr) required qty was 03, the firm quoted 01.
Remarks:		Not Eligible	Not Eligible
<b>PART – II</b> <b>KNOCKED DOWN CRITERIA</b> <b>(PRODUCT EVALUATION)</b>			
Item Ad. No.	General Purpose Ultrasound Machine (dplr) & Portable Ultrasound Machine (dplr)	M/s Biotech Services	M/s Medequips (SMC) Pvt. Ltd.
	Brand	Esaote	Konica Minolta & Canon Medical
42	Quoted Model	N/A	N/A
1.	Country of Manufacturer	Italy	Japan
2.	Country of Origin of Product	Italy	Japan
3.	Compliance with defined Quality Standards	N/A	N/A
4.	Specifications Compliance Feature Wise	N/A	N/A
5.	Technical Eligibility of Product	N/A	N/A
6.	Technical Eligibility of Firm	Not Eligible	Not Eligible
Technical Status:		<b>NON RESPONSIVE</b>	<b>NON RESPONSIVE</b>



Item No. 43  
Item Name: External Pacemakers (Temporary Pace Maker (Single Chamber) & Temporary Pace Maker Double Chamber)  
Tender Name: Procurement of Bio-Medical Equipment Under PSDP Project "Establishment of 50 Bedded Cardiac Hospital, Gilgit".

PART – I KNOCKED DOWN CRITERIA (BID EVALUATION & VENDOR EVALUATION)		
Sr. #	Evaluation Parameters	M/s Jamil Enterprises
1.	Bid Form	Yes
2.	Bidder's Information Form	Yes
3.	Documentary proof active of National Tax Number (NTN) and General Sales Tax (GST) Number and Registration Professional Tax Certificate.	Yes
4.	Original Bid Security (In form of CDR) & Form	Yes
5.	Bid validity time is as per demand	Yes
6.	Warranty Form as per Tender	Yes
7.	Delivery as per Schedule of Requirement	Yes
8.	Undertaking on legal valid and attested stamp paper Rs. 100/- that the firm is not blacklisted by any of Provincial or Federal Government Department, Agency, Organization or autonomous body or Private Sector Organization anywhere in Pakistan.	Yes
9.	Embassy Attested Exclusively Authorized Agent / Sole Distributor Certificate by the Original Equipment Manufacturer (OEM).	Yes (For M/s Claris Medical)
10.	Sub-Distributor Certificate from Sole Distributor having consent of OEM. (Where Applicable)	Yes
11.	Proof for having minimum 5 years of experience in supply, installation, commissioning of biomedical equipment.	Yes
12.	At least 05 verifiable Past Performance Certificates by the Procuring Agency of quoted products issued against the same Supply Orders.	Yes
13.	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes
14.	Originally signed documentary evidence from the Manufacturer that the quoted brand is the Original Equipment Manufacturer (OEM) with indication of manufacturing site and its location.	Yes
15.	Up to Date After Sales Service Certificate from OEM/Principal	Yes
16.	Availability of relevant Tools and Testing / Calibration Equipment.	Yes
17.	A Certificate from the OEM that the installation will be conducted in conformity with the system requirements.	Yes
18.	Consortiums or Joint Ventures (JVs) (Where applicable)	N/A
19.	Complete Package / Tender	Yes
Remarks:		Eligible
PART – II KNOCKED DOWN CRITERIA		

(PRODUCT EVALUATION)		
Item Ad. No.	External Pacemakers (Temporary Pace Maker (Single Chamber) & Temporary Pace Maker Double Chamber)	M/s Jamil Enterprises
	Brand	Osypka Medical
43	Quoted Model	Pace 101 & Pace 203
1.	Country of Manufacturer	Germany
2.	Country of Origin of Product	Germany
3.	Compliance with defined Quality Standards	(Compliance) DRAP, CE MDD & FDA10K
4.	Specifications Compliance Feature Wise	Compliance
5.	Technical Eligibility of Product	Eligible
6.	Technical Eligibility of Firm	Eligible
Technical Status:		RESPONSIVE

Handwritten signatures and initials in blue ink, including a large signature at the top left, a signature with a checkmark at the top center, a signature with a checkmark at the top right, a signature with a checkmark at the bottom left, a signature with a checkmark at the bottom center, a signature with a checkmark at the bottom right, and a signature with a checkmark at the bottom left.



Item No. 44

Item Name: Blood Gas Analyzer

Tender Name: Procurement Of Bio-Medical Equipment Under PSDP Project "Establishment of 50 Bedded Cardiac Hospital, Gilgit".

PART – I KNOCKED DOWN CRITERIA (BID EVALUATION & VENDOR EVALUATION)					
Sr. #	Evaluation Parameters	M/s Irfan Brothers	M/s Jamil Enterprises	M/s Hoora Pharma	M/s Sahaab International
1.	Bid Form	Yes	Yes	No (Same Bid Form for Bid No. 41, 44, 47, 48, 51.)	Yes
2.	Bidder's Information Form	Yes	Yes	Yes	Yes
3.	Documentary proof active of National Tax Number (NTN) and General Sales Tax (GST) Number and Registration Professional Tax Certificate.	Yes	Yes	Yes	Yes
4.	Original Bid Security (In form of CDR) & Form	Yes	Yes	No	No
5.	Bid validity time is as per demand	Yes	Yes	Yes	Yes
6.	Warranty Form as per Tender	Yes	Yes	Yes	Yes
7.	Delivery as per Schedule of Requirement	Yes	Yes	Yes	Yes
8.	Undertaking on legal valid and attested stamp paper Rs. 100/- that the firm is not blacklisted by any of Provincial or Federal Government Department, Agency, Organization or autonomous body or Private Sector Organization anywhere in Pakistan.	Yes	Yes	Yes	No
9.	Embassy Attested Exclusively Authorized Agent / Sole Distributor Certificate by the Original Equipment Manufacturer (OEM).	No (For M/s Human Healthcare)	Yes (For M/s Meditec Instruments & Co.)	Yes	No (Embassy attested Certificate not provided)
10.	Sub-Distributor Certificate from Sole Distributor having consent of OEM. (Where Applicable)	Yes	Yes	N/A	N/A

11.	Proof for having minimum 5 years of experience in supply, installation, commissioning of biomedical equipment.	Yes	Yes	Yes	Yes
12.	At least 05 verifiable Past Performance Certificates by the Procuring Agency of quoted products issued against the same Supply Orders.	No	Yes	Yes	No
13.	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes	Yes	Yes	Yes
14.	Originally signed documentary evidence from the Manufacturer that the quoted brand is the Original Equipment Manufacturer (OEM) with indication of manufacturing site and its location.	Yes	Yes	No	No
15.	Up to Date After Sales Service Certificate from OEM/Principal	Yes	Yes	Yes	Yes
16.	Availability of relevant Tools and Testing / Calibration Equipment.	Yes	Yes	Yes	Yes
17.	A Certificate from the OEM that the installation will be conducted in conformity with the system requirements.	Yes	Yes	Yes	Yes
18.	Consortiums or Joint Ventures (JVs) (Where applicable)	N/A	N/A	N/A	N/A
19.	Complete Package / Tender	No	Yes	No	No
Remarks:		Not Eligible	Eligible	Not Eligible	Not Eligible
<b>PART – II</b> <b>KNOCKED DOWN CRITERIA</b> <b>(PRODUCT EVALUATION)</b>					
Item Ad. No.	Blood Gas Analyzer	M/s Irfan Brothers	M/s Jamil Enterprises	M/s Hoora Pharma	M/s Sahaab International
	Brand	VITANIX D.O.O	Medica	Siemens Healthineers	Medallion Equipment
44	Quoted Model	VITAGAS	Easystat	Rapid Point 500	MPT-100
1.	Country of Manufacturer	Slovenia	USA	USA	UK
2.	Country of Origin of Product	Slovenia	USA	UK	UK



3.	Compliance with defined Quality Standards	(Non Complied) CE (MDD) certificate not from authorized EU body is attached.	Compliance FDA 510K	(Compliance) CE (MDD) & FDA CFG	(Non Complied) (Certificate could not verified online) and (CE (MDD) from authorized EU body not available.)
4.	Specifications Compliance Feature Wise	(Non Complied)  3. Na+, K+, Cl- and HCT is not offered in Technical Offer. 4. No Embassy Certificate is attached. 5. No CE (MDD) certificate from authorized EU body is attached.	(Non Complied)  1. Storage memory is 64 sample instead of required for minimum of 150.	(Non Complied)  1. The country of origin is not same as country of manufacturer.	(Non Complied)  6. The quoted model is not available on official website of quoted manufacturer and The required specifications were copied on product catalog attached to the technical offer.
5.	Technical Eligibility of Product	Not Eligible	Not Eligible	Not Eligible	Not Eligible
6.	Technical Eligibility of Firm	Not Eligible	Eligible	Not Eligible	Not Eligible
Technical Status		NON RESPONSIVE	NON RESPONSIVE	NON RESPONSIVE	NON RESPONSIVE

Item No. 45  
Item Name: Elisa System  
Tender Name: Procurement of Bio-Medical Equipment Under PSDP Project  
"Establishment of 50 Bedded Cardiac Hospital, Gilgit".

PART – I KNOCKED DOWN CRITERIA (BID EVALUATION & VENDOR EVALUATION)			
Sr. #	Evaluation Parameters	M/s Sahaab International.	M/s Irfan Brothers Intl.
1.	Bid Form	Yes	Yes
2.	Bidder's Information Form	Yes	Yes
3.	Documentary proof active of National Tax Number (NTN) and General Sales Tax (GST) Number and Registration Professional Tax Certificate.	Yes	Yes
4.	Original Bid Security (In form of CDR) & Form	Yes	No
5.	Bid validity time is as per demand	Yes	Yes
6.	Warranty Form as per Tender	Yes	Yes
7.	Delivery as per Schedule of Requirement	Yes	Yes
8.	Undertaking on legal valid and attested stamp paper Rs. 100/- that the firm is not blacklisted by any of Provincial or Federal Government Department, Agency, Organization or autonomous body or Private Sector Organization anywhere in Pakistan.	No	Yes
9.	Embassy Attested Exclusively Authorized Agent / Sole Distributor Certificate by the Original Equipment Manufacturer (OEM).	No (Embassy attested certificate not provided)	No (Embassy attested certificate not provided)
10.	Sub-Distributor Certificate from Sole Distributor having consent of OEM. (Where Applicable)	N/A	N/A
11.	Proof for having minimum 5 years of experience in supply, installation, commissioning of biomedical equipment.	Yes	Yes
12.	At least 05 verifiable Past Performance Certificates by the Procuring Agency of quoted products issued against the same Supply Orders.	No	No
13.	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes	Yes
14.	Originally signed documentary evidence from the Manufacturer that the quoted brand is the Original Equipment Manufacturer (OEM) with indication of manufacturing site and its location.	Yes	Yes
15.	Up to Date After Sales Service Certificate from OEM/Principal	Yes	Yes
16.	Availability of relevant Tools and Testing / Calibration Equipment.	Yes	Yes
17.	A Certificate from the OEM that the installation will be conducted in conformity with the system requirements.	Yes	Yes

102



18.	Consortiums or Joint Ventures (JVs) (Where applicable)	N/A	N/A
19.	Complete Package / Tender	No	No
Remarks:		Not Eligible	Not Eligible
<b>PART – II</b> <b>KNOCKED DOWN CRITERIA</b> <b>(PRODUCT EVALUATION)</b>			
Item Ad. No.	Elisa System	M/s Sahaab International.	M/s Irfan Brothers Intl.
	Brand	Medallion Equipment	TKA Teknolabo A.S.S.I. Srl
45	Quoted Model	MES 2000	Bellini
1.	Country of Manufacturer	UK	Italy
2.	Country of Origin of Product	UK	Italy
3.	Compliance with defined Quality Standards	(Non Complied) (Certificate could not verified online) and (CE (MDD) from authorized EU body not available.)	(Non Complied) No (CE (MDD) Third Party Certificate is not Attached)
4.	Specifications Compliance Feature Wise	(Non Complied)  The required specifications were copied on product catalog attached to the technical offer; moreover, the product information also could not be verified from the manufacturer's website.	(Non Complied)  LAN/ WLAN port to share the data with hospital network cannot be verified from product data sheet.
5.	Technical Eligibility of Product	Not Eligible	Not Eligible
6.	Technical Eligibility of Firm	Not Eligible	Not Eligible
Technical Status:		<b>NON RESPONSIVE</b>	<b>NON RESPONSIVE</b>

Item No. 46  
Item Name: ESR Machine  
Tender Name: Procurement of Bio-Medical Equipment Under PSDP Project  
"Establishment of 50 Bedded Cardiac Hospital, Gilgit".

PART – I KNOCKED DOWN CRITERIA (BID EVALUATION & VENDOR EVALUATION)		
Sr. #	Evaluation Parameters	M/s Sahaab International Pvt. Ltd.
1.	Bid Form	Yes
2.	Bidder's Information Form	Yes
3.	Documentary proof active of National Tax Number (NTN) and General Sales Tax (GST) Number and Registration Professional Tax Certificate.	Yes
4.	Original Bid Security (In form of CDR) & Form	Yes
5.	Bid validity time is as per demand	Yes
6.	Warranty Form as per Tender	Yes
7.	Delivery as per Schedule of Requirement	Yes
8.	Undertaking on legal valid and attested stamp paper Rs. 100/- that the firm is not blacklisted by any of Provincial or Federal Government Department, Agency, Organization or autonomous body or Private Sector Organization anywhere in Pakistan.	No
9.	Embassy Attested Exclusively Authorized Agent / Sole Distributor Certificate by the Original Equipment Manufacturer (OEM).	No (Embassy attested certificate not provided)
10.	Sub-Distributor Certificate from Sole Distributor having consent of OEM. (Where Applicable)	N/A
11.	Proof for having minimum 5 years of experience in supply, installation, commissioning of biomedical equipment.	Yes
12.	At least 05 verifiable Past Performance Certificates by the Procuring Agency of quoted products issued against the same Supply Orders.	No
13.	Technical & Engineering capability (As defined for the specific Bid in specifications)	No
14.	Originally signed documentary evidence from the Manufacturer that the quoted brand is the Original Equipment Manufacturer (OEM) with indication of manufacturing site and its location.	No
15.	Up to Date After Sales Service Certificate from OEM/Principal	No
16.	Availability of relevant Tools and Testing / Calibration Equipment.	No
17.	A Certificate from the OEM that the installation will be conducted in conformity with the system requirements.	No
18.	Consortiums or Joint Ventures (JVs) (Where applicable)	N/A
19.	Complete Package / Tender	No
Remarks:		Not Eligible
PART – II		



KNOCKED DOWN CRITERIA (PRODUCT EVALUATION)		
Item Ad. No.	ESR Machine	M/s Sahaab International Pvt. Ltd.
		Medallion Equipment
46	Quoted Model	MAB-140
1.	Country of Manufacturer	UK
2.	Country of Origin of Product	UK
3.	Compliance with defined Quality Standards	No (Non Complied)
4.	Specifications Compliance Feature Wise	(Non Complied)  The required specifications were copied on product catalog attached to the technical offer; moreover, the product information also could not be verified from the manufacturer's website.
5.	Technical Eligibility of Product	Not Eligible
6.	Technical Eligibility of Firm	Not Eligible
Technical Status:		NON RESPONSIVE

Handwritten signatures and initials in blue ink, including a large signature on the left, a checkmark, and several other marks.

Item No. 47  
Item Name: Cell Separator  
Tender Name: Procurement of Bio-Medical Equipment Under PSDP Project  
"Establishment of 50 Bedded Cardiac Hospital, Gilgit".

PART – I KNOCKED DOWN CRITERIA (BID EVALUATION & VENDOR EVALUATION)		
Sr. #	Evaluation Parameters	M/s Hoora Pharma
1.	Bid Form	No (Same Bid Form for Bid No. 41, 44, 47, 48, 51.)
2.	Bidder's Information Form	Yes
3.	Documentary proof active of National Tax Number (NTN) and General Sales Tax (GST) Number and Registration Professional Tax Certificate.	Yes
4.	Original Bid Security (In form of CDR) & Form	No (Not Attached with Technical Offer)
5.	Bid validity time is as per demand	Yes
6.	Warranty Form as per Tender	Yes
7.	Delivery as per Schedule of Requirement	Yes
8.	Undertaking on legal valid and attested stamp paper Rs. 100/- that the firm is not blacklisted by any of Provincial or Federal Government Department, Agency, Organization or autonomous body or Private Sector Organization anywhere in Pakistan.	No
9.	Embassy Attested Exclusively Authorized Agent / Sole Distributor Certificate by the Original Equipment Manufacturer (OEM).	No
10.	Sub-Distributor Certificate from Sole Distributor having consent of OEM. (Where Applicable)	No
11.	Proof for having minimum 5 years of experience in supply, installation, commissioning of biomedical equipment.	Yes
12.	At least 05 verifiable Past Performance Certificates by the Procuring Agency of quoted products issued against the same Supply Orders.	No
13.	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes
14.	Originally signed documentary evidence from the Manufacturer that the quoted brand is the Original Equipment Manufacturer (OEM) with indication of manufacturing site and its location.	No
15.	Up to Date After Sales Service Certificate from OEM/Principal	No
16.	Availability of relevant Tools and Testing / Calibration Equipment.	Yes
17.	A Certificate from the OEM that the installation will be conducted in conformity with the system requirements.	No
18.	Consortiums or Joint Ventures (JVs) (Where applicable)	N/A
19.	Complete Package / Tender	No
Remarks:		Not Eligible



PART – II KNOCKED DOWN CRITERIA (PRODUCT EVALUATION)		
Item Ad. No.	Cell Separator	M/s Hoora Pharma
	Brand	Terumo-BCT
47	Quoted Model	Spectra Optia
1.	Country of Manufacturer	USA
2.	Country of Origin of Product	USA
3.	Compliance with defined Quality Standards	(Compliance) CE MDD & FDA10K (Non Complied)
4.	Specifications Compliance Feature Wise	8. Centrifuge Speed: 3000-7000 RPM instead 500-3000 RPM. 9. Working on open Apheresis kits not available. 10. Continuous Flow instead of Discontinuous Flow. 11. Remaining parameters cannot be verified from the attached catalogue.
5.	Technical Eligibility of Product	Not Eligible
6.	Technical Eligibility of Firm	Not Eligible
Technical Status:		NON RESPONSIVE

Item No. 48  
Item Name: Chemistry Analyzer  
Tender Name: Procurement of Bio-Medical Equipment Under PSDP Project "Establishment of 50 Bedded Cardiac Hospital, Gilgit".

PART – I KNOCKED DOWN CRITERIA (BID EVALUATION & VENDOR EVALUATION)				
Sr. #	Evaluation Parameters	M/s Sahaab International	M/s Irfan Brothers International	M/s Hoora Pharma
1.	Bid Form	Yes	Yes	No (Same Bid Form for Bid No. 41, 44, 47, 48, 51.)
2.	Bidder's Information Form	Yes	Yes	Yes
3.	Documentary proof active of National Tax Number (NTN) and General Sales Tax (GST) Number and Registration Professional Tax Certificate.	Yes	Yes	Yes
4.	Original Bid Security (In form of CDR) & Form	Yes	No	No Not Attached with Technical Offer
5.	Bid validity time is as per demand	Yes	Yes	Yes
6.	Warranty Form as per Tender	Yes	Yes	Yes
7.	Delivery as per Schedule of Requirement	Yes	Yes	Yes
8.	Undertaking on legal valid and attested stamp paper Rs. 100/- that the firm is not blacklisted by any of Provincial or Federal Government Department, Agency, Organization or autonomous body or Private Sector Organization anywhere in Pakistan.	No	Yes	Yes
9.	Embassy Attested Exclusively Authorized Agent / Sole Distributor Certificate by the Original Equipment Manufacturer (OEM).	No (Embassy attested certificate not provided)	Yes (M/s Human Healthcare)	Yes
10.	Sub-Distributor Certificate from Sole Distributor having consent of OEM. (Where Applicable)	N/A	No	N/A
11.	Proof for having minimum 5 years of experience in supply, installation, commissioning of biomedical equipment.	Yes	Yes	Yes



12.	At least 05 verifiable Past Performance Certificates by the Procuring Agency of quoted products issued against the same Supply Orders.	No	Yes	No
13.	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes	Yes	Yes
14.	Originally signed documentary evidence from the Manufacturer that the quoted brand is the Original Equipment Manufacturer (OEM) with indication of manufacturing site and its location.	Yes	Yes	No
15.	Up to Date After Sales Service Certificate from OEM/Principal	Yes	Yes	Yes
16.	Availability of relevant Tools and Testing / Calibration Equipment.	Yes	Yes	Yes
17.	A Certificate from the OEM that the installation will be conducted in conformity with the system requirements.	Yes	Yes	Yes
18.	Consortiums or Joint Ventures (JVs) (Where applicable)	N/A	N/A	N/A
19.	Complete Package / Tender	No	No Bid Pages are not signed and stamped as required in BD ITB Clause 19.	No
Remarks:		Not Eligible	Not Eligible	Not Eligible
<b>PART – II</b> <b>KNOCKED DOWN CRITERIA</b> <b>(PRODUCT EVALUATION)</b>				
Item Ad. No.	Chemistry Analyzer	M/s Sahaab International	M/s Irfan Brothers International	M/s Hoora Pharma
	Brand	Medallion Equipment	Vitanix D.O.O	Siemens Healthineers
48	Quoted Model	MCS-480	Vitachem 200	EXL 200
1.	Country of Manufacturer	UK	Slovenia	USA
2.	Country of Origin of Product	Not Mentioned	Slovenia	USA
3.	Compliance with defined Quality Standards	(Non Complied) (Expired Certificate and (CE (MDD) from	(Non Complied) CE (MDD) certificate not from authorized EU body is attached.	(Compliance) CE (MDD) & FDA CFG

		authorized EU body not available.) (Non Complied)		
4.	Specifications Compliance Feature Wise	1- The required specifications were copied on product catalog attached to the technical offer; moreover, the product information also could not be verified from the manufacturer's website. 2- Capability to re-run with automatic on-board sample dilution not available. 3- Automatic Sample Quality Analysis for Lipemia, Hemolysis, Icterus not available. 4- 400 tests/hr or above for a range of chemistries (Other than ISE i.e. Na <sup>+</sup> , K <sup>+</sup> Cl <sup>-</sup> ) 5- Compatible RO System and UPS not quoted.	(Non Complied) 1. Spectral Range 340-800 instead of 320-690 nm. 2. Touch Screen Display Monitor not available.	Compliance




5.	Technical Eligibility of Product	Not Eligible	Not Eligible	Eligible
6.	Technical Eligibility of Firm	Not Eligible	Not Eligible	Not Eligible
<b>Technical Status:</b>		<b>NON RESPONSIVE</b>	<b>NON RESPONSIVE</b>	<b>NON RESPONSIVE</b>

Handwritten signatures and initials in blue ink, including a large signature at the top left, a signature in the center, and several smaller initials and marks scattered around.

Item No.  
Item Name:  
Tender Name:

49  
Coagulometer  
Procurement of Bio-Medical Equipment Under PSDP Project  
"Establishment of 50 Bedded Cardiac Hospital, Gilgit".

PART – I KNOCKED DOWN CRITERIA (BID EVALUATION & VENDOR EVALUATION)			
Sr. #	Evaluation Parameters	M/s Sahaab International.	M/s Irfan Brothers International
1.	Bid Form	Yes	Yes
2.	Bidder's Information Form	Yes	Yes
3.	Documentary proof active of National Tax Number (NTN) and General Sales Tax (GST) Number and Registration Professional Tax Certificate.	Yes	Yes
4.	Original Bid Security (In form of CDR) & Form	Yes	No
5.	Bid validity time is as per demand	Yes	Yes
6.	Warranty Form as per Tender	Yes	Yes
7.	Delivery as per Schedule of Requirement	Yes	Yes
8.	Undertaking on legal valid and attested stamp paper Rs. 100/- that the firm is not blacklisted by any of Provincial or Federal Government Department, Agency, Organization or autonomous body or Private Sector Organization anywhere in Pakistan.	No	Yes
9.	Embassy Attested Exclusively Authorized Agent / Sole Distributor Certificate by the Original Equipment Manufacturer (OEM).	No (Embassy attested certificate not provided)	No (Embassy attested certificate not provided)
10.	Sub-Distributor Certificate from Sole Distributor having consent of OEM. (Where Applicable)	N/A	N/A
11.	Proof for having minimum 5 years of experience in supply, installation, commissioning of biomedical equipment.	Yes	Yes
12.	At least 05 verifiable Past Performance Certificates by the Procuring Agency of quoted products issued against the same Supply Orders.	No	No
13.	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes	Yes
14.	Originally signed documentary evidence from the Manufacturer that the quoted brand is the Original Equipment Manufacturer (OEM) with indication of manufacturing site and its location.	Yes	Yes
15.	Up to Date After Sales Service Certificate from OEM/Principal	Yes	Yes
16.	Availability of relevant Tools and Testing / Calibration Equipment.	Yes	Yes
17.	A Certificate from the OEM that the installation will be conducted in conformity with the system requirements.	Yes	Yes





18.	Consortiums or Joint Ventures (JVs) (Where applicable)	N/A	N/A
19.	Complete Package / Tender	No Bid Pages are not signed and stamped as required in BD ITB Clause 19.	No Bid Pages are not signed and stamped as required in BD ITB Clause 19.
Remarks:		Not Eligible	Not Eligible
<b>PART – II</b> <b>KNOCKED DOWN CRITERIA</b> <b>(PRODUCT EVALUATION)</b>			
Item Ad. No.	Coagulometer	M/s Sahaab International.	M/s Irfan Brothers International
	Brand	Medallion Equipment	LABITEC
49	Quoted Model	MAC-050	COALAB 1000
1.	Country of Manufacturer	UK	Germany
2.	Country of Origin of Product	Not Mentioned	Germany
3.	Compliance with defined Quality Standards	(Non Complied) (Expired Certificate and (CE (MDD) from authorized EU body not available.)	(Non Complied) CE (MDD) certificate not from authorized EU body is attached.
4.	Specifications Compliance Feature Wise	(Non Complied)  The quoted model is not available on official website of quoted manufacturer and the required specifications were copied on product catalog attached to the technical offer.	(Non Complied)  1. On Board storage capacity with separate positions for STAT not verified on product data sheet online.
5.	Technical Eligibility of Product	Not Eligible	Not Eligible
6.	Technical Eligibility of Firm	Not Eligible	Not Eligible
Technical Status:		NON RESPONSIVE	NON RESPONSIVE

Item No. 50  
Item Name: Spectrophotometer (Semi Auto Chem)  
Tender Name: Procurement of Bio-Medical Equipment Under PSDP Project  
"Establishment of 50 Bedded Cardiac Hospital, Gilgit".

PART – I KNOCKED DOWN CRITERIA (BID EVALUATION & VENDOR EVALUATION)			
Sr. #	Evaluation Parameters	M/s Sahaab International.	M/s Irfan Brothers International
1.	Bid Form	Yes	Yes
2.	Bidder's Information Form	Yes	Yes
3.	Documentary proof active of National Tax Number (NTN) and General Sales Tax (GST) Number and Registration Professional Tax Certificate.	Yes	Yes
4.	Original Bid Security (In form of CDR) & Form	Yes	No
5.	Bid validity time is as per demand	Yes	Yes
6.	Warranty Form as per Tender	Yes	Yes
7.	Delivery as per Schedule of Requirement	Yes	Yes
8.	Undertaking on legal valid and attested stamp paper Rs. 100/- that the firm is not blacklisted by any of Provincial or Federal Government Department, Agency, Organization or autonomous body or Private Sector Organization anywhere in Pakistan.	No	Yes
9.	Embassy Attested Exclusively Authorized Agent / Sole Distributor Certificate by the Original Equipment Manufacturer (OEM).	No (Embassy attested certificate not provided)	No (Embassy attested certificate not provided)
10.	Sub-Distributor Certificate from Sole Distributor having consent of OEM. (Where Applicable)	N/A	N/A
11.	Proof for having minimum 5 years of experience in supply, installation, commissioning of biomedical equipment.	Yes	Yes
12.	At least 05 verifiable Past Performance Certificates by the Procuring Agency of quoted products issued against the same Supply Orders.	No	Yes
13.	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes	Yes
14.	Originally signed documentary evidence from the Manufacturer that the quoted brand is the Original Equipment Manufacturer (OEM) with indication of manufacturing site and its location.	Yes	Yes
15.	Up to Date After Sales Service Certificate from OEM/Principal	Yes	Yes
16.	Availability of relevant Tools and Testing / Calibration Equipment.	Yes	Yes
17.	A Certificate from the OEM that the installation will be conducted in conformity with the system requirements.	Yes	Yes
18.	Consortiums or Joint Ventures (JVs) (Where applicable)	N/A	N/A



19.	Complete Package / Tender	No	No Bid Pages are not signed and stamped as required in BD ITB Clause 19.
Remarks:		Not Eligible	Not Eligible
PART – II KNOCKED DOWN CRITERIA (PRODUCT EVALUATION)			
Item Ad. No.	Spectrophotometer (Semi Auto Chem)	M/s Sahaab International.	M/s Irfan Brothers International
	Brand	Medallion Equipment	Sanymed
50	Quoted Model	Chemaxcel	SA-200
1.	Country of Manufacturer	UK	Italy
2.	Country of Origin of Product	Not Mentioned	Italy
3.	Compliance with defined Quality Standards	(Non Complied) (Expired Certificate and (CE (MDD) from authorized EU body not available.)	(Non Complied) CE (MDD) certificate not from authorized EU body is attached.
4.	Specifications Compliance Feature Wise	(Non Complied)  The required specifications were copied on product catalog attached to the technical offer; moreover, the product information also could not be verified from the manufacturer's website.	(Non Complied)  2. Freely programmable against all commercially available chemistries not available.
5.	Technical Eligibility of Product	Not Eligible	Not Eligible
6.	Technical Eligibility of Firm	Not Eligible	Not Eligible
Technical Status:		NON RESPONSIVE	NON RESPONSIVE

Item No. 51  
Item Name: Hematology Analyzer  
Tender Name: Procurement of Bio-Medical Equipment Under PSDP Project "Establishment of 50 Bedded Cardiac Hospital, Gilgit".

PART - I KNOCKED DOWN CRITERIA (BID EVALUATION & VENDOR EVALUATION)				
Sr. #	Evaluation Parameters	M/s Irfan Brothers International	M/s Hoora Pharma	M/s Sahaab International
1.	Bid Form	Yes	No (Same Bid Form for Bid No. 41, 44, 47, 48, 51.)	Yes
2.	Bidder's Information Form	Yes	Yes	Yes
3.	Documentary proof active of National Tax Number (NTN) and General Sales Tax (GST) Number and Registration Professional Tax Certificate.	Yes	Yes	Yes
4.	Original Bid Security (In form of CDR) & Form	Yes	No (Not Attached with Technical Offer)	No
5.	Bid validity time is as per demand	Yes	Yes	Yes
6.	Warranty Form as per Tender	Yes	Yes	Yes
7.	Delivery as per Schedule of Requirement	Yes	Yes	Yes
8.	Undertaking on legal valid and attested stamp paper Rs. 100/- that the firm is not blacklisted by any of Provincial or Federal Government Department, Agency, Organization or autonomous body or Private Sector Organization anywhere in Pakistan.	Yes	Yes	No
9.	Embassy Attested Exclusively Authorized Agent / Sole Distributor Certificate by the Original Equipment Manufacturer (OEM).	Yes (M/s Human Healthcare)	Yes	No (Embassy attested certificate not provided)
10.	Sub-Distributor Certificate from Sole Distributor having consent of OEM. (Where Applicable)	Yes	N/A	N/A

116



11.	Proof for having minimum 5 years of experience in supply, installation, commissioning of biomedical equipment.	Yes	Yes	Yes
12.	At least 05 verifiable Past Performance Certificates by the Procuring Agency of quoted products issued against the same Supply Orders.	Yes	Yes	No
13.	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes	Yes	Yes
14.	Originally signed documentary evidence from the Manufacturer that the quoted brand is the Original Equipment Manufacturer (OEM) with indication of manufacturing site and its location.	Yes	No	Yes
15.	Up to Date After Sales Service Certificate from OEM/Principal	Yes	Yes	Yes
16.	Availability of relevant Tools and Testing / Calibration Equipment.	Yes	Yes	Yes
17.	A Certificate from the OEM that the installation will be conducted in conformity with the system requirements.	Yes	Yes	Yes
18.	Consortiums or Joint Ventures (JVs) (Where applicable)	N/A	N/A	N/A
19.	Complete Package / Tender	Yes	No	No
Remarks:		Eligible	Not Eligible	Not Eligible
<b>PART – II</b> <b>KNOCKED DOWN CRITERIA</b> <b>(PRODUCT EVALUATION)</b>				
Item Ad. No.	Hematology Analyzer	M/s Irfan Brothers International	M/s Hoora Pharma	M/s Sahaab International
	Brand	Vitanix D.O.O	Siemens Healthineer	Medallion Equipment
51	Quoted Model	Vitacount 5P	Advia 560	MF-9600
1.	Country of Manufacturer	Slovenia	USA	UK
2.	Country of Origin of Product	Slovenia	USA/EU	UK
3.	Compliance with defined Quality Standards	(Non Complied)	(Compliance) CE (MDD)	(Non Complied)  (Expired Certificate and (CE (MDD) from

		CE (MDD) certificate not from authorized EU body is attached.		authorized EU body not available.)
4.	Specifications Compliance Feature Wise	(Non Complied) 1. LCD display with touch screen not available.	(Non Complied) 1. The offered through put 60 samples/hour instead of 80. 2. Difference among the data pages attached with the technical bid, as on a page is mentioned that equipment measure 20 parameters 5-part blood cell differential while on ther page it is 26 and the demand was of Minimum 24 parameters 5-Part differential Hb, WBC, RBC, HCT, PLT, and derived values.	(Non Complied) 1. The required specifications were copied on product catalog attached to the technical offer; moreover, the product information also could not be verified from the manufacturer's website. 2. Automated loader quoted as optional. 3. The quoted model is not available on official website of quoted manufacturer. 4. Specifications were copied on product catalog attached to the technical offer; moreover, the product information also



				could not be verified from the manufacturer's website.
5.	Technical Eligibility of Product	Not Eligible	Not Eligible	Not Eligible
6.	Technical Eligibility of Firm	Eligible	Not Eligible	Not Eligible
<b>Technical Status:</b>		<b>NON RESPONSIVE</b>	<b>NON RESPONSIVE</b>	<b>NON RESPONSIVE</b>

Item No. 52  
Item Name: Table Top Autoclave 22L  
Tender Name: Procurement of Bio-Medical Equipment Under PSDP Project "Establishment of 50 Bedded Cardiac Hospital, Gilgit".

PART – I KNOCKED DOWN CRITERIA (BID EVALUATION & VENDOR EVALUATION)				
Sr. #	Evaluation Parameters	M/s Total Technologies Pvt. Ltd.	M/s Zodiac International	M/s Sahaab International
1.	Bid Form	Yes	Yes	Yes
2.	Bidder's Information Form	Yes	Yes	Yes
3.	Documentary proof active of National Tax Number (NTN) and General Sales Tax (GST) Number and Registration Professional Tax Certificate.	Yes	Yes	Yes
4.	Original Bid Security (In form of CDR) & Form	Yes	No (Less Amount CDR)	No
5.	Bid validity time is as per demand	Yes	Yes	Yes
6.	Warranty Form as per Tender	Yes	No (Conditional Warranty)	Yes
7.	Delivery as per Schedule of Requirement	Yes	Yes	Yes
8.	Undertaking on legal valid and attested stamp paper Rs. 100/- that the firm is not blacklisted by any of Provincial or Federal Government Department, Agency, Organization or autonomous body or Private Sector Organization anywhere in Pakistan.	Yes	No (Photocopy attached)	Yes
9.	Embassy Attested Exclusively Authorized Agent / Sole Distributor Certificate by the Original Equipment Manufacturer (OEM).	Yes	No (Embassy attested certificate not provided)	No (Embassy attested certificate not provided)
10.	Sub-Distributor Certificate from Sole Distributor having consent of OEM. (Where Applicable)	N/A	N/A	N/A
11.	Proof for having minimum 5 years of experience in supply, installation, commissioning of biomedical equipment.	Yes	Yes	Yes



12.	At least 05 verifiable Past Performance Certificates by the Procuring Agency of quoted products issued against the same Supply Orders.	Yes	Yes	No
13.	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes	Yes	Yes
14.	Originally signed documentary evidence from the Manufacturer that the quoted brand is the Original Equipment Manufacturer (OEM) with indication of manufacturing site and its location.	Yes	No	Yes
15.	Up to Date After Sales Service Certificate from OEM/Principal	Yes	Yes	Yes
16.	Availability of relevant Tools and Testing / Calibration Equipment.	Yes	Yes	Yes
17.	A Certificate from the OEM that the installation will be conducted in conformity with the system requirements.	Yes	Yes	Yes
18.	Consortiums or Joint Ventures (JVs) (Where applicable)	N/A	N/A	N/A
19.	Complete Package / Tender	Yes	No	No
Remarks:		Eligible	Not Eligible	Not Eligible
<b>PART – II</b> <b>KNOCKED DOWN CRITERIA</b> <b>(PRODUCT EVALUATION)</b>				
Item Ad. No.	Table Top Autoclave 22L	M/s Total Technologies Pvt. Ltd.	M/s Zodiac International	M/s Sahaab International
	Brand	MELAG	Euronda	Medallion Equipment
52	Quoted Model	Vacuklave 43B+ Evolution	E8 Bultin Printer	MTA-B23
1.	Country of Manufacturer	Germany	Italy	UK
2.	Country of Origin of Product	Germany	Italy	UK
3.	Compliance with defined Quality Standards	(Compliance) CE (MDD)	(Compliance) CE (MDD)	(Non Complied)  (Expired Certificate and (CE (MDD) from authorized EU body not available.)

4.	Specifications Compliance Feature Wise	Compliance	Compliance	(Non Complied)  The required specifications were copied on product catalog attached to the technical offer; moreover, the product information also could not be verified from the manufacturer's website.
5.	Technical Eligibility of Product	Eligible	Eligible	Not Eligible
6.	Technical Eligibility of Firm	Eligible	Not Eligible	Not Eligible
<b>Technical Status:</b>		<b>RESPONSIVE</b>	<b>NON RESPONSIVE</b>	<b>NON RESPONSIVE</b>



Item No.  
Item Name:  
Tender Name:

53  
Laboratory Sterilizer  
Procurement of Bio-Medical Equipment Under PSDP Project  
"Establishment of 50 Bedded Cardiac Hospital, Gilgit".

PART – I KNOCKED DOWN CRITERIA (BID EVALUATION & VENDOR EVALUATION)			
Sr. #	Evaluation Parameters	M/s Sahaab International.	M/s Zodiac International
1.	Bid Form	Yes	Yes
2.	Bidder's Information Form	Yes	Yes
3.	Documentary proof active of National Tax Number (NTN) and General Sales Tax (GST) Number and Registration Professional Tax Certificate.	Yes	Yes
4.	Original Bid Security (In form of CDR) & Form	Yes	No (Less Amount CDR)
5.	Bid validity time is as per demand	Yes	Yes
6.	Warranty Form as per Tender	Yes	No (Conditional Warranty)
7.	Delivery as per Schedule of Requirement	Yes	Yes
8.	Undertaking on legal valid and attested stamp paper Rs. 100/- that the firm is not blacklisted by any of Provincial or Federal Government Department, Agency, Organization or autonomous body or Private Sector Organization anywhere in Pakistan.	No	No (Photocopy attached)
9.	Embassy Attested Exclusively Authorized Agent / Sole Distributor Certificate by the Original Equipment Manufacturer (OEM).	No (Embassy attested certificate not provided)	No (Embassy attested certificate not provided)
10.	Sub-Distributor Certificate from Sole Distributor having consent of OEM. (Where Applicable)	N/A	N/A
11.	Proof for having minimum 5 years of experience in supply, installation, commissioning of biomedical equipment.	Yes	Yes
12.	At least 05 verifiable Past Performance Certificates by the Procuring Agency of quoted products issued against the same Supply Orders.	No	No
13.	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes	Yes
14.	Originally signed documentary evidence from the Manufacturer that the quoted brand is the Original Equipment Manufacturer (OEM) with indication of manufacturing site and its location.	Yes	No
15.	Up to Date After Sales Service Certificate from OEM/Principal	Yes	Yes
16.	Availability of relevant Tools and Testing / Calibration Equipment.	Yes	Yes

17.	A Certificate from the OEM that the installation will be conducted in conformity with the system requirements.	Yes	Yes
18.	Consortiums or Joint Ventures (JVs) (Where applicable)	N/A	N/A
19.	Complete Package / Tender	No Bid Pages are not signed and stamped as required in BD ITB Clause 19.	No
Remarks:		Not Eligible	Not Eligible
<b>PART – II</b> <b>KNOCKED DOWN CRITERIA</b> <b>(PRODUCT EVALUATION)</b>			
Item Ad. No.	Laboratory Sterilizer	M/s Sahaab International.	M/s Irfan Brothers International
	Brand	Medallion Equipment	Summer
53	Quoted Model	MVA-15	ADELA 115
1.	Country of Manufacturer	UK	Turkey
2.	Country of Origin of Product	Not Mentioned	Not Mentioned
3.	Compliance with defined Quality Standards	(Non Complied) CE (MDD) certificate not from authorized EU body is attached.	(Non Complied) CE (MDD) certificate not from authorized EU body is attached.
4.	Specifications Compliance Feature Wise	(Non Complied)  The quoted model is not available on official website of quoted manufacturer and the required specifications were copied on product catalog attached to the technical offer	(Non Complied)  3. The country of manufacturer and origin is not as per requirement.
5.	Technical Eligibility of Product	Not Eligible	Not Eligible
6.	Technical Eligibility of Firm	Not Eligible	Not Eligible
Technical Status:		NON RESPONSIVE	NON RESPONSIVE



Item No. 54  
Item Name: CSSD  
Tender Name: Procurement of Bio-Medical Equipment Under PSDP Project  
"Establishment of 50 Bedded Cardiac Hospital, Gilgit".

PART – I KNOCKED DOWN CRITERIA (BID EVALUATION & VENDOR EVALUATION)			
Sr. #	Evaluation Parameters	M/s Mediland Pakistan Pvt. Ltd.	M/s Total Technologies Pvt. Ltd.
1.	Bid Form	Yes	Yes
2.	Bidder's Information Form	Yes	Yes
3.	Documentary proof active of National Tax Number (NTN) and General Sales Tax (GST) Number and Registration Professional Tax Certificate.	Yes	Yes
4.	Original Bid Security (In form of CDR) & Form	Yes	Yes
5.	Bid validity time is as per demand	Yes	Yes
6.	Warranty Form as per Tender	Yes	Yes
7.	Delivery as per Schedule of Requirement	Yes	Yes
8.	Undertaking on legal valid and attested stamp paper Rs. 100/- that the firm is not blacklisted by any of Provincial or Federal Government Department, Agency, Organization or autonomous body or Private Sector Organization anywhere in Pakistan.	Yes	Yes
9.	Embassy Attested Exclusively Authorized Agent / Sole Distributor Certificate by the Original Equipment Manufacturer (OEM).	Yes	Yes
10.	Sub-Distributor Certificate from Sole Distributor having consent of OEM. (Where Applicable)	N/A	N/A
11.	Proof for having minimum 5 years of experience in supply, installation, commissioning of biomedical equipment.	Yes	Yes
12.	At least 05 verifiable Past Performance Certificates by the Procuring Agency of quoted products issued against the same Supply Orders.	Yes	Yes
13.	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes	Yes
14.	Originally signed documentary evidence from the Manufacturer that the quoted brand is the Original Equipment Manufacturer (OEM) with indication of manufacturing site and its location.	Yes	Yes
15.	Up to Date After Sales Service Certificate from OEM/Principal	Yes	Yes
16.	Availability of relevant Tools and Testing / Calibration Equipment.	Yes	Yes
17.	A Certificate from the OEM that the installation will be conducted in conformity with the system requirements.	Yes	Yes
18.	Consortiums or Joint Ventures (JVs) (Where applicable)	N/A	N/A
19.	Complete Package / Tender	Yes	Yes

Remarks:		Eligible	Eligible
PART – II KNOCKED DOWN CRITERIA (PRODUCT EVALUATION)			
Item Ad. No.	CSSD	M/s Mediland Pakistan Pvt. Ltd.	M/s Total Technologies Pvt. Ltd.
	Brand		
54	Quoted Model	As per Compliance Sheet	As per Compliance Sheet
1.	Country of Manufacturer		
2.	Country of Origin of Product		
3.	Compliance with defined Quality Standards	(Compliance) CE/ISO	(Compliance) CE/ISO
4.	Specifications Compliance Feature Wise	Compliance	Compliance
5.	Technical Eligibility of Product	Eligible	Eligible
6.	Technical Eligibility of Firm	Eligible	Eligible
Technical Status:		RESPONSIVE	RESPONSIVE

COMPLIANCE SHEET			
No.	Parameters	M/s Radiant Medical Pvt. Ltd.	M/s Total Technologies Pvt. Ltd.
1	Cleaning & Washing Unit	Model: Hupfer Germany	Model: 10547162+10547186+ 10547187+010547180 MMM Group Germany
2	High End Washing /Disinfection / Drying Unit	Model: S8668 Getinge AB Sweden	Model: Uniclean PL- II15-2 MMM Group Germany
3	Heavy Duty Steam Sterilizer	Model: GSS67H13 Getinge AB Sweden	Model: Selectomat PL 6612-2 CL MMM Group Germany
4	Heavy Duty Steam Sterilizer	Model: GSS67H0 Getinge AB Sweden	Model: Selectomat PL 666-2 CL MMM Group Germany
5	Ultrasonic Cleaner	Model: Elamsonic Med 500 Getinge AB Sweden	Model: Sonica 5300 EPI Soltec Italy
6	Oil Free Air Compressor	Model: 600300382 Getinge AB Sweden	Model: Air Med 258-50 FIAC Italy
7	Reverse Osmosis System	Local	Local
8	Plasma Sterilizer	Model: Poladas Getinge AB Sweden	Model: Hyper LTS 150- 2 MMM Group Germany
9	Sterilant	Quoted	Quoted
10	Transport And Distribution Trolley	Model: Hupfer Germany	Model: 10032061 MMM Group Germany/Italy
11	Paper Dispensing Trolley	Model: Hupfer Germany	Model: 000154271+010002646



			MMM Group Germany/Italy
12	<b>Paper Sealing Machine</b>	Model: Pro Seal Premium Getting AB Germany	Model: HM950DC-V Hawo Germany
13	<b>Cutting Device</b>	Model: L=600 mm Getting AB Germany	Model: HM631S Hawo Germany
14	<b>Tape Dispenser</b>	Model: 561570701 Getting AB Germany	Model: 10585106/Interlock MMM Group Germany
15	<b>Basket Trolley</b>	Model: Hupfer Germany	Model: 010546762 MMM Germany /Italy
16	<b>Free Standing Baskets Rack</b>	Model: Hupfer Germany	Model: 10546912 MMM Germany/Italy
17	<b>Packing Table</b>	Model: Hupfer Germany	Model: 0105547+010537077 MMM Germany/Italy
18	<b>Linen Table</b>	Model: Hupfer Germany	Model: 10546970 MMM Germany/Italy
19	<b>Modular Wire Basket</b>	Model: 5618636 Germany/EU	Model: 164580 MMM Germany/Italy
20	<b>Closed Transport Trolley</b>	Model: Hupfer Germany	Model: 10082747 MMM Germany/Italy
21	<b>Clean Room Sterilization System</b>	Air Purification Unit Pro 130 Mediclean Air Belgium Supplied by Getinge	Model: Air Manager EOLIS 1200S Naterosante France
22	<b>Provision To Integrate with Documentation Software (Optional)</b>	Quoted	Quoted

Item No. 55

Item Name: Hospital Laundry

Tender Name: Procurement Of Bio-Medical Equipment Under PSDP Project "Establishment of 50 Bedded Cardiac Hospital, Gilgit".

PART - I KNOCKED DOWN CRITERIA (BID EVALUATION & VENDOR EVALUATION)					
Sr. #	Evaluation Parameters	M/s Total Technologies Pvt. Ltd.	M/s Pakistan Oxygen Limited	M/s Mediland Pakistan Pvt. Ltd.	M/s Med Engineering
1.	Bid Form	Yes	Yes	Yes	Yes
2.	Bidder's Information Form	Yes	Yes	Yes	Yes
3.	Documentary proof active of National Tax Number (NTN) and General Sales Tax (GST) Number and Registration Professional Tax Certificate.	Yes	Yes	Yes	Yes
4.	Original Bid Security (In form of CDR) & Form	Yes	Yes	Yes	No (Original Copy not attached)
5.	Bid validity time is as per demand	Yes	Yes	Yes	Yes
6.	Warranty Form as per Tender	Yes	Yes	Yes	Yes
7.	Delivery as per Schedule of Requirement		Yes	Yes	Yes
8.	Undertaking on legal valid and attested stamp paper Rs. 100/- that the firm is not blacklisted by any of Provincial or Federal Government Department, Agency, Organization or autonomous body or Private Sector Organization anywhere in Pakistan.	Yes	Yes	Yes	Yes
9.	Embassy Attested Exclusively Authorized Agent / Sole Distributor Certificate by the Original Equipment Manufacturer (OEM).	No	No	Yes	Yes
10.	Sub-Distributor Certificate from Sole Distributor having consent of OEM. (Where Applicable)	N/A	Yes	N/A	N/A
11.	Proof for having minimum 5 years of experience in supply, installation, commissioning of biomedical equipment.	Yes	Yes	Yes	Yes



12.	At least 05 verifiable Past Performance Certificates by the Procuring Agency of quoted products issued against the same Supply Orders.	No	No	Yes	Yes
13.	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes	Yes	Yes	Yes
14.	Originally signed documentary evidence from the Manufacturer that the quoted brand is the Original Equipment Manufacturer (OEM) with indication of manufacturing site and its location.	Yes	Yes	Yes	Yes
15.	Up to Date After Sales Service Certificate from OEM/Principal	Yes	Yes	Yes	Yes
16.	Availability of relevant Tools and Testing / Calibration Equipment.	Yes	Yes	Yes	Yes
17.	A Certificate from the OEM that the installation will be conducted in conformity with the system requirements.	Yes	Yes	Yes	Yes
18.	Consortiums or Joint Ventures (JVs) (Where applicable)	N/A	N/A	N/A	N/A
19.	Complete Package / Tender	No	No	Yes	No
Remarks:		Not Eligible	Not Eligible	Eligible	Not Eligible
<b>PART – II</b> <b>KNOCKED DOWN CRITERIA</b> <b>(PRODUCT EVALUATION)</b>					
Item Ad. No.	HOSPITAL LAUNDRY	M/s Total Technologies Pvt. Ltd.	M/s Pakistan Oxygen	M/s Mediland Pakistan Pvt. Ltd.	M/s Med Engineering
	Brand	Girbau, Ghidini, Juki	Ghidini	Domus	Primer
55	Quoted Model	As Per Compliance Sheet	As Per Compliance Sheet	As Per Compliance Sheet	As Per Compliance Sheet
1.	Country of Manufacturer				
2.	Country of Origin of Product				
3.	Compliance with defined Quality Standards	(Compliance) CE/ISO	(Non Complied) Not Attached	(Compliance) CE/ISO	(Compliance) CE/ISO

4.	Specifications Compliance Feature Wise	Compliance	(Non Complied) 1. No product catalogue attached. 2. Country of Manufacturer and Origin of Products not mentioned in technical offer.	Compliance	(Non Complied) 1. The furniture and accessory items quoted locally.
5.	Technical Eligibility of Product	Eligible	Not Eligible	Eligible	Not Eligible
6.	Technical Eligibility of Firm	Not Eligible	Not Eligible	Eligible	Not Eligible
<b>Technical Status:</b>		<b>NON RESPONSIVE</b>	<b>NON RESPONSIVE</b>	<b>RESPONSIVE</b>	<b>NON RESPONSIVE</b>

COMPLIANCE SHEET					
No.	Parameters	M/s Total Technologies Pvt. Ltd.	M/s Pakistan Oxygen	M/s Mediland Pakistan Pvt. Ltd.	M/s Med Engineering
1	Washer Extractor	Model: HS 6057 Girbau, Spain	Model: N/A Lapauw	Model: DHS-60 Domus, Spain	Model: LS-60C T2 Primer, Spain
2	Drying Tumbler	Model: ED 900 Girbau, Spain	Model: N/A Lapauw	Model: DTT-45 Domus, Spain	Model: DO45 T2 Primer, Spain
3	Flat Work Dryer Ironer	Model: PB 5119 Girbau, Spain	Model: N/A Lapauw	Model: CM-2650 Touch II Domus, Spain	Model: C-200/50 Primer, Spain
4	Spotting Table	Model: GBS 88F Ghidini, Italy	Local	Model: DES-F Domus, Spain	Model: DES-F Primer, Spain
5	Utility Press for Shirt Collars and Cuffs	Model: P88CP Ghidini, Italy	Model: N/A Ghidini	Model: PUN-C C&P Domus, Spain	Model: PRE-CF Primer, Spain
6	Utility Press	Model: P88 Ghidini, Italy	Model: N/A Ghidini	Model: PUN-C Domus, Spain	Model: PRE-U Primer, Spain



7	<b>Hand Iron with Ironing Table</b>	Model: AMBRIA Ghidini, Italy	Model: N/A Ghidini	Model: TPC-A Domus, Spain	Model: FDA/A Primer, Spain
8	<b>Weighing Scale with Ramp (Locally Supplied)</b>	Local	Local	Local	Local
9	<b>Reverse Osmosis System (Locally Supplied) In Case of Electric Heated Laundry</b>	Local	Local	Local	Local
10	<b>Water Softener System (Optional)</b>	Local Quoted	Not mentioned	Local Quoted	Local Quoted
11	<b>Sewing Machine</b>	Model: DDL-7000AS-7-NBK Juki/Japan	Local	Local	Imported Locally Supplied Singer, Juki Jack etc.
12	<b>Packing/Sorting Table</b>	Model: 10547139 MMM, Germany	Model: N/A Mediteck	Model: N/A Hupfer, Germany	Customized-1 Med Engineering (Local)
13	<b>Dirty Linen Collection Trolleys</b>	Model: RB 416 Girbau, Spain	Model: N/A Mediteck	Model: N/A Domus, Spain	Customized-2 Med Engineering (Local)
14	<b>Dirty Linen Trolleys for Working Inside Dirty Area</b>	Model: WET linen trolley 400L Girbau, Spain	Model: N/A Mediteck	Model: N/A Domus, Spain	Customized-3 Med Engineering (Local)
15	<b>Clean Linen Trolleys for Working Inside Clean Area</b>	Model: WET linen trolley 400L Girbau, Spain	Model: N/A Mediteck	Model: N/A Domus, Spain	Customized-4 Med Engineering (Local)
16	<b>Storage Racks</b>	Model: Shelf Trolley with drop down front Girbau, Spain	Model: N/A Mediteck	Model: N/A Hupfer, Germany	Customized-5 Med Engineering (Local)

Item No. 56

Item Name: Medical Gases

Tender Name: Procurement Of Bio-Medical Equipment Under PSDP Project "Establishment of 50 Bedded Cardiac Hospital, Gilgit".

PART - I KNOCKED DOWN CRITERIA (BID EVALUATION & VENDOR EVALUATION)					
Sr. #	Evaluation Parameters	M/s Pakistan Oxygen Limited	M/s Jamil Enterprises	M/s Shiga Trader	M/s Multan Chemicals
1.	Bid Form	Yes	Yes	Yes	Yes
2.	Bidder's Information Form	Yes	Yes	Yes	Yes
3.	Documentary proof active of National Tax Number (NTN) and General Sales Tax (GST) Number and Registration Professional Tax Certificate.	Yes	Yes	Yes	Yes
4.	Original Bid Security (In form of CDR) & Form	Yes	Yes	Yes	Yes
5.	Bid validity time is as per demand	No	Yes	No	Yes
6.	Warranty Form as per Tender	Yes	Yes	Yes	Yes
7.	Delivery as per Schedule of Requirement		Yes	Yes	Yes
8.	Undertaking on legal valid and attested stamp paper Rs. 100/- that the firm is not blacklisted by any of Provincial or Federal Government Department, Agency, Organization or autonomous body or Private Sector Organization anywhere in Pakistan.	Yes	Yes	Yes	Yes
9.	Embassy Attested Exclusively Authorized Agent / Sole Distributor Certificate by the Original Equipment Manufacturer (OEM).	No (Embassy attested Certificate not provided)	Yes (For M/s Total Technologies Pvt. Ltd., M/s Claris Medical)	Yes (Only for available for BERG Kompressoren GmbH) not for other quoted brands	No (Embassy attested Certificate not provided)
10.	Sub-Distributor Certificate from Sole Distributor having consent of OEM. (Where Applicable)	N/A	Yes	N/A	N/A



11.	Proof for having minimum 5 years of experience in supply, installation, commissioning of biomedical equipment.	Yes	Yes	Yes	No
12.	At least 05 verifiable Past Performance Certificates by the Procuring Agency of quoted products issued against the same Supply Orders.	Yes	Yes	No	No (Not a single Oxygen Generation Plant and Medical Air Compressors reference available for the said firm).
13.	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes	Yes	Yes	Yes
14.	Originally signed documentary evidence from the Manufacturer that the quoted brand is the Original Equipment Manufacturer (OEM) with indication of manufacturing site and its location.	Yes	Yes	Yes	Yes
15.	Up to Date After Sales Service Certificate from OEM/Principal	Yes	Yes	Yes	Yes
16.	Availability of relevant Tools and Testing / Calibration Equipment.	Yes	Yes	Yes	Yes
17.	A Certificate from the OEM that the installation will be conducted in conformity with the system requirements.	Yes	Yes	Yes	Yes
18.	Consortiums or Joint Ventures (JVs) (Where applicable)	N/A	N/A	No (Not as per the format of BD) M/s Shiga Traders and Ask Associates)	N/A
19.	Complete Package / Tender	No	Yes	No	No
Remarks:		Not Eligible	Eligible	Not Eligible	Not Eligible
<b>PART – II</b> <b>KNOCKED DOWN CRITERIA</b> <b>(PRODUCT EVALUATION)</b>					

Item Ad. No.	Medical Gases	M/s Pakistan Oxygen Limited	M/s Jamil Enterprises	M/s Shiga Trader	M/s Multan Chemicals
	Brand	Amico, ACBM, Sumer, Lawton	Precision, LM Medical, Worthington, Airsep Corporation	CPX, BERG Kompressoren GmbH & Oxywise Single Line ESA Medical Oxygen Generator,	Phoenix, Connect Medical System, DZ, Oxyone, Tonoli, Hydro Gaz Med, Mz Liberc
56	Quoted Model	As per Technical Offer	As per Technical Offer	As per Technical Offer	As per Technical Offer
1.	Country of Manufacturer	USA/China	UK, Italy, Austria, USA	UK, Japan, Slovakia, Germany	UK, Italy, Poland
2.	Country of Origin of Product	USA/China	UK, Italy, Austria, USA	UK, Japan, Slovakia, Germany	UK, Italy, Poland
3.	Compliance with defined Quality Standards	(Compliance) CE/ISO	(Compliance) ISO/FDA	(Compliance) CE/ISO	(Compliance) CE/ISO
4.	Specifications Compliance Feature Wise	(Non Complied)  6. The PSA is from Chines manufacturer, which isn't the requirement. 7. Heat Exchanger for heat utilization not quoted.	Compliance	(Non Complied)  1. Dual offer for PSA, i.e., BERG Kompressoren GmbH & Oxywise Single Line ESA Medical Oxygen Generator 0150. 2. Medical gas pipeline brand not quoted. 3. Heat Exchanger for heat utilization not	(Non Complied)  7. No past performance reference for Oxygen Generation Plant and Medical Air Compressors. reference available for the said firm. 8. Heat Exchanger for heat utilization not quoted.



				quoted. 4. Medical gas pipeline brand not quoted.	
5.	Technical Eligibility of Product	Not Eligible	Eligible	Not Eligible	Not Eligible
6.	Technical Eligibility of Firm	Not Eligible	Eligible	Not Eligible	Not Eligible
<b>Technical Status:</b>		<b>NON RESPONSIVE</b>	<b>RESPONSIVE</b>	<b>NON RESPONSIVE</b>	<b>NON RESPONSIVE</b>

Item No. 58

Item Name: Dental Treatment

Tender Name: Procurement Of Bio-Medical Equipment Under PSDP Project "Establishment of 50 Bedded Cardiac Hospital, Gilgit".

PART – I KNOCKED DOWN CRITERIA (BID EVALUATION & VENDOR EVALUATION)					
Sr. #	Evaluation Parameters	M/s Irfan Brothers International	M/s Zodiac International	M/s Jamil Enterprises	M/s Executive Enterprise
1.	Bid Form	Yes	Yes	Yes	Yes
2.	Bidder's Information Form	Yes	Yes	Yes	Yes
3.	Documentary proof active of National Tax Number (NTN) and General Sales Tax (GST) Number and Registration Professional Tax Certificate.	Yes	Yes	Yes	N/A
4.	Original Bid Security (In form of CDR) & Form	Yes	Yes	Yes	Yes
5.	Bid validity time is as per demand	Yes	Yes	Yes	Yes
6.	Warranty Form as per Tender	Yes	No (Conditional Warranty)	Yes	Yes
7.	Delivery as per Schedule of Requirement		Yes	Yes	Yes
8.	Undertaking on legal valid and attested stamp paper Rs. 100/- that the firm is not blacklisted by any of Provincial or Federal Government Department, Agency, Organization or autonomous body or Private Sector Organization anywhere in Pakistan.	No (Unsigned attached)	Yes	Yes	No (From M/s Combined Engineering not from M/s Executive Enterprise)
9.	Embassy Attested Exclusively Authorized Agent / Sole Distributor Certificate by the Original Equipment Manufacturer (OEM).	Yes (For Human Healthcare)	No (Embassy attested certificate not provided) (Expired certificate for Owandy Radiology)	No (For M/s Total Technologies Pvt Ltd. & M/s Advance Western)	Yes (For Combined Engineering)



10	Sub-Distributor Certificate from Sole Distributor having consent of OEM. (Where Applicable)	Yes	N/A	Yes	N/A
11	Proof for having minimum 5 years of experience in supply, installation, commissioning of biomedical equipment.	Yes	Yes	Yes	Yes
12	At least 05 verifiable Past Performance Certificates by the Procuring Agency of quoted products issued against the same Supply Orders.	Yes	No	Yes	Yes
13	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes	Yes	Yes	Yes
14	Originally signed documentary evidence from the Manufacturer that the quoted brand is the Original Equipment Manufacturer (OEM) with indication of manufacturing site and its location.	Yes	No	Yes	Yes
15	Up to Date After Sales Service Certificate from OEM/Principal	Yes	Yes	Yes	Yes
16	Availability of relevant Tools and Testing / Calibration Equipment.	Yes	Yes	Yes	Yes
17	A Certificate from the OEM that the installation will be conducted in conformity with the system requirements.	Yes	Yes	Yes	Yes
18	Consortiums or Joint Ventures (JVs) (Where applicable)	N/A	N/A	N/A	N/A
19	Complete Package / Tender	No	No	Yes	No
Remarks:		Not Eligible	Not Eligible	Eligible	Not Eligible

**PART – II**  
**KNOCKED DOWN CRITERIA**  
**(PRODUCT EVALUATION)**

Item Ad. No.	Dental Treatment	M/s Irfan Brothers International	M/s Zodiac International	M/s Jamil Enterprises	M/s Executive Enterprise
	Brand	As per Compliance Sheet	As per Compliance Sheet	As per Compliance Sheet	As per Compliance Sheet
58	Quoted Model				
1.	Country of Manufacturer				

2.	Country of Origin of Product				
3.	Compliance with defined Quality Standards	(Non Complied) CE (MDD) & other certification for Dental Unit not available.	(Non Complied) CE (MDD) & other certification for Dental Unit not available.	(Compliance) CE (MDD) & MHLW for Dental Unit	(Non Complied) CE (MDD) & other certification for Dental Unit not available.
4.	Specifications Compliance Feature Wise	(Non Complied) 1. For OPG with ceph is offered 61-85 KV, 3-10 MA instead of 60-90 KV, 3-16 mA – or more	(Non Complied) 1. Country of manufacturer for LED Cure Light is from R.O.C instead of USA or Europe or Japan. 2. Ultrasonic cleaner 28 lit not quoted.	Compliance	Compliance
2.	Technical Eligibility of Product	Not Eligible	Not Eligible	Eligible	Not Eligible
3.	Technical Eligibility of Firm	Not Eligible	Not Eligible	Eligible	Not Eligible
Technical Status:		NON RESPONSIVE	NON RESPONSIVE	RESPONSIVE	NON RESPONSIVE

COMPLIANCE SHEET					
No.	Parameters	M/s Irfan Brothers International	M/s Zodiac International	M/s Jamil Enterprises	M/s Executive Enterprise
1	Dental Treatment Unit with Scaler, Electric Motor, Air Turbine	Model: Ultimate Comfort Ritter Concept GmbH Germany	Model: Friend Easy Swident Italy Model: M3 Champion Marathon South Korea	Model: Clesta EIII Belmont Japan	Model: R7 Anthos Italy Model: Xsmart Plus Dentply Sirona Germany



2	<b>Intra Oral X-ray</b>	Model: RIX 70 DC Trident S.r.l. Italy	Model: RX Owandy France	Model: PHT-XIIS Belmont Japan	Model: best X DC New Life Radiology Italy
3	<b>Intra Oral Digital Sensor Size 2</b>	Model: I-View Gold Trident S.r.l. Italy	Model: Opteao with Sensor Position Holder Owandy France	Model: SOPIX2 Aceton Imaging France	Model: Digisens New Life Radiology Italy
4	<b>Light Cure Unit</b>	Model: Curing Light LED Ritter Concept GmbH Germany	Model: B-Cure R.O.C	Model: LED E Carlo de Girogi Italy	Model: T-LED New Life Radiology Italy
5	<b>Oil Free Air Compressor</b>	Model: S45 Ritter Concept GmbH Germany	Model: MIRAGE70 4 Tek Italy	Model: Genesis s MGF Italy	Model: Air 30 Light Meta Austria
6	<b>Electric Suction Unit</b>	Model: VAC1S Ritter Concept GmbH Germany	Model: A006/L 4 Tek Italy	Model: Durr Dental V600 Durr Dental Germany	Model: Excom Hybrid Meta Austria
7	<b>Class B touch screen autoclave 23 lit with short cycle time</b>	Model: Clean Tec CB 23 Ritter Concept GmbH Germany	Model: E8 Biltin Printer Euronda Italy	Model: Vacuclave 123 Melag Germany	Model: B22 Classic Mocom Italy
8	<b>Ultrasonic cleaner 28 lit</b>	Model: VU28H1 SMEG Instruments Italy	Not Quoted	Model: NLX Nano NXK Japan	Model: B22 Classic Mocom Italy
9	<b>OPG with ceph</b>	Model: X View 2D Pan Trident S.r.l. Italy	Model: IMAX Ceph Pro 20 Owandy France	Model: Orthophos S 2D Ceph Dentsply Sirona Germany	Hyprion X5 Myray Cefla SC Italy
10	<b>Cabinet system with sink</b>	Cabinet System with Sink Pak Made (Local)	Cabinet System with Sink Pak Made (Local)	Cabinet System with Sink Pak Made (Local)	Cabinet System with Sink Pak Made (Local)

Item No. 59  
Item Name: Complete Set of Instruments and Surgical Instruments (Complete Cardiac Surgical Set)  
Tender Name: Procurement Of Bio-Medical Equipment Under PSDP Project "Establishment of 50 Bedded Cardiac Hospital, Gilgit".

PART – I KNOCKED DOWN CRITERIA (BID EVALUATION & VENDOR EVALUATION)					
Sr. #	Evaluation Parameters	M/s Itefaque Multi Traders	M/s Zodiac International	M/s Jamil Enterprises	M/s Executive Enterprise
1.	Bid Form	No (Incomplete offer)	Yes	Yes	Yes
2.	Bidder's Information Form	Yes	Yes	Yes	Yes
3.	Documentary proof active of National Tax Number (NTN) and General Sales Tax (GST) Number and Registration Professional Tax Certificate.	Yes	Yes	Yes	Yes
4.	Original Bid Security (In form of CDR) & Form	Yes	No (Less Amount CDR)	Yes	Yes
5.	Bid validity time is as per demand	Yes	Yes	No	Yes
6.	Warranty Form as per Tender	Yes	No (Conditional Warranty)	Yes	Yes
7.	Delivery as per Schedule of Requirement		Yes	Yes	Yes
8.	Undertaking on legal valid and attested stamp paper Rs. 100/- that the firm is not blacklisted by any of Provincial or Federal Government Department, Agency, Organization or autonomous body or Private Sector Organization anywhere in Pakistan.	Yes	No (Photocopy attached)	Yes	Yes
9.	Embassy Attested Exclusively Authorized Agent / Sole Distributor Certificate by the Original Equipment Manufacturer (OEM).	Yes (For Claris Medical)	Yes	No (For M/s Idea Medical) (The manufacturer Swantia Medical Pvt Ltd. brand quoted)	No (The manufacturer Swantia Medical Pvt Ltd. brand quoted)



10	Sub-Distributor Certificate from Sole Distributor having consent of OEM. (Where Applicable)	Yes	N/A	Yes	N/A
11	Proof for having minimum 5 years of experience in supply, installation, commissioning of biomedical equipment.	Yes	Yes	Yes	Yes
12	At least 05 verifiable Past Performance Certificates by the Procuring Agency of quoted products issued against the same Supply Orders.	Yes	No	Yes	Yes
13	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes	Yes	Yes	Yes
14	Originally signed documentary evidence from the Manufacturer that the quoted brand is the Original Equipment Manufacturer (OEM) with indication of manufacturing site and its location.	Yes	No	Yes	Yes
15	Up to Date After Sales Service Certificate from OEM/Principal	Yes	Yes	Yes	Yes
16	Availability of relevant Tools and Testing / Calibration Equipment.	Yes	Yes	Yes	Yes
17	A Certificate from the OEM that the installation will be conducted in conformity with the system requirements.	Yes	Yes	Yes	Yes
18	Consortiums or Joint Ventures (JVs) (Where applicable)	N/A	N/A	N/A	N/A
19	Complete Package / Tender	No (Complete Set of Instruments related to dental not quoted)	No	No	No
Remarks:		Not Eligible	Not Eligible	Not Eligible	Not Eligible
<b>PART – II</b> <b>KNOCKED DOWN CRITERIA</b> <b>(PRODUCT EVALUATION)</b>					
Item Ad. No.	Complete Set of Instruments and Surgical Instruments (Complete Cardiac Surgical Set)	M/s Itefaque Multi Traders	M/s Zodiac International	M/s Jamil Enterprises	M/s Executive Enterprise

	Brand	Asanus Medizintechnik GmbH	Riaz & Sajjad Surgical Pvt Ltd.	Swantial Medical Pvt Ltd.	Swantia Medical Pvt Ltd.
59	Quoted Model	As per Technical Offer	As per Technical Offer	As per Technical Offer	As per Technical Offer
1.	Country of Manufacturer	Germany	Pakistan	Pakistan	Pakistan
2.	Country of Origin of Product				
3.	Compliance with defined Quality Standards	(Compliance) CE(MDR)	(Non complied) No relevant certificate attached	(Non complied) FDA Listing instead of FDA 510K / CE	(Non complied) FDA Listing instead of FDA 510K / CE
4.	Specifications Compliance Feature Wise	Compliance	Compliance	(Non complied)  Both M/s Jamil Enterprises and M/s Executive Enterprise quoted M/s Swantia Medial Pvt Ltd. brand.	(Non complied)  Both M/s Jamil Enterprises and M/s Executive Enterprise quoted M/s Swantia Medial Pvt Ltd. brand.
5.	Technical Eligibility of Product	Eligible	Not Eligible	Not Eligible	Not Eligible
6.	Technical Eligibility of Firm	Not Eligible	Not Eligible	Not Eligible	Not Eligible
Technical Status:		<b>NON RESPONSIVE</b>	<b>NON RESPONSIVE</b>	<b>NON RESPONSIVE</b>	<b>NON RESPONSIVE</b>



Item No. 60

Item Name: Gyne Instrument Complete

Tender Name: Procurement Of Bio-Medical Equipment Under PSDP Project "Establishment of 50 Bedded Cardiac Hospital, Gilgit".












PART - I KNOCKED DOWN CRITERIA (BID EVALUATION & VENDOR EVALUATION)						
Sr. #	Evaluation Parameters	M/s Itefaque Multi Traders	M/s Zodiac International	M/s Jamil Enterprises	M/s Executive Enterprise	M/s Sahaab International
1.	Bid Form	Yes	Yes	Yes	Yes	Yes
2.	Bidder's Information Form	Yes	Yes	Yes	Yes	Yes
3.	Documentary proof active of National Tax Number (NTN) and General Sales Tax (GST) Number and Registration Professional Tax Certificate.	N/A	Yes	Yes	Yes	Yes
4.	Original Bid Security (In form of CDR) & Form	Yes	No (Less Amount CDR)	Yes	Yes	Yes
5.	Bid validity time is as per demand	Yes	Yes	No	Yes	Yes
6.	Warranty Form as per Tender	Yes	No (Conditional Warranty)	Yes	Yes	Yes
7.	Delivery as per Schedule of Requirement		Yes	Yes	Yes	Yes
8.	Undertaking on legal valid and attested stamp paper Rs. 100/- that the firm is not blacklisted by any of Provincial or Federal Government Department, Agency, Organization or autonomous body or Private Sector Organization anywhere in Pakistan.	Yes	No (Photocopy attached)	Yes	Yes	Yes
9.	Embassy Attested Exclusively Authorized Agent / Sole Distributor Certificate by the Original Equipment Manufacturer (OEM).	Yes (For Claris Medical)	Yes	No (For M/s Idea Medical) (The manufacturer	No (The manufacturer Swantia	No

				Swantia Medical Pvt Ltd. brand quoted)	Medical Pvt Ltd. brand quoted)	
10.	Sub-Distributor Certificate from Sole Distributor having consent of OEM. (Where Applicable)	Yes	N/A	Yes	N/A	N/A
11.	Proof for having minimum 5 years of experience in supply, installation, commissioning of biomedical equipment.	Yes	Yes	Yes	Yes	Yes
12.	At least 05 verifiable Past Performance Certificates by the Procuring Agency of quoted products issued against the same Supply Orders.	Yes	No	Yes	Yes	No
13.	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes	Yes	Yes	Yes	Yes
14.	Originally signed documentary evidence from the Manufacturer that the quoted brand is the Original Equipment Manufacturer (OEM) with indication of manufacturing site and its location.	Yes	No	Yes	Yes	No
15.	Up to Date After Sales Service Certificate from OEM/Principal	Yes	Yes	Yes	Yes	Yes
16.	Availability of relevant Tools and Testing / Calibration Equipment.	Yes	Yes	Yes	Yes	Yes
17.	A Certificate from the OEM that the installation will be conducted in conformity with the system requirements.	Yes	Yes	Yes	Yes	Yes
18.	Consortiums or Joint Ventures (JVs) (Where applicable)	N/A	N/A	N/A	N/A	N/A
19.	Complete Package / Tender	Yes	No	No	No	No Bid Pages are not signed and stamped as













						required in BD ITB Clause 19.
	<b>Remarks:</b>	<b>Eligible</b>	<b>Not Eligible</b>	<b>Not Eligible</b>	<b>Not Eligible</b>	<b>Not Eligible</b>
<b>PART – II</b> <b>KNOCKED DOWN CRITERIA</b> <b>(PRODUCT EVALUATION)</b>						
Item Ad. No.	<b>GYNE INSTRUMENT COMPLETE</b>	<b>M/s Itefaque Multi Traders</b>	<b>M/s Zodiac International</b>	<b>M/s Jamil Enterprises</b>	<b>M/s Executive Enterprise</b>	<b>M/s Sahaab International</b>
	Brand	Asanus Medizintechnik GmbH	Riaz & Sajjad Surgical Pvt Ltd.	Swantial Medical Pvt Ltd.	Swantia Medical Pvt Ltd.	Medalion Equipment
<b>60</b>	Quoted Model	As per Technical Offer	As per Technical Offer	As per Technical Offer	As per Technical Offer	As per Technical Offer
<b>1.</b>	Country of Manufacturer	Germany	Pakistan	Pakistan	Pakistan	UK
<b>2.</b>	Country of Origin of Product					
<b>3.</b>	Compliance with defined Quality Standards	(Compliance) CE (MDR)	(Non complied) No relevant certificate attached	(Non complied) FDA Listing instead of FDA 510K / CE	(Non complied) FDA Listing instead of FDA 510K / CE	(Non complied) CE (MDD) certificate not from authorized EU body is attached.
<b>4.</b>	Specifications Compliance Feature Wise	Compliance	Compliance	(Non complied) Both M/s Jamil Enterprises and M/s Executive Enterprise quoted M/s Swantia Medial Pvt Ltd. brand.	(Non complied) Both M/s Jamil Enterprises and M/s Executive Enterprise quoted M/s Swantia Medial Pvt Ltd. brand.	(Non Complied) 6- The quoted model is not available on official website of quoted manufacturer and the required specifications

						were copied on product catalog attached to the technical offer.
5.	Technical Eligibility of Product	Eligible	Not Eligible	Not Eligible	Not Eligible	Not Eligible
6.	Technical Eligibility of Firm	Eligible	Not Eligible	Not Eligible	Not Eligible	Not Eligible
<b>Technical Status:</b>		<b>RESPONSIVE</b>	<b>NON RESPONSIVE</b>	<b>NON RESPONSIVE</b>	<b>NON RESPONSIVE</b>	<b>NON RESPONSIVE</b>



<b>13</b>	<b>RECOMMENDATIONS OF TECHNICAL EVALUATION:</b>
	It is hereby recommended to open the Financial Proposals of responsive firms after fulfilling codal formalities as per GB PPRA Rules.

<b>S#</b>	<b>Name</b>	<b>Designation</b>	<b>Signatures</b>
<b>1</b>	<b>Dr Syed Sadiq Shah</b>	Director, EPI-GB	
<b>2</b>	<b>Engr. Shahid Masood</b>	PD ,50 bedded Cardiac Hospital Gilgit	
<b>3</b>	<b>Syed Ishtiaq Hussain</b>	Director, SHPI-GB	
<b>4</b>	<b>Dr. Faheem Iqbal</b>	Consultant Cardiologist 50 bedded Cardiac Hospital Gilgit	
<b>5</b>	<b>Aziz Haider</b>	DD(P), DHS Baltistan	
<b>6</b>	<b>Engr. Shabbir Hussain</b>	BME DHS(P), Gilgit	
<b>7</b>	<b>Muhammad Ali</b>	Finance Officer, FD-GB	
<b>8</b>	<b>Muhammad Harith</b>	RO, P&D D GB	
<b>9</b>	<b>Engr. Mohsin Ali</b>	BME Shifa Int. Hosp Islamabad	
<b>10</b>	<b>Engr. Rizwan</b>	BME RIC Rawalpindi	
<b>11</b>	<b>Engr. Umair Kabeer</b>	BME PIC Lahore	