

**Technical Evaluation Report****As per GBPPRA Rule 38**

<b>Name of Procuring agency</b>	Medical Superintendent PHQ Hospital Gilgit, Health Department Gilgit-Baltistan
<b>Title of Procurement</b>	Procurement Of Anesthesia Machine With All Accessories For PHQ Hospital Gilgit Under Development Scheme " <b><u>CONSTRUCTION OF 'C' CLASS DISPENSARY AT PAHOTE AND MISSING FACILITIES PHQ HOSPITAL GILGIT</u></b> "
<b>Method of Procurement</b>	Single Stage Two Envelope Method
<b>PPRA Advertisement (TS) No</b>	TSE - 20240923230
<b>Newspaper Tender Advertisement No</b>	National English paper -DAILY Pakistan Today (Islamabad)No.IDGB-697/24 dated 17/9/2024 National Urdu paper Daily Dunia (Islamabad No.IDGB-697/24, dated 17/09/2024 Local paper Daily K-2 GB No. IDGB (GLT) 697/24
<b>Tender Submission date and time</b>	8-10-2024 at 11:00 am
<b>Tender Opening date and time</b>	8-10-2024 at 11:30am ( Technical Proposal)
<b>NO. of Bid received</b>	01
<b>Detail of evaluation</b>	As mentioned below






**Administrative Evaluation to Determine the Strength of the Bidder/Pre-Qualification Clauses**

Name of Bidder : M/S Armaghan Trader JV M/S Jamal Enterprises			
S.#	Evaluation Parameters	Relevant Document	Leading firm: M/S Armaghan Traders JV Associate firm: M/S Jamal Enterprises
1.	<b>Valid NTN (where applicable)</b>	NTN Certificate	Not Applicable
2.	<b>Valid GST Registration (where applicable)</b>	GST Certificate	Not Applicable
3.	<b>Firm Registration</b>	PEC/ Form-C/ Form- D/SECP certificate	Form -C Submitted
4.	<b>Tender Fee</b>	Tender Purchase Receipt	Tender Purchase Receipt Submitted
5.	<b>Bid Security ( CDR)</b>	Original bid Security (CDR) with Financial Offer & a photocopy of bid security must be attach with technical offer.	photocopy of bid security Submitted

6.	<b>NON-Conditional BID</b>	Certificate from the Bidder/Firm that the bidder/firm has no any condition against the submitted bid.	Accepted All the terms and conditions of bid document.
7.	<b>The bidding firm shall provide audited financial statements for the last years where taxes are applicable. Bank Statement of last years where taxes are not applicable.</b>	a. Audited Financial Statement for the year 2022-2023. b. bank Statement of last one year till one week prior to bid opening date is mandatory.	Submitted Bank Statement
8.	<b>NON- Blacklisting Firm</b>	An affidavit of Rs.100/- from bidder and all partners of JV stating that their firm is not blacklisted by any of the Federal and Provincial Government or organizations of the State/ Central Government in Pakistan.	Submitted Non Blacklisting under taking on An affidavit of Rs.100/-
9.	<b>Authorization letter of the quoted product</b> (Verifiable having valid email address and contact No)	a. Original Manufacturer: must provide Prove of Orig Manufacturer Certificate b. Sole Distributor: valid legal enforceable exclusive Authorization letter of product from Manufacturer c. Sub Distributor: Authorization letter of Product from sole Distributor and authorization letter of product from Manufacturer to sole distributor as well.	Submitted Authorization letter of Product from sole Distributor.
9.	<b>The bidding firm shall have a minimum of three (03) years of experience in the supply, delivery, installation, testing and commissioning, training and maintenance of medical Equipment in Health Sector of Pakistan.</b>	The bidding firm should provide at least three (03) verifiable past supply orders from Health Sector of Pakistan as proof of experience.	Submitted 3 supply orders from Health department of GB by leading firm Armaghan Traders.
10.	<b>Joint Venture Form</b> (in case of JV)	Copy of JV Agreement in original with lead bidder declaration form.	Leading firm: M/S Armaghan Traders JV Associate firm: M/S Jamal Enterprises , JV Agreement in original submitted
11	<b>Acceptance of the terms &amp; conditions of bidding document.</b>	Accepting terms and conditions, of bidding documents dully singed and stamp must be submitted.	<u>submitted</u>

Handwritten signatures and stamps are present at the bottom of the page, including a large signature on the left, a smaller signature in the center, and a stamp on the right.

13.	<b>Engineering and Technical capabilities</b>	(Affidavit on Stamp paper Rs 50, with the technical engineer staff detail must be submitted). At least ONE biomedical engineer having PEC license in Biomedical engineering.	Affidavit on Stamp paper Rs 50 submitted including detail of biomedical engineer having PEC license in Biomedical engineering
<b><u>Eligibility for further evaluation:</u></b>		<b>Eligible for further evaluation Subject to Provision of Authorization Letter Of Product from Manufacturer to sole distributor.</b>	
<b><u>Note:</u></b>		The procuring agency shall verify the documents submitted by the bidder and shall disqualify the supplier or contractor if it finds, at any time, that the information submitted by him concerning his qualification as supplier or contractor was false and materially inaccurate or incomplete under GBPPRA rule 21.	

Name	Designation		Signature
Dr. Syed Sartaj Hussain	Medical Superintendent PHQ Hospital Gilgit	Chairman	
Dr. Matam Shah	Chief Consultant Anesthetics PHQ Hospital Gilgit	Member	
Engr. Shabbir Hussain	Biomedical Engineer DHS P&P Gilgit Health Department	Member	
Mr. Zafar Iqbal	Section Officer Health Secretariat	Member	
Mr. Sohrab Khan	Store officer DHS Gilgit Health Department	Member	

9



**Category 2: Compliance to Purchaser's Technical Specifications & Requirements**

<b>A</b>	<b>Name of Model</b>	<b>A9800 ( Anesthesia Work Station)</b>		
<b>B</b>	<b>Name of Manufacturer/ Make</b>	<b>ORICARE Inc.</b>		
<b>C</b>	<b>Country of Manufacturer</b>	<b>USA</b>		
<b>D</b>	<b>Country of Origin Of Product</b>	<b>China</b>		
<b>S. No</b>	<b>Compliance to Purchaser's Technical Specifications (only one set of marks according to the highest obtained, after the comparison of tender specifications with the bidder's offered specifications of equipment, will be considered for evaluation)</b>	<b>Total points (100)</b>	<b>81/100</b>	
1	Full compliance to the Purchaser's specifications	50	50	50
	Compliant with one (01) major deviation	40		
	Compliant with two (02) major deviations	30		
2	Country of Origin of product : (Must attach Boucher clearly mentioning country as prove)		15	China 5
	USA, Japan, Europe	15		
	China, Korea & others	5		
	No Country Mentioned	K0		
3	Country of Manufacturer :		15	USA 15
	USA, Japan, Europe	15		
	China, Korea & others	5		
	No Country Mentioned/identified	K0		
3	Valid Product Certification: (Must attach Valid Product certificate as prove)		10	FDA & CE 5
	Certificates: US FDA with (K num), MHLW, CEMDD	Points		
	All three (3)	10		
	Any two (2)	5		
	Any One (1)	3		
	Any Fake/Altered Certificate/missing Certificate	K0		
4	Experience: The bidding firm should provide verifiable past supply orders from Health Sector of Pakistan as proof of experience		10	(6)

<table><tr><td>More than 6 years of Product Experience</td><td>10</td></tr><tr><td>More than 4 years and Less than 6 years Product experience</td><td>7</td></tr><tr><td>3 years of Product Experience</td><td>4</td></tr><tr><td>General Experience in Supply of Medical Equipment</td><td>2</td></tr></table>	More than 6 years of Product Experience	10	More than 4 years and Less than 6 years Product experience	7	3 years of Product Experience	4	General Experience in Supply of Medical Equipment	2		3 Years' experience of Leading Firm + General Experience in supply of medical testing equipment by Associate Firm.
More than 6 years of Product Experience	10									
More than 4 years and Less than 6 years Product experience	7									
3 years of Product Experience	4									
General Experience in Supply of Medical Equipment	2									
Total Obtained Marks/Secure		81/100								
Detail of Deviations in Technical Specifications: 1) The quoted model Oricare "A9800" Anesthesia Machine workstation provides general inhalation Anesthesia and ventilator support for Adolescents to Adults, And not covers the patient population of Peads & Neonate with deviation of having Tidal volume 20ml to 1500ml against the tender specification of tidal volume 5ml to 1400ml. As per FDA510K it is Substantially Equivalent to Tender specification of tidal volume 5ml to 1400ml.										
<b>Note: The quoted model Oricare "A9800" Anesthesia Machine workstation is considered technically responsive subject to provision of following technical features/specifications mandatory as per tender specifications.</b> 2) The quoted model Oricare "A9800" Anesthesia Machine workstation is consider technically responsive subject to provision of all ventilation modes including • Volume Preset Time Cycled Ventilation or Volume Control, Manual and spontaneous, Pressure Mode (PCV), Synchronized pressure controlled ventilation (SIMV-PC), Synchronized volume controlled ventilation (SIMV), PS with apnea back up, Breathing Mode Selection (Standby, Volume, Spontaneous and Pressure). 3) The quoted model Oricare "A9800" Anesthesia Machine workstation is consider technically responsive subject to provision of Oxygen Senor: Paramagnetic Technology 4) Note: Annual maintenance kits (needs to replace) will be included in the warranty period as per manufacturer's guidelines. 5) The quoted model Oricare "A9800" Anesthesia Machine workstation is consider technically responsive subject to provision of <b>VITAL SIGN MONITOR:</b> <ul style="list-style-type: none"><li>• Size of display minimum 17" touch screen or more for monitoring of vital sign parameters of Neonates, Peads and adults.</li><li>• Measurement of ECG 5 or 6 leads.</li><li>• NIBP with re-usable single hose cuff for neonates, Peads and Adults.</li><li>• SpO2 (Masimo Technology or Equivalent motion tolerant technology) with re-usable cable and sensors for neonates, Peads &amp; adult</li><li>• Temperature with nasal and skin probe.</li><li>• Respiration.</li><li>• Four Channel NIBP.</li><li>• Trend data 72 hours or more.</li><li>• Printer module for printing waveforms and patient data.</li><li>• Anesthetic Agent, CO2, O2 and N2O Monitoring (With monitor or within Anesthesia Machine) along with Gas consumption and uptake data.</li><li>• Monitoring of xMAC or MAC value. (with monitor or with in the anesthesia machine).</li><li>• EtCO2 main or side stream (Complete with all sensors probes, reusable).</li><li>• Provision of communication port for sharing and transfer of data to HIMS or HIS through HL7 protocol.</li><li>• 220V, 50 Hz operated.</li></ul>										

